

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062415\
 Data File : BE089940.D
 Acq On : 24 Jun 2015 19:38
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 25 03:54:08 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 24 09:27:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	94	0.00
2	1,4-Dioxane	1.000	0.966	3.4	95	0.00
3	n-Nitrosodimethylamine	1.000	0.890	11.0	91	0.00
4 S	2-Fluorophenol	1.000	0.792	20.8	79	0.00
5 S	Phenol-d6	1.000	0.815	18.5	82	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	95	0.00
7 S	Nitrobenzene-d5	1.000	0.861	13.9	91	0.00
8	Nitrobenzene	1.000	0.864	13.6	92	0.00
9	Naphthalene	1.000	0.922	7.8	95	0.00
10	Hexachlorobutadiene	1.000	0.938	6.2	96	0.00
11	2-Methylnaphthalene	1.000	0.899	10.1	94	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	93	0.00
13 S	2,4,6-Tribromophenol	1.000	0.766	23.4	72	0.00
14 S	2-Fluorobiphenyl	1.000	0.906	9.4	94	0.00
15	Acenaphthylene	1.000	0.928	7.2	93	0.00
16	Acenaphthene	1.000	0.950	5.0	94	0.00
17	Fluorene	1.000	0.993	0.7	93	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	91	0.00
19	4-Bromophenyl-phenylether	1.000	0.885	11.5	93	0.00
20	Hexachlorobenzene	1.000	0.913	8.7	93	0.00
21	Pentachlorophenol	1.000	0.777	22.3	76	-0.02
22	Phenanthrene	1.000	0.906	9.4	92	0.00
23	Anthracene	1.000	0.880	12.0	89	0.00
24	Fluoranthene	1.000	0.895	10.5	90	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	90	0.00
26	Pyrene	1.000	0.943	5.7	89	0.00
27 S	Terphenyl-d14	1.000	0.907	9.3	89	0.00
28	Benzo(a)anthracene	1.000	0.894	10.6	88	0.00
29	Chrysene	1.000	0.895	10.5	89	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.763	23.7	81	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.898	10.2	85	0.00
32 I	Perylene-d12	5.000	5.000	0.0	88	0.00
33	Benzo(b)fluoranthene	1.000	0.938	6.2	87	0.00
34	Benzo(k)fluoranthene	1.000	0.937	6.3	87	0.00
35 C	Benzo(a)pyrene	1.000	0.903	9.7	85	0.00
36	Dibenzo(a,h)anthracene	1.000	0.906	9.4	86	0.00
37	Benzo(g,h,i)perylene	1.000	0.919	8.1	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0