

Data Path : Z:\HPCHEM1\BNA_E\Data\BE062415\
 Data File : BE089940.D
 Acq On : 24 Jun 2015 19:38
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 25 03:54:08 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jun 24 09:27:42 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.675	0.586	13.2	95	0.00
3	n-Nitrosodimethylamine	0.899	0.800	11.0	91	0.00
4 S	2-Fluorophenol	0.903	0.715	20.8	79	0.00
5 S	Phenol-d6	1.383	1.128	18.4	82	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	95	0.00
7 S	Nitrobenzene-d5	0.372	0.321	13.7	91	0.00
8	Nitrobenzene	0.396	0.342	13.6	92	0.00
9	Naphthalene	1.112	1.026	7.7	95	0.00
10	Hexachlorobutadiene	0.176	0.165	6.2	96	0.00
11	2-Methylnaphthalene	0.761	0.684	10.1	94	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
13 S	2,4,6-Tribromophenol	0.064	0.041	35.9#	72	0.00
14 S	2-Fluorobiphenyl	1.457	1.321	9.3	94	0.00
15	Acenaphthylene	8.767	8.139	7.2	93	0.00
16	Acenaphthene	1.288	1.223	5.0	94	0.00
17	Fluorene	1.750	1.609	8.1	93	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
19	4-Bromophenyl-phenylether	0.204	0.180	11.8	93	0.00
20	Hexachlorobenzene	0.894	0.816	8.7	93	0.00
21	Pentachlorophenol	0.033	0.021	36.4#	76	-0.02
22	Phenanthrene	1.254	1.136	9.4	92	0.00
23	Anthracene	1.164	1.025	11.9	89	0.00
24	Fluoranthene	1.361	1.218	10.5	90	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
26	Pyrene	1.266	1.193	5.8	89	0.00
27 S	Terphenyl-d14	0.863	0.783	9.3	89	0.00
28	Benzo(a)anthracene	1.195	1.068	10.6	88	0.00
29	Chrysene	1.306	1.170	10.4	89	0.00
30	Bis(2-ethylhexyl)phthalate	0.102	0.078	23.5	81	0.00
31	Indeno(1,2,3-cd)pyrene	0.960	0.862	10.2	85	0.00
32 I	Perylene-d12	1.000	1.000	0.0	88	0.00
33	Benzo(b)fluoranthene	1.058	0.993	6.1	87	0.00
34	Benzo(k)fluoranthene	1.199	1.124	6.3	87	0.00
35 C	Benzo(a)pyrene	0.940	0.849	9.7	85	0.00
36	Dibenzo(a,h)anthracene	0.875	0.792	9.5	86	0.00
37	Benzo(g,h,i)perylene	0.941	0.865	8.1	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0