

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
 Data File : BE093356.D
 Acq On : 24 Jun 2017 11:32
 Operator : SJ/JU
 Sample : I3653-10DL 5X
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F5C53DL

Manual Integrations
 APPROVED

mohammad
 6/29/2017 8:33:59 AM

Quant Time: Jun 26 01:25:53 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 26 01:13:30 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	1401	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	7148	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.36	164	4297	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	11085	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	11651	0.40	ng/ul	0.00
20) Perylene-d12	23.66	264	11201	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.10	152	458	0.04	ng/ul	0.00
14) Fluoranthene-d10	19.12	212	1291	0.04	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.56	128	84565	4.768	ng/ul#	91
5) 2-Methylnaphthalene	12.18	142	13506	1.114	ng/ul	98
7) Acenaphthylene	14.08	152	741	0.040	ng/ul#	78
8) Acenaphthene	14.42	153	8951	0.602	ng/ul	97
9) Fluorene	15.42	166	9962	0.571	ng/ul	96
12) Phenanthrene	17.14	178	75266	2.468	ng/ul#	91
13) Anthracene	17.22	178	5499	0.207	ng/ul#	91
15) Fluoranthene	19.15	202	33722	0.902	ng/ul	83
17) Pyrene	19.51	202	24555	0.715	ng/ul#	88
18) Benzo(a)anthracene	21.24	228	7057	0.227	ng/ul#	85
19) Chrysene	21.29	228	6830	0.212	ng/ul	99
21) Benzo(b)fluoranthene	22.93	252	4940m	0.145	ng/ul	
22) Benzo(k)fluoranthene	22.97	252	2038m	0.056	ng/ul	
23) Benzo(a)pyrene	23.55	252	3234	0.101	ng/ul#	91
24) Indeno(1,2,3-cd)pyrene	26.18	276	1386	0.037	ng/ul#	92
26) Benzo(g,h,i)perylene	26.96	276	1049	0.033	ng/ul#	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
Data File : BE093356.D
Acq On : 24 Jun 2017 11:32
Operator : SJ/JU
Sample : I3653-10DL 5X
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5C53DL

Manual Integrations
APPROVED

mohammad
6/29/2017 8:33:59 AM

Quant Time: Jun 26 01:25:53 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 26 01:13:30 2017
Response via : Initial Calibration

