

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
 Data File : BE093360.D
 Acq On : 24 Jun 2017 14:01
 Operator : SJ/JU
 Sample : I3625-09
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F5D03

Manual Integrations
 APPROVED

mohammad
 6/29/2017 8:34:10 AM

Quant Time: Jun 26 01:41:03 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 26 01:13:30 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	1386	0.40	ng/ul	0.00
2) Naphthalene-d8	10.52	136	7293	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.36	164	4645	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	11462	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	11999	0.40	ng/ul	0.00
20) Perylene-d12	23.66	264	9656m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.10	152	3443	0.32	ng/ul	0.00
14) Fluoranthene-d10	19.12	212	13796	0.44	ng/ul	0.00
Target Compounds				Qvalue		
3) Naphthalene	10.56	128	6633	0.367	ng/ul#	93
5) 2-Methylnaphthalene	12.18	142	1742	0.141	ng/ul	99
7) Acenaphthylene	14.08	152	15600	0.783	ng/ul	97
8) Acenaphthene	14.42	153	4811	0.299	ng/ul	99
9) Fluorene	15.41	166	6463	0.343	ng/ul	91
12) Phenanthrene	17.13	178	149643	4.745	ng/ul#	89
13) Anthracene	17.22	178	41086	1.496	ng/ul#	90
15) Fluoranthene	19.15	202	392806	10.157	ng/ul	84
17) Pyrene	19.51	202	396016	11.191	ng/ul#	88
18) Benzo(a)anthracene	21.24	228	239478	7.482	ng/ul#	85
19) Chrysene	21.29	228	251785	7.584	ng/ul	99
21) Benzo(b)fluoranthene	22.93	252	353495	12.063	ng/ul	87
22) Benzo(k)fluoranthene	22.97	252	110921m	3.564	ng/ul	
23) Benzo(a)pyrene	23.56	252	209210m	7.551	ng/ul	
24) Indeno(1,2,3-cd)pyrene	26.18	276	129969m	4.051	ng/ul	
25) Dibenzo(a,h)anthracene	26.19	278	37891m	1.402	ng/ul	
26) Benzo(g,h,i)perylene	26.96	276	101395	3.685	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
Data File : BE093360.D
Acq On : 24 Jun 2017 14:01
Operator : SJ/JU
Sample : I3625-09
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F5D03

Manual Integrations
APPROVED

mohammad
6/29/2017 8:34:10 AM

Quant Time: Jun 26 01:41:03 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 26 01:13:30 2017
Response via : Initial Calibration

