

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_E\Data\BE062417\
 Data File : BE093353.D
 Acq On : 24 Jun 2017 9:32
 Operator : SJ/JU
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.424

Quant Time: Jun 26 01:11:59 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jun 24 04:29:07 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	42	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	44	0.00
3	Naphthalene	0.992	0.986	0.6	45	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.610	-4.1	46	0.00
5	2-Methylnaphthalene	0.679	0.697	-2.7	46	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	46	0.00
7	Acenaphthylene	1.717	1.803	-5.0	49	0.00
8 C	Acenaphthene	1.385	1.380	0.4	46	0.00
9	Fluorene	1.624	1.576	3.0	45	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	42	0.00
11	Pentachlorophenol	0.059	0.064	-8.5	49	0.00
12	Phenanthrene	1.101	1.084	1.5	43	0.00
13	Anthracene	0.958	1.036	-8.1	47	0.00
14 SURR	Fluoranthene-d10	1.088	1.116	-2.6	44	0.00
15 C	Fluoranthene	1.350	1.356	-0.4	43	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	41	0.00
17	Pyrene	1.180	1.289	-9.2	42	0.00
18	Benzo(a)anthracene	1.067	1.142	-7.0	44	0.00
19	Chrysene	1.107	1.094	1.2	41	0.00
20 I	Perylene-d12	1.000	1.000	0.0	41	0.00
21	Benzo(b)fluoranthene	1.214	1.218	-0.3	43	0.00
22	Benzo(k)fluoranthene	1.289	1.240	3.8	41	0.00
23 C	Benzo(a)pyrene	1.148	1.168	-1.7	44	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.348	-1.4	44	0.00
25	Dibenzo(a,h)anthracene	1.119	1.111	0.7	42	0.00
26	Benzo(g,h,i)perylene	1.140	1.092	4.2	41	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0