

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
 Data File : BE093366.D
 Acq On : 24 Jun 2017 17:45
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.425

Quant Time: Jun 26 01:15:37 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 26 01:13:30 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	40	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	42	0.00
3	Naphthalene	0.992	0.986	0.6	43	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.607	-3.6	44	0.00
5	2-Methylnaphthalene	0.679	0.706	-4.0	44	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	45	0.00
7	Acenaphthylene	1.717	1.841	-7.2	49	0.00
8 C	Acenaphthene	1.385	1.364	1.5	45	0.00
9	Fluorene	1.624	1.571	3.3	44	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	42	0.00
11	Pentachlorophenol	0.059	0.070	-18.6	53	0.00
12	Phenanthrene	1.101	1.082	1.7	42	0.00
13	Anthracene	0.958	1.076	-12.3	48	0.00
14 SURR	Fluoranthene-d10	1.088	1.118	-2.8	44	0.00
15 C	Fluoranthene	1.350	1.354	-0.3	42	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	39	0.00
17	Pyrene	1.180	1.299	-10.1	41	0.00
18	Benzo(a)anthracene	1.067	1.154	-8.2	42	0.00
19	Chrysene	1.107	1.077	2.7	38	0.00
20 I	Perylene-d12	1.000	1.000	0.0	39	0.00
21	Benzo(b)fluoranthene	1.214	1.252	-3.1	42	0.00
22	Benzo(k)fluoranthene	1.289	1.215	5.7	38	0.00
23 C	Benzo(a)pyrene	1.148	1.183	-3.0	42	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.348	-1.4	41	0.00
25	Dibenzo(a,h)anthracene	1.119	1.126	-0.6	41	0.00
26	Benzo(g,h,i)perylene	1.140	1.113	2.4	40	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0