

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
 Data File : BE093371.D
 Acq On : 24 Jun 2017 20:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.426

Quant Time: Jun 26 01:58:47 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 26 01:58:03 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	40	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	43	0.00
3	Naphthalene	0.992	0.983	0.9	43	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.603	-2.9	44	0.00
5	2-Methylnaphthalene	0.679	0.703	-3.5	45	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	46	0.00
7	Acenaphthylene	1.717	1.805	-5.1	48	0.00
8 C	Acenaphthene	1.385	1.366	1.4	45	0.00
9	Fluorene	1.624	1.574	3.1	45	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	41	0.00
11	Pentachlorophenol	0.059	0.066	-11.9	50	0.00
12	Phenanthrene	1.101	1.091	0.9	42	0.00
13	Anthracene	0.958	1.061	-10.8	47	0.00
14 SURR	Fluoranthene-d10	1.088	1.123	-3.2	44	0.00
15 C	Fluoranthene	1.350	1.359	-0.7	42	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	39	0.00
17	Pyrene	1.180	1.312	-11.2	41	0.00
18	Benzo(a)anthracene	1.067	1.149	-7.7	42	0.00
19	Chrysene	1.107	1.080	2.4	38	0.00
20 I	Perylene-d12	1.000	1.000	0.0	39	0.00
21	Benzo(b)fluoranthene	1.214	1.245	-2.6	41	0.00
22	Benzo(k)fluoranthene	1.289	1.250	3.0	39	0.00
23 C	Benzo(a)pyrene	1.148	1.188	-3.5	42	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.356	-2.0	41	0.00
25	Dibenzo(a,h)anthracene	1.119	1.129	-0.9	40	0.00
26	Benzo(g,h,i)perylene	1.140	1.104	3.2	39	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0