

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
 Data File : BE093371.D
 Acq On : 24 Jun 2017 20:50
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTD0.426

Quant Time: Jun 26 01:58:47 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jun 26 01:58:03 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.73	152	1548	0.40	ng/ul	0.00
2) Naphthalene-d8	10.51	136	7863	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.36	164	4592	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.10	188	11754	0.40	ng/ul	0.00
16) Chrysene-d12	21.26	240	13286	0.40	ng/ul	0.00
20) Perylene-d12	23.66	264	11160	0.40	ng/ul	0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.11	152	4743	0.41	ng/ul	0.00
14) Fluoranthene-d10	19.12	212	13198	0.41	ng/ul	0.00

Target Compounds				Qvalue		
3) Naphthalene	10.56	128	7733	0.396	ng/ul#	92
5) 2-Methylnaphthalene	12.18	142	5524	0.414	ng/ul	99
7) Acenaphthylene	14.08	152	8287	0.421	ng/ul	97
8) Acenaphthene	14.42	153	6271	0.394	ng/ul	99
9) Fluorene	15.41	166	7226	0.388	ng/ul#	92
11) Pentachlorophenol	16.75	266	779	0.447	ng/ul	98
12) Phenanthrene	17.14	178	12821	0.396	ng/ul#	91
13) Anthracene	17.22	178	12466	0.443	ng/ul#	91
15) Fluoranthene	19.15	202	15972	0.403	ng/ul	84
17) Pyrene	19.51	202	17429	0.445	ng/ul#	89
18) Benzo(a)anthracene	21.24	228	15267	0.431	ng/ul#	86
19) Chrysene	21.29	228	14352	0.390	ng/ul	97
21) Benzo(b)fluoranthene	22.92	252	13896	0.410	ng/ul	89
22) Benzo(k)fluoranthene	22.97	252	13949	0.388	ng/ul#	90
23) Benzo(a)pyrene	23.55	252	13259	0.414	ng/ul#	83
24) Indeno(1,2,3-cd)pyrene	26.18	276	15136	0.408	ng/ul#	94
25) Dibenzo(a,h)anthracene	26.20	278	12605	0.404	ng/ul#	87
26) Benzo(g,h,i)perylene	26.96	276	12317	0.387	ng/ul	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE062417\
Data File : BE093371.D
Acq On : 24 Jun 2017 20:50
Operator : SJ/JU
Sample : SSTDCCC0.4EC
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTD0.426

Quant Time: Jun 26 01:58:47 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jun 26 01:58:03 2017
Response via : Initial Calibration

