

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089967.D
 Acq On : 27 Jun 2015 9:21
 Operator : TP/IZ
 Sample : G2782-27
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0700204-150624

Quant Time: Jun 29 01:18:15 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	25444	5.00	ng	0.00
6) Naphthalene-d8	10.39	136	125097	5.00	ng	0.00
12) Acenaphthene-d10	14.24	164	75777	5.00	ng	0.00
18) Phenanthrene-d10	16.98	188	179073	5.00	ng	0.00
25) Chrysene-d12	21.13	240	202309	5.00	ng	0.00
32) Perylene-d12	23.46	264	182832	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.26	112	66675	14.51	ng	0.00
5) Phenol-d6	6.81	99	108776	15.45	ng	0.00
7) Nitrobenzene-d5	8.77	82	72274	7.76	ng	0.00
13) 2,4,6-Tribromophenol	15.74	330	19373	14.24	ng	0.00
14) 2-Fluorobiphenyl	12.87	172	169458	7.67	ng	0.00
27) Terphenyl-d14	19.60	244	259485	7.43	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
21) Pentachlorophenol	16.63	266	595	0.76	ng	93
22) Phenanthrene	17.01	178	1827	0.04	ng	# 87
24) Fluoranthene	19.02	202	3904	0.08	ng	# 96
26) Pyrene	19.38	202	3926	0.08	ng	96
28) Benzo(a)anthracene	21.12	228	3379	0.07	ng	96
29) Chrysene	21.17	228	2897	0.05	ng	93
30) Bis(2-ethylhexyl)phthalate	21.07	149	5432	1.32	ng	100
31) Indeno(1,2,3-cd)pyrene	25.88	276	1806	0.05	ng	# 100
33) Benzo(b)fluoranthene	22.75	252	3391	0.09	ng	# 82
34) Benzo(k)fluoranthene	22.80	252	1487	0.03	ng	# 62
35) Benzo(a)pyrene	23.36	252	2193	0.06	ng	# 80
36) Dibenzo(a,h)anthracene	25.90	278	773	0.02	ng	# 50
37) Benzo(g,h,i)perylene	26.62	276	1797	0.05	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
Data File : BE089967.D
Acq On : 27 Jun 2015 9:21
Operator : TP/IZ
Sample : G2782-27
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0700204-150624

Quant Time: Jun 29 01:18:15 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 25 16:51:27 2015
Response via : Initial Calibration

