

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089976.D
 Acq On : 27 Jun 2015 14:43
 Operator : TP/IZ
 Sample : G2782-26 2X
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0620204-150624

Manual Integrations
 APPROVED

apatel
 6/29/2015 5:04:56 PM

Quant Time: Jun 29 01:46:27 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	23404	5.00	ng	0.00
6) Naphthalene-d8	10.39	136	117691	5.00	ng	0.00
12) Acenaphthene-d10	14.24	164	71390	5.00	ng	0.00
18) Phenanthrene-d10	16.98	188	164163	5.00	ng	0.00
25) Chrysene-d12	21.14	240	181710	5.00	ng	0.00
32) Perylene-d12	23.47	264	169503	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.26	112	26015	6.16	ng	0.00
5) Phenol-d6	6.81	99	44941	6.94	ng	0.00
7) Nitrobenzene-d5	8.77	82	28038	3.20	ng	0.00
13) 2,4,6-Tribromophenol	15.74	330	8387	6.71	ng	0.00
14) 2-Fluorobiphenyl	12.86	172	66807	3.21	ng	0.00
27) Terphenyl-d14	19.60	244	95587	3.05	ng	0.00
Target Compounds						Qvalue
15) Acenaphthylene	13.95	152	10210	0.08	ng	99
16) Acenaphthene	14.30	154	498	0.03	ng	92
22) Phenanthrene	17.01	178	14646	0.36	ng	99
23) Anthracene	17.12	178	4411m	0.12	ng	
24) Fluoranthene	19.02	202	61430	1.37	ng	100
26) Pyrene	19.38	202	60423	1.31	ng	96
28) Benzo(a)anthracene	21.12	228	44117	1.02	ng	98
29) Chrysene	21.17	228	49040	1.03	ng	97
30) Bis(2-ethylhexyl)phthalate	21.07	149	4323	1.17	ng	94
31) Indeno(1,2,3-cd)pyrene	25.88	276	31543	0.90	ng	# 100
33) Benzo(b)fluoranthene	22.76	252	67897	1.89	ng	96
34) Benzo(k)fluoranthene	22.80	252	23310m	0.57	ng	
35) Benzo(a)pyrene	23.36	252	43377	1.36	ng	98
36) Dibenzo(a,h)anthracene	25.90	278	7477	0.25	ng	# 89
37) Benzo(g,h,i)perylene	26.62	276	34680	1.09	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
Data File : BE089976.D
Acq On : 27 Jun 2015 14:43
Operator : TP/IZ
Sample : G2782-26 2X
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0620204-150624

Manual Integrations
APPROVED

apatel
6/29/2015 5:04:56 PM

Quant Time: Jun 29 01:46:27 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 25 16:51:27 2015
Response via : Initial Calibration

