

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089942.D
 Acq On : 26 Jun 2015 13:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 29 04:05:41 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.675	0.598	11.4	106	0.00
3	n-Nitrosodimethylamine	0.899	0.781	13.1	97	0.00
4 S	2-Fluorophenol	0.903	0.783	13.3	94	0.00
5 S	Phenol-d6	1.383	1.192	13.8	95	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	103	0.00
7 S	Nitrobenzene-d5	0.372	0.333	10.5	103	0.00
8	Nitrobenzene	0.396	0.348	12.1	101	0.00
9	Naphthalene	1.112	1.030	7.4	103	0.00
10	Hexachlorobutadiene	0.176	0.166	5.7	104	0.00
11	2-Methylnaphthalene	0.761	0.685	10.0	102	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	0.064	0.050	21.9	94	0.00
14 S	2-Fluorobiphenyl	1.457	1.316	9.7	101	0.00
15	Acenaphthylene	8.767	8.114	7.4	100	0.00
16	Acenaphthene	1.288	1.209	6.1	100	0.00
17	Fluorene	1.750	1.608	8.1	100	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	98	0.00
19	4-Bromophenyl-phenylether	0.204	0.177	13.2	98	-0.02
20	Hexachlorobenzene	0.894	0.820	8.3	100	0.00
21	Pentachlorophenol	0.033	0.028	15.2	108	-0.02
22	Phenanthrene	1.254	1.133	9.6	99	-0.02
23	Anthracene	1.164	1.038	10.8	97	0.00
24	Fluoranthene	1.361	1.232	9.5	98	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	99	0.00
26	Pyrene	1.266	1.177	7.0	97	0.00
27 S	Terphenyl-d14	0.863	0.765	11.4	97	0.00
28	Benzo(a)anthracene	1.195	1.101	7.9	100	0.00
29	Chrysene	1.306	1.164	10.9	98	0.00
30	Bis(2-ethylhexyl)phthalate	0.102	0.086	15.7	100	0.00
31	Indeno(1,2,3-cd)pyrene	0.960	0.932	2.9	102	0.00
32 I	Perylene-d12	1.000	1.000	0.0	100	0.00
33	Benzo(b)fluoranthene	1.058	1.022	3.4	102	0.00
34	Benzo(k)fluoranthene	1.199	1.148	4.3	102	0.00
35 C	Benzo(a)pyrene	0.940	0.875	6.9	100	0.00
36	Dibenzo(a,h)anthracene	0.875	0.832	4.9	102	0.00
37	Benzo(g,h,i)perylene	0.941	0.888	5.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0