

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089942.D
 Acq On : 26 Jun 2015 13:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 29 04:05:41 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	0.985	1.5	106	0.00
3	n-Nitrosodimethylamine	1.000	0.869	13.1	97	0.00
4 S	2-Fluorophenol	1.000	0.868	13.2	94	0.00
5 S	Phenol-d6	1.000	0.862	13.8	95	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	103	0.00
7 S	Nitrobenzene-d5	1.000	0.893	10.7	103	0.00
8	Nitrobenzene	1.000	0.879	12.1	101	0.00
9	Naphthalene	1.000	0.926	7.4	103	0.00
10	Hexachlorobutadiene	1.000	0.943	5.7	104	0.00
11	2-Methylnaphthalene	1.000	0.901	9.9	102	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
13 S	2,4,6-Tribromophenol	1.000	0.860	14.0	94	0.00
14 S	2-Fluorobiphenyl	1.000	0.903	9.7	101	0.00
15	Acenaphthylene	1.000	0.925	7.5	100	0.00
16	Acenaphthene	1.000	0.939	6.1	100	0.00
17	Fluorene	1.000	0.992	0.8	100	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	98	0.00
19	4-Bromophenyl-phenylether	1.000	0.866	13.4	98	-0.02
20	Hexachlorobenzene	1.000	0.917	8.3	100	0.00
21	Pentachlorophenol	1.000	0.979	2.1	108	-0.02
22	Phenanthrene	1.000	0.903	9.7	99	-0.02
23	Anthracene	1.000	0.892	10.8	97	0.00
24	Fluoranthene	1.000	0.905	9.5	98	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	99	0.00
26	Pyrene	1.000	0.930	7.0	97	0.00
27 S	Terphenyl-d14	1.000	0.886	11.4	97	0.00
28	Benzo(a)anthracene	1.000	0.922	7.8	100	0.00
29	Chrysene	1.000	0.891	10.9	98	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.846	15.4	100	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.971	2.9	102	0.00
32 I	Perylene-d12	5.000	5.000	0.0	100	0.00
33	Benzo(b)fluoranthene	1.000	0.966	3.4	102	0.00
34	Benzo(k)fluoranthene	1.000	0.957	4.3	102	0.00
35 C	Benzo(a)pyrene	1.000	0.930	7.0	100	0.00
36	Dibenzo(a,h)anthracene	1.000	0.951	4.9	102	0.00
37	Benzo(g,h,i)perylene	1.000	0.944	5.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0