

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089953.D
 Acq On : 26 Jun 2015 21:25
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 27 01:01:53 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	86	0.00
2	1,4-Dioxane	1.000	1.019	-1.9	93	0.00
3	n-Nitrosodimethylamine	1.000	0.908	9.2	86	0.00
4 S	2-Fluorophenol	1.000	1.070	-7.0	98	0.00
5 S	Phenol-d6	1.000	1.604	-60.4#	149	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	88	0.00
7 S	Nitrobenzene-d5	1.000	1.065	-6.5	105	0.00
8	Nitrobenzene	1.000	1.052	-5.2	104	0.00
9	Naphthalene	1.000	1.097	-9.7	104	0.00
10	Hexachlorobutadiene	1.000	0.995	0.5	94	0.00
11	2-Methylnaphthalene	1.000	1.239	-23.9	120	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	1.000	1.567	-56.7#	189	0.00
14 S	2-Fluorobiphenyl	1.000	1.297	-29.7#	127	0.00
15	Acenaphthylene	1.000	1.285	-28.5#	121	0.00
16	Acenaphthene	1.000	1.313	-31.3#	122	0.00
17	Fluorene	1.000	1.408	-40.8#	124	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
19	4-Bromophenyl-phenylether	1.000	1.254	-25.4#	125	-0.02
20	Hexachlorobenzene	1.000	1.273	-27.3#	123	0.00
21	Pentachlorophenol	1.000	1.628	-62.8#	211	-0.02
22	Phenanthrene	1.000	1.259	-25.9#	123	-0.02
23	Anthracene	1.000	1.286	-28.6#	124	0.00
24	Fluoranthene	1.000	1.308	-30.8#	126	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	95	0.00
26	Pyrene	1.000	1.267	-26.7#	126	0.00
27 S	Terphenyl-d14	1.000	1.262	-26.2#	131	0.00
28	Benzo(a)anthracene	1.000	1.334	-33.4#	138	0.00
29	Chrysene	1.000	1.283	-28.3#	135	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.535	-53.5#	172	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.437	-43.7#	144	0.00
32 I	Perylene-d12	5.000	5.000	0.0	95	0.00
33	Benzo(b)fluoranthene	1.000	1.388	-38.8#	139	0.00
34	Benzo(k)fluoranthene	1.000	1.398	-39.8#	142	0.00
35 C	Benzo(a)pyrene	1.000	1.405	-40.5#	144	0.00
36	Dibenzo(a,h)anthracene	1.000	1.441	-44.1#	148	0.00
37	Benzo(g,h,i)perylene	1.000	1.413	-41.3#	144	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1