

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089962.D
 Acq On : 27 Jun 2015 4:34
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 27 06:30:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.675	0.638	5.5	95	0.00
3	n-Nitrosodimethylamine	0.899	0.744	17.2	78	0.00
4 S	2-Fluorophenol	0.903	1.230	-36.2#	124	0.00
5 S	Phenol-d6	1.383	3.187	-130.4#	213#	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	93	0.00
7 S	Nitrobenzene-d5	0.372	0.519	-39.5#	145	0.00
8	Nitrobenzene	0.396	0.300	24.2	79	0.00
9	Naphthalene	1.112	0.923	17.0	83	0.00
10	Hexachlorobutadiene	0.176	0.150	14.8	85	0.00
11	2-Methylnaphthalene	0.761	0.600	21.2	80	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
13 S	2,4,6-Tribromophenol	0.064	0.129	-101.6#	221#	0.00
14 S	2-Fluorobiphenyl	1.457	2.616	-79.5#	184#	0.00
15	Acenaphthylene	8.767	6.878	21.5	77	0.00
16	Acenaphthene	1.288	1.041	19.2	78	0.00
17	Fluorene	1.750	1.344	23.2	76	0.00
18 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	0.204	0.154	24.5	75	-0.02
20	Hexachlorobenzene	0.894	0.734	17.9	79	0.00
21	Pentachlorophenol	0.033	0.021	36.4#	72	-0.02
22	Phenanthrene	1.254	0.980	21.9	75	-0.02
23	Anthracene	1.164	0.883	24.1	72	0.00
24	Fluoranthene	1.361	1.007	26.0#	70	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	81	0.00
26	Pyrene	1.266	1.041	17.8	70	0.00
27 S	Terphenyl-d14	0.863	1.671	-93.6#	173#	0.00
28	Benzo(a)anthracene	1.195	0.911	23.8	68	0.00
29	Chrysene	1.306	0.990	24.2	68	0.00
30	Bis(2-ethylhexyl)phthalate	0.102	0.077	24.5	73	0.00
31	Indeno(1,2,3-cd)pyrene	0.960	0.715	25.5#	64	0.00
32 I	Perylene-d12	1.000	1.000	0.0	77	0.00
33	Benzo(b)fluoranthene	1.058	0.850	19.7	65	0.00
34	Benzo(k)fluoranthene	1.199	0.968	19.3	66	0.00
35 C	Benzo(a)pyrene	0.940	0.718	23.6#	63	0.00
36	Dibenzo(a,h)anthracene	0.875	0.682	22.1	65	0.00
37	Benzo(g,h,i)perylene	0.941	0.729	22.5	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1