

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089962.D
 Acq On : 27 Jun 2015 4:34
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 27 06:30:22 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	86	0.00
2	1,4-Dioxane	1.000	1.054	-5.4	95	0.00
3	n-Nitrosodimethylamine	1.000	0.828	17.2	78	0.00
4 S	2-Fluorophenol	1.000	1.362	-36.2#	124	0.00
5 S	Phenol-d6	1.000	2.304	-130.4#	213	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	93	0.00
7 S	Nitrobenzene-d5	1.000	1.395	-39.5#	145	0.00
8	Nitrobenzene	1.000	0.758	24.2	79	0.00
9	Naphthalene	1.000	0.830	17.0	83	0.00
10	Hexachlorobutadiene	1.000	0.856	14.4	85	0.00
11	2-Methylnaphthalene	1.000	0.788	21.2	80	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	92	0.00
13 S	2,4,6-Tribromophenol	1.000	1.721	-72.1#	221	0.00
14 S	2-Fluorobiphenyl	1.000	1.795	-79.5#	184	0.00
15	Acenaphthylene	1.000	0.785	21.5	77	0.00
16	Acenaphthene	1.000	0.808	19.2	78	0.00
17	Fluorene	1.000	0.829	17.1	76	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	1.000	0.754	24.6	75	-0.02
20	Hexachlorobenzene	1.000	0.821	17.9	79	0.00
21	Pentachlorophenol	1.000	0.849	15.1	72	-0.02
22	Phenanthrene	1.000	0.781	21.9	75	-0.02
23	Anthracene	1.000	0.758	24.2	72	0.00
24	Fluoranthene	1.000	0.740	26.0#	70	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	81	0.00
26	Pyrene	1.000	0.823	17.7	70	0.00
27 S	Terphenyl-d14	1.000	1.935	-93.5#	173	0.00
28	Benzo(a)anthracene	1.000	0.763	23.7	68	0.00
29	Chrysene	1.000	0.758	24.2	68	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.757	24.3	73	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.745	25.5#	64	0.00
32 I	Perylene-d12	5.000	5.000	0.0	77	0.00
33	Benzo(b)fluoranthene	1.000	0.803	19.7	65	0.00
34	Benzo(k)fluoranthene	1.000	0.807	19.3	66	0.00
35 C	Benzo(a)pyrene	1.000	0.764	23.6#	63	0.00
36	Dibenzo(a,h)anthracene	1.000	0.780	22.0	65	0.00
37	Benzo(g,h,i)perylene	1.000	0.775	22.5	64	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1