

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062615\
 Data File : BE089964.D
 Acq On : 27 Jun 2015 7:33
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 29 01:12:37 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 25 16:51:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	81	0.00
2	1,4-Dioxane	1.000	1.030	-3.0	88	0.00
3	n-Nitrosodimethylamine	1.000	0.850	15.0	75	0.00
4 S	2-Fluorophenol	1.000	0.796	20.4	69	0.00
5 S	Phenol-d6	1.000	0.908	9.2	80	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	81	0.00
7 S	Nitrobenzene-d5	1.000	0.813	18.7	74	0.00
8	Nitrobenzene	1.000	0.820	18.0	75	0.00
9	Naphthalene	1.000	0.931	6.9	82	0.00
10	Hexachlorobutadiene	1.000	0.930	7.0	81	0.00
11	2-Methylnaphthalene	1.000	0.902	9.8	80	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	81	0.00
13 S	2,4,6-Tribromophenol	1.000	0.852	14.8	75	0.00
14 S	2-Fluorobiphenyl	1.000	0.898	10.2	81	0.00
15	Acenaphthylene	1.000	0.913	8.7	79	0.00
16	Acenaphthene	1.000	0.946	5.4	81	0.00
17	Fluorene	1.000	0.985	1.5	80	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	80	0.00
19	4-Bromophenyl-phenylether	1.000	0.858	14.2	78	-0.02
20	Hexachlorobenzene	1.000	0.917	8.3	81	0.00
21	Pentachlorophenol	1.000	2.478	-147.8#	330	-0.02
22	Phenanthrene	1.000	0.912	8.8	81	-0.02
23	Anthracene	1.000	0.869	13.1	76	0.00
24	Fluoranthene	1.000	0.881	11.9	77	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	79	0.00
26	Pyrene	1.000	0.938	6.2	78	0.00
27 S	Terphenyl-d14	1.000	0.894	10.6	78	0.00
28	Benzo(a)anthracene	1.000	0.901	9.9	78	0.00
29	Chrysene	1.000	0.871	12.9	77	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.924	7.6	87	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.931	6.9	78	0.00
32 I	Perylene-d12	5.000	5.000	0.0	76	0.00
33	Benzo(b)fluoranthene	1.000	0.957	4.3	77	0.00
34	Benzo(k)fluoranthene	1.000	0.955	4.5	77	0.00
35 C	Benzo(a)pyrene	1.000	0.911	8.9	75	0.00
36	Dibenzo(a,h)anthracene	1.000	0.962	3.8	79	0.00
37	Benzo(g,h,i)perylene	1.000	0.932	6.8	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0