

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100198.D
 Acq On : 27 Jun 2019 10:08
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 28 14:33:08 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 25 15:40:27 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	154#	0.00
2	1,4-Dioxane	0.626	0.595	5.0	134	0.00
3	n-Nitrosodimethylamine	0.232	0.226	2.6	179#	0.00
4 S	2-Fluorophenol	0.889	0.931	-4.7	154#	0.00
5 S	Phenol-d6	1.223	1.250	-2.2	166#	0.00
6	bis(2-Chloroethyl)ether	1.120	1.137	-1.5	170#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	149	0.00
8 S	Nitrobenzene-d5	0.398	0.394	1.0	147	0.00
9	Naphthalene	4.365	4.595	-5.3	148	0.00
10	Hexachlorobutadiene	0.451	0.460	-2.0	141	0.00
11 SURR	2-Methylnaphthalene-d10	0.750	0.771	-2.8	146	0.00
12	2-Methylnaphthalene	0.703	0.717	-2.0	142	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	154#	0.00
14 S	2,4,6-Tribromophenol	0.232	0.177	23.7	122	0.00
15 S	2-Fluorobiphenyl	1.604	1.585	1.2	146	0.00
16	Acenaphthylene	1.950	1.956	-0.3	145	0.00
17	Acenaphthene	1.093	1.112	-1.7	145	0.00
18	Fluorene	1.581	1.651	-4.4	148	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	155#	0.00
20	4-Bromophenyl-phenylether	0.248	0.258	-4.0	154#	0.00
21	Hexachlorobenzene	0.272	0.273	-0.4	146	0.00
22	Pentachlorophenol	0.064	0.021	67.2#	64	0.20
23	Phenanthrene	0.947	0.995	-5.1	152#	0.00
24	Anthracene	0.800	0.830	-3.7	158#	0.00
25 SURR	Fluoranthene-d10	5.765	5.620	2.5	146	0.00
26	Fluoranthene	1.541	1.519	1.4	148	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	144	0.00
28	Pvrene	1.024	1.179	-15.1	148	0.00
29 S	Terphenyl-d14	0.785	0.880	-12.1	145	0.00
30	Benzo(a)anthracene	1.179	1.251	-6.1	144	0.00
31	Chrysene	1.383	1.428	-3.3	143	0.00
32	Bis(2-ethylhexyl)phthalate	1.355	1.332	1.7	151#	0.00
33	Indeno(1,2,3-cd)pyrene	1.969	1.556	21.0	123	0.00
34 I	Perylene-d12	1.000	1.000	0.0	131	0.00
35	Benzo(b)fluoranthene	1.076	1.131	-5.1	131	0.00
36	Benzo(k)fluoranthene	1.287	1.309	-1.7	123	0.00
37 C	Benzo(a)pyrene	1.115	1.151	-3.2	129	0.00
38	Dibenzo(a,h)anthracene	1.192	1.062	10.9	121	0.00
39	Benzo(a,h,i)perylene	1.245	1.162	6.7	122	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100198.D
Acq On : 27 Jun 2019 10:08
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 28 14:33:08 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 25 15:40:27 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			