

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100206.D
 Acq On : 27 Jun 2019 16:03
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 27 19:17:25 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 19:12:41 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	156#	0.00
2	1,4-Dioxane	0.626	0.608	2.9	140	0.00
3	n-Nitrosodimethylamine	0.232	0.253	-9.1	203#	-0.01
4 S	2-Fluorophenol	0.889	1.079	-21.4	182#	0.00
5 S	Phenol-d6	1.223	1.314	-7.4	177#	-0.01
6	bis(2-Chloroethyl)ether	1.120	1.187	-6.0	180#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	158#	0.00
8 S	Nitrobenzene-d5	0.398	0.420	-5.5	167#	-0.01
9	Naphthalene	4.365	4.580	-4.9	157#	0.00
10	Hexachlorobutadiene	0.451	0.458	-1.6	150	0.00
11 SURR	2-Methylnaphthalene-d10	0.750	0.773	-3.1	155#	-0.01
12	2-Methylnaphthalene	0.703	0.744	-5.8	156#	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	171#	0.00
14 S	2,4,6-Tribromophenol	0.232	0.223	3.9	171#	-0.01
15 S	2-Fluorobiphenyl	1.604	1.541	3.9	159#	-0.01
16	Acenaphthylene	1.950	1.891	3.0	156#	0.00
17	Acenaphthene	1.093	1.112	-1.7	162#	0.00
18	Fluorene	1.581	1.637	-3.5	164#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	165#	0.00
20	4-Bromophenyl-phenylether	0.248	0.249	-0.4	158#	0.00
21	Hexachlorobenzene	0.272	0.281	-3.3	159#	0.00
22	Pentachlorophenol	0.064	0.060	6.3	193#	0.01
23	Phenanthrene	0.947	1.003	-5.9	163#	0.00
24	Anthracene	0.800	0.873	-9.1	177#	0.00
25 SURR	Fluoranthene-d10	5.765	5.638	2.2	155#	0.00
26	Fluoranthene	1.541	1.526	1.0	158#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	154#	0.00
28	Pvrene	1.024	1.190	-16.2	161#	0.00
29 S	Terphenyl-d14	0.785	0.872	-11.1	155#	0.00
30	Benzo(a)anthracene	1.179	1.310	-11.1	162#	0.00
31	Chrysene	1.383	1.390	-0.5	149	0.00
32	Bis(2-ethylhexyl)phthalate	1.355	2.034	-50.1#	247#	0.00
33	Indeno(1,2,3-cd)pyrene	1.969	1.553	21.1	132	0.00
34 I	Perylene-d12	1.000	1.000	0.0	140	0.00
35	Benzo(b)fluoranthene	1.076	1.142	-6.1	141	0.00
36	Benzo(k)fluoranthene	1.287	1.261	2.0	126	0.00
37 C	Benzo(a)pyrene	1.115	1.127	-1.1	135	0.00
38	Dibenzo(a,h)anthracene	1.192	1.073	10.0	130	0.00
39	Benzo(a,h,i)perylene	1.245	1.204	3.3	135	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100206.D
Acq On : 27 Jun 2019 16:03
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 27 19:17:25 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 19:12:41 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			