

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100208.D
 Acq On : 27 Jun 2019 17:14
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 27 19:19:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 19:14:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	150	0.00
2	1,4-Dioxane	0.400	0.412	-3.0	142	0.00
3	n-Nitrosodimethylamine	0.400	0.409	-2.2	183	-0.01
4 S	2-Fluorophenol	0.400	0.414	-3.5	149	0.00
5 S	Phenol-d6	0.400	0.397	0.8	157	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.422	-5.5	172	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	147	0.00
8 S	Nitrobenzene-d5	0.400	0.424	-6.0	156	-0.01
9	Naphthalene	0.400	0.418	-4.5	145	0.00
10	Hexachlorobutadiene	0.400	0.413	-3.2	141	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.402	-0.5	141	-0.01
12	2-Methylnaphthalene	0.400	0.423	-5.7	145	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	157	0.00
14 S	2,4,6-Tribromophenol	0.400	0.348	13.0	142	-0.01
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	147	0.00
16	Acenaphthylene	0.400	0.389	2.8	143	0.00
17	Acenaphthene	0.400	0.399	0.3	145	0.00
18	Fluorene	0.400	0.420	-5.0	152	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	150	0.00
20	4-Bromophenyl-phenylether	0.400	0.408	-2.0	146	0.00
21	Hexachlorobenzene	0.400	0.413	-3.2	145	0.00
22	Pentachlorophenol	0.400	0.464	-16.0	131	0.03
23	Phenanthrene	0.400	0.420	-5.0	148	0.00
24	Anthracene	0.400	0.426	-6.5	158	0.00
25 SURR	Fluoranthene-d10	0.400	0.385	3.8	139	0.00
26	Fluoranthene	0.400	0.393	1.8	143	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	137	0.00
28	Pvrene	0.400	0.464	-16.0	142	0.00
29 S	Terphenyl-d14	0.400	0.443	-10.7	138	0.00
30	Benzo(a)anthracene	0.400	0.422	-5.5	136	0.00
31	Chrysene	0.400	0.409	-2.2	135	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.372	7.0	136	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.320	20.0	119	0.00
34 I	Perylene-d12	0.400	0.400	0.0	128	0.00
35	Benzo(b)fluoranthene	0.400	0.420	-5.0	128	0.00
36	Benzo(k)fluoranthene	0.400	0.409	-2.2	121	0.00
37 C	Benzo(a)pyrene	0.400	0.397	0.8	121	0.00
38	Dibenzo(a,h)anthracene	0.400	0.352	12.0	117	-0.01
39	Benzo(a,h,i)perylene	0.400	0.384	4.0	122	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100208.D
Acq On : 27 Jun 2019 17:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 27 19:19:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 19:14:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				