

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100208.D
 Acq On : 27 Jun 2019 17:14
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 27 19:19:23 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jun 27 19:14:38 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	150	0.00
2	1,4-Dioxane	0.626	0.645	-3.0	142	0.00
3	n-Nitrosodimethylamine	0.232	0.237	-2.2	183#	-0.01
4 S	2-Fluorophenol	0.889	0.920	-3.5	149	0.00
5 S	Phenol-d6	1.223	1.213	0.8	157#	-0.01
6	bis(2-Chloroethyl)ether	1.120	1.182	-5.5	172#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	147	0.00
8 S	Nitrobenzene-d5	0.398	0.422	-6.0	156#	-0.01
9	Naphthalene	4.365	4.559	-4.4	145	0.00
10	Hexachlorobutadiene	0.451	0.466	-3.3	141	0.00
11 SURR	2-Methylnaphthalene-d10	0.750	0.754	-0.5	141	-0.01
12	2-Methylnaphthalene	0.703	0.744	-5.8	145	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	157#	0.00
14 S	2,4,6-Tribromophenol	0.232	0.202	12.9	142	-0.01
15 S	2-Fluorobiphenyl	1.604	1.564	2.5	147	0.00
16	Acenaphthylene	1.950	1.897	2.7	143	0.00
17	Acenaphthene	1.093	1.091	0.2	145	0.00
18	Fluorene	1.581	1.660	-5.0	152#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	150#	0.00
20	4-Bromophenyl-phenylether	0.248	0.253	-2.0	146	0.00
21	Hexachlorobenzene	0.272	0.281	-3.3	145	0.00
22	Pentachlorophenol	0.064	0.044	31.3#	131	0.03
23	Phenanthrene	0.947	0.996	-5.2	148	0.00
24	Anthracene	0.800	0.852	-6.5	158#	0.00
25 SURR	Fluoranthene-d10	5.765	5.546	3.8	139	0.00
26	Fluoranthene	1.541	1.513	1.8	143	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	137	0.00
28	Pvrene	1.024	1.187	-15.9	142	0.00
29 S	Terphenyl-d14	0.785	0.871	-11.0	138	0.00
30	Benzo(a)anthracene	1.179	1.243	-5.4	136	0.00
31	Chrysene	1.383	1.415	-2.3	135	0.00
32	Bis(2-ethylhexyl)phthalate	1.355	1.260	7.0	136	-0.01
33	Indeno(1,2,3-cd)pyrene	1.969	1.575	20.0	119	0.00
34 I	Perylene-d12	1.000	1.000	0.0	128	0.00
35	Benzo(b)fluoranthene	1.076	1.130	-5.0	128	0.00
36	Benzo(k)fluoranthene	1.287	1.317	-2.3	121	0.00
37 C	Benzo(a)pyrene	1.115	1.105	0.9	121	0.00
38	Dibenzo(a,h)anthracene	1.192	1.050	11.9	117	-0.01
39	Benzo(a,h,i)perylene	1.245	1.194	4.1	122	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100208.D
Acq On : 27 Jun 2019 17:14
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 27 19:19:23 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jun 27 19:14:38 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			