

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100224.D
 Acq On : 28 Jun 2019 2:53
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jun 28 06:43:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 05:41:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	220	0.00
2	1,4-Dioxane	0.400	0.323	19.3	164	0.00
3	n-Nitrosodimethylamine	0.400	0.452	-13.0	296	-0.02
4 S	2-Fluorophenol	0.400	0.487	-21.7	257	0.00
5 S	Phenol-d6	0.400	0.468	-17.0	272	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	221	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	236	0.00
8 S	Nitrobenzene-d5	0.400	0.390	2.5	230	-0.01
9	Naphthalene	0.400	0.406	-1.5	227	0.00
10	Hexachlorobutadiene	0.400	0.386	3.5	212	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.417	-4.2	235	-0.02
12	2-Methylnaphthalene	0.400	0.429	-7.2	236	-0.02
13 I	Acenaphthene-d10	0.400	0.400	0.0	228	0.00
14 S	2,4,6-Tribromophenol	0.400	0.468	-17.0	277	-0.01
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	220	-0.02
16	Acenaphthylene	0.400	0.419	-4.7	225	0.00
17	Acenaphthene	0.400	0.406	-1.5	215	0.00
18	Fluorene	0.400	0.404	-1.0	213	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	205	0.00
20	4-Bromophenyl-phenylether	0.400	0.426	-6.5	208	0.00
21	Hexachlorobenzene	0.400	0.408	-2.0	195	0.00
22	Pentachlorophenol	0.400	0.658	-64.5#	357	0.01
23	Phenanthrene	0.400	0.414	-3.5	198	0.00
24	Anthracene	0.400	0.453	-13.2	228	0.00
25 SURR	Fluoranthene-d10	0.400	0.386	3.5	191	0.00
26	Fluoranthene	0.400	0.381	4.8	188	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	175	0.01
28	Pvrene	0.400	0.488	-22.0	191	0.00
29 S	Terphenyl-d14	0.400	0.487	-21.7	193	0.00
30	Benzo(a)anthracene	0.400	0.461	-15.3	190	0.00
31	Chrysene	0.400	0.403	-0.8	170	0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.653	-63.2#	305	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.335	16.3	159	0.00
34 I	Perylene-d12	0.400	0.400	0.0	170	0.00
35	Benzo(b)fluoranthene	0.400	0.431	-7.7	173	0.00
36	Benzo(k)fluoranthene	0.400	0.360	10.0	140	0.00
37 C	Benzo(a)pyrene	0.400	0.396	1.0	160	0.00
38	Dibenzo(a,h)anthracene	0.400	0.381	4.8	167	0.00
39	Benzo(a,h,i)perylene	0.400	0.379	5.3	160	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100224.D
Acq On : 28 Jun 2019 2:53
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 28 06:43:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 05:41:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			