

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100224.D
 Acq On : 28 Jun 2019 2:53
 Operator : JU/SJ
 Sample : SSTDC00.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDC00.4

Quant Time: Jun 28 06:43:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 05:41:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	220#	0.00
2	1,4-Dioxane	0.626	0.506	19.2	164#	0.00
3	n-Nitrosodimethylamine	0.232	0.262	-12.9	296#	-0.02
4 S	2-Fluorophenol	0.889	1.084	-21.9	257#	0.00
5 S	Phenol-d6	1.223	1.432	-17.1	272#	-0.01
6	bis(2-Chloroethyl)ether	1.120	1.036	7.5	221#	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	236#	0.00
8 S	Nitrobenzene-d5	0.398	0.388	2.5	230#	-0.01
9	Naphthalene	4.365	4.434	-1.6	227#	0.00
10	Hexachlorobutadiene	0.451	0.435	3.5	212#	0.00
11 SURR	2-Methylnaphthalene-d10	0.750	0.782	-4.3	235#	-0.02
12	2-Methylnaphthalene	0.703	0.754	-7.3	236#	-0.02
13 I	Acenaphthene-d10	1.000	1.000	0.0	228#	0.00
14 S	2,4,6-Tribromophenol	0.232	0.271	-16.8	277#	-0.01
15 S	2-Fluorobiphenyl	1.604	1.601	0.2	220#	-0.02
16	Acenaphthylene	1.950	2.043	-4.8	225#	0.00
17	Acenaphthene	1.093	1.108	-1.4	215#	0.00
18	Fluorene	1.581	1.595	-0.9	213#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	205#	0.00
20	4-Bromophenyl-phenylether	0.248	0.264	-6.5	208#	0.00
21	Hexachlorobenzene	0.272	0.277	-1.8	195#	0.00
22	Pentachlorophenol	0.064	0.089	-39.1#	357#	0.01
23	Phenanthrene	0.947	0.981	-3.6	198#	0.00
24	Anthracene	0.800	0.907	-13.4	228#	0.00
25 SURR	Fluoranthene-d10	5.765	5.568	3.4	191#	0.00
26	Fluoranthene	1.541	1.468	4.7	188#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	175#	0.01
28	Pvrene	1.024	1.249	-22.0	191#	0.00
29 S	Terphenyl-d14	0.785	0.956	-21.8	193#	0.00
30	Benzo(a)anthracene	1.179	1.359	-15.3	190#	0.00
31	Chrysene	1.383	1.395	-0.9	170#	0.01
32	Bis(2-ethylhexyl)phthalate	1.355	2.212	-63.2#	305#	0.00
33	Indeno(1,2,3-cd)pyrene	1.969	1.648	16.3	159#	0.00
34 I	Perylene-d12	1.000	1.000	0.0	170#	0.00
35	Benzo(b)fluoranthene	1.076	1.160	-7.8	173#	0.00
36	Benzo(k)fluoranthene	1.287	1.159	9.9	140	0.00
37 C	Benzo(a)pyrene	1.115	1.102	1.2	160#	0.00
38	Dibenzo(a,h)anthracene	1.192	1.135	4.8	167#	0.00
39	Benzo(a,h,i)perylene	1.245	1.178	5.4	160#	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100224.D
Acq On : 28 Jun 2019 2:53
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jun 28 06:43:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 05:41:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			