

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
 Data File : BE100224.D
 Acq On : 28 Jun 2019 2:53
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Jun 28 06:43:18 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 05:41:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.65	152	1063	0.40	ng	0.00
7) Naphthalene-d8	10.45	136	3846	0.40	ng	0.00
13) Acenaphthene-d10	14.33	164	2823	0.40	ng	0.00
19) Phenanthrene-d10	17.07	188	7968	0.40	ng	0.00
27) Chrysene-d12	21.26	240	9516	0.40	ng	0.01
34) Perylene-d12	23.68	264	11154	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.22	112	1152	0.49	ng	0.00
5) Phenol-d6	6.84	99	1522	0.47	ng	-0.01
8) Nitrobenzene-d5	8.82	82	1491	0.39	ng	-0.01
11) 2-Methylnaphthalene-d10	12.05	152	3009	0.42	ng	-0.02
14) 2,4,6-Tribromophenol	15.82	330	765	0.47	ng	-0.01
15) 2-Fluorobiphenyl	12.94	172	4521	0.40	ng	-0.02
25) Fluoranthene-d10	19.10	212	44362	0.39	ng	0.00
29) Terphenyl-d14	19.70	244	9099	0.49	ng	0.00


Target Compounds				Qvalue		
2) 1,4-Dioxane	3.12	88	538	0.323	ng	# 87
3) n-Nitrosodimethylamine	3.47	42	278	0.452	ng	# 29
6) bis(2-Chloroethyl)ether	7.09	93	1101	0.370	ng	# 82
9) Naphthalene	10.50	128	17053	0.406	ng	# 96
10) Hexachlorobutadiene	10.77	225	1673	0.386	ng	100
12) 2-Methylnaphthalene	12.12	142	2901	0.429	ng	96
16) Acenaphthylene	14.04	152	5767	0.419	ng	99
17) Acenaphthene	14.38	154	3129	0.406	ng	98
18) Fluorene	15.37	166	4504	0.404	ng	99
20) 4-Bromophenyl-phenylether	16.26	248	2105	0.426	ng	# 88
21) Hexachlorobenzene	16.37	284	2211	0.408	ng	98
22) Pentachlorophenol	16.73	266	706	0.658	ng	85
23) Phenanthrene	17.11	178	7814	0.414	ng	99
24) Anthracene	17.20	178	7224	0.453	ng	98
26) Fluoranthene	19.13	202	11695	0.381	ng	99
28) Pyrene	19.50	202	11882	0.488	ng	99
30) Benzo(a)anthracene	21.23	228	12929	0.461	ng	97
31) Chrysene	21.29	228	13272	0.403	ng	98
32) Bis(2-ethylhexyl)phthalate	21.16	149	21053	0.653	ng	100
33) Indeno(1,2,3-cd)pyrene	26.23	276	15686	0.335	ng	# 95
35) Benzo(b)fluoranthene	22.93	252	12939	0.431	ng	98
36) Benzo(k)fluoranthene	22.98	252	12933	0.360	ng	98
37) Benzo(a)pyrene	23.57	252	12295	0.396	ng	97
38) Dibenzo(a,h)anthracene	26.26	278	12655	0.381	ng	98
39) Benzo(g,h,i)perylene	27.02	276	13142	0.379	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE062719\
Data File : BE100224.D
Acq On : 28 Jun 2019 2:53
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Jun 28 06:43:18 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062519.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 05:41:07 2019
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.65 min Scan# 416
Delta R.T. 0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:152 Resp: 1063
Ion Ratio Lower Upper
152 100
150 130.0 113.4 170.2
115 57.5 55.8 83.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#2


1,4-Dioxane
Concen: 0.323 ng
RT: 3.12 min Scan# 22
Delta R.T. 0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Tgt Ion: 88 Resp: 538
Ion Ratio Lower Upper
88 100
58 50.0 43.8 65.6
43 37.2 19.4 29.2#



Abundance Ion 88.00 (87.70 to 88.70): BE1
Ion 58.00 (57.70 to 58.70): BE1
Ion 43.00 (42.70 to 43.70): BE1





#3

n-Nitrosodimethylamine

Concen: 0.452 ng

RT: 3.47 min Scan# 53

Delta R.T. -0.02 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion: 42 Resp: 278

Ion Ratio Lower Upper




42 100

74 165.1 193.2 289.8#

44 0.0 114.5 171.7#



Abundance Ion 42.00 (41.70 to 42.70): BE1
Ion 74.00 (73.70 to 74.70): BE1
Ion 44.00 (43.70 to 44.70): BE1



#4

2-Fluorophenol

Concen: 0.487 ng

RT: 5.22 min Scan# 205

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion: 112 Resp: 1152

Ion Ratio Lower Upper



112 100




64 53.9 40.7 61.1

63 124.4 93.2 139.8



Abundance Ion 112.00 (111.70 to 112.70): BE1
Ion 64.00 (63.70 to 64.70): BE1
Ion 63.00 (62.70 to 63.70): BE1





#5

Phenol-d6

Concen: 0.468 ng

RT: 6.84 min Scan# 345

Delta R.T. -0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC





Tot Ion: 99 Resp: 1522

Ion Ratio Lower Upper

99 100

42 17.1 10.4 15.6#

71 39.8 32.4 48.6



#6

bis(2-Chloroethyl)ether

Concen: 0.370 ng

RT: 7.09 min Scan# 367

Delta R.T. -0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion: 93 Resp: 1101


Ion Ratio Lower Upper

93 100

63 248.8 0.0 0.0#

95 35.7 21.0 31.6#





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.45 min Scan# 637

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC

Tot Ion:136 Resp: 3846





Ion Ratio Lower Upper

136 100

137 12.4 13.8 20.8#

54 24.7 35.4 53.2#

68 5.9 8.8 13.2#



#8

Nitrobenzene-d5

Concen: 0.390 ng

RT: 8.82 min Scan# 512

Delta R.T. -0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53




Tgt Ion: 82 Resp: 1491


Ion Ratio Lower Upper

82 100

128 169.8 144.6 217.0

54 148.3 163.0 244.6#









#9
Naphthalene
Concen: 0.406 ng
RT: 10.50 min Scan# 641
Delta R.T. 0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tot Ion:128 Resp: 17053
Ion Ratio Lower Upper
128 100
129 3.2 3.6 5.4#
127 3.6 3.8 5.8#



#10
Hexachlorobutadiene
Concen: 0.386 ng
RT: 10.77 min Scan# 662
Delta R.T. 0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Tgt Ion:225 Resp: 1673
Ion Ratio Lower Upper
225 100
223 62.8 50.1 75.1
227 63.7 50.9 76.3





#11

2-Methylnaphthalene-d10

Concen: 0.417 ng

RT: 12.05 min Scan# 752

Delta R.T. -0.02 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :


SSTDCCC0.4EC

Tot Ion:152 Resp: 3009

Ion Ratio Lower Upper

152 100

151 19.6 16.3 24.5



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

12.05



1500

1000

500

0

Time--> 11.90 12.00 12.10 12.20



#12

2-Methylnaphthalene

Concen: 0.429 ng

RT: 12.12 min Scan# 757

Delta R.T. -0.02 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:142 Resp: 2901

Ion Ratio Lower Upper

142 100

141 93.7 71.8 107.8

115 48.2 40.4 60.6



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E

2000


1500


1000

500

0

Time--> 12.10 12.20









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.33 min Scan# 910
 Delta R.T. -0.00 min
 Lab File: BE100224.D
 Acq: 28 Jun 2019 2:53

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC

Tot Ion:164 Resp: 2823
 Ion Ratio Lower Upper
 164 100
 162 98.9 78.6 118.0
 160 48.9 46.9 70.3




Abundance Ion 164.00 (163.70 to 164.70): E
 2500 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E






#14
 2,4,6-Tribromophenol
 Concen: 0.468 ng
 RT: 15.82 min Scan# 1015
 Delta R.T. -0.01 min
 Lab File: BE100224.D
 Acq: 28 Jun 2019 2:53

Tgt Ion:330 Resp: 765
 Ion Ratio Lower Upper
 330 100
 332 100.8 77.7 116.5
 141 27.5 18.9 28.3



Abundance Ion 330.00 (329.70 to 330.70): E
 600 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.399 ng
RT: 12.94 min Scan# 814
Delta R.T. -0.02 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53




Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC


Tot Ion:172 Resp: 4521
Ion Ratio Lower Upper
172 100
171 36.4 31.0 46.4
170 27.1 23.9 35.9



#16
Acenaphthylene
Concen: 0.419 ng
RT: 14.04 min Scan# 890
Delta R.T. -0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Tgt Ion:152 Resp: 5767
Ion Ratio Lower Upper
152 100
151 19.3 15.4 23.0
153 13.1 10.8 16.2





#17

Acenaphthene

Concen: 0.406 ng

RT: 14.38 min Scan# 914

Delta R.T. -0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:154 Resp: 3129

Ion Ratio Lower Upper

154 100

153 118.9 95.3 142.9




152 62.1 52.9 79.3



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18

Fluorene

Concen: 0.404 ng

RT: 15.37 min Scan# 981

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:166 Resp: 4504

Ion Ratio Lower Upper

166 100

165 102.5 81.0 121.4



167 13.8 11.4 17.0




Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.07 min Scan# 1108

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:188 Resp: 7968

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0




80 5.9 6.4 9.6#



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE1

Ion 80.00 (79.70 to 80.70): BE1



#20

4-Bromophenyl-phenylether

Concen: 0.426 ng

RT: 16.26 min Scan# 1048

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:248 Resp: 2105

Ion Ratio Lower Upper

248 100

250 101.7 77.1 115.7



141 32.4 39.5 59.3#




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E





#21

Hexachlorobenzene

Concen: 0.408 ng

RT: 16.37 min Scan# 1056

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:284 Resp: 2211

Ion Ratio Lower Upper

284 100

142 22.9 20.1 30.1




249 32.6 26.2 39.4



Abundance Ion 283.80 (283.50 to 284.50): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E



#22

Pentachlorophenol

Concen: 0.658 ng

RT: 16.73 min Scan# 1083

Delta R.T. 0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:266 Resp: 706

Ion Ratio Lower Upper

266 100

264 72.1 67.7 101.5



268 71.4 69.8 104.6




Abundance Ion 266.00 (265.70 to 266.70): E

Ion 264.00 (263.70 to 264.70): E

Ion 268.00 (267.70 to 268.70): E





#23

Phenanthrene

Concen: 0.414 ng

RT: 17.11 min Scan# 1111

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:178 Resp: 7814

Ion Ratio Lower Upper

178 100

176 19.9 15.5 23.3




179 15.3 12.2 18.2



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#24

Anthracene

Concen: 0.453 ng

RT: 17.20 min Scan# 1118

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:178 Resp: 7224

Ion Ratio Lower Upper

178 100

176 18.9 15.7 23.5



179 14.8 12.9 19.3




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#25

Fluoranthene-d10

Concen: 0.386 ng

RT: 19.10 min Scan# 1364

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:212 Resp: 44362

Ion Ratio Lower Upper

212 100

106 1.6 1.3 1.9




104 1.0 0.7 1.1



Abundance Ion 212.00 (211.70 to 212.70): E

Ion 106.00 (105.70 to 106.70): E

Ion 104.00 (103.70 to 104.70): E



#26

Fluoranthene

Concen: 0.381 ng

RT: 19.13 min Scan# 1369

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:202 Resp: 11695

Ion Ratio Lower Upper

202 100

101 5.7 4.5 6.7



203 16.7 13.7 20.5




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E









#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.26 min Scan# 1620
Delta R.T. 0.01 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:240 Resp: 9516
Ion Ratio Lower Upper
240 100
120 3.7 4.1 6.1#
236 27.7 23.0 34.6




Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E






#28
Pyrene
Concen: 0.488 ng
RT: 19.50 min Scan# 1432
Delta R.T. 0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Tgt Ion:202 Resp: 11882
Ion Ratio Lower Upper
202 100
200 20.2 15.5 23.3
203 17.7 13.9 20.9



Abundance Ion 202.00 (201.70 to 202.70): E
Ion 200.00 (199.70 to 200.70): E
Ion 203.00 (202.70 to 203.70): E









#29
 Terphenyl-d14
 Concen: 0.487 ng
 RT: 19.70 min Scan# 1468
 Delta R.T. -0.01 min
 Lab File: BE100224.D
 Acq: 28 Jun 2019 2:53




Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4EC


Tot Ion:244 Resp: 9099
 Ion Ratio Lower Upper
 244 100
 212 32.1 30.7 46.1
 122 6.0 5.8 8.6



#30
 Benzo(a)anthracene
 Concen: 0.461 ng
 RT: 21.23 min Scan# 1618
 Delta R.T. -0.00 min
 Lab File: BE100224.D
 Acq: 28 Jun 2019 2:53

Tgt Ion:228 Resp: 12929
 Ion Ratio Lower Upper
 228 100
 226 28.9 21.5 32.3
 229 21.1 17.2 25.8





#31

Chrysene

Concen: 0.403 ng

RT: 21.29 min Scan# 1623

Delta R.T. 0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:228 Resp: 13272

Ion Ratio Lower Upper

228 100

226 30.0 23.7 35.5




229 22.0 16.5 24.7



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.653 ng

RT: 21.16 min Scan# 1612

Delta R.T. -0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:149 Resp: 21053

Ion Ratio Lower Upper

149 100

167 7.5 5.9 8.9



279 1.5 1.0 1.6




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#33

Indeno(1,2,3-cd)pyrene

Concen: 0.335 ng

RT: 26.23 min Scan# 2856

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:276 Resp: 15686

Ion Ratio Lower Upper




276 100

138 11.5 3.7 5.5#

277 24.2 18.7 28.1



Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E



#34

Perylene-d12

Concen: 0.400 ng

RT: 23.68 min Scan# 2145

Delta R.T. 0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:264 Resp: 11154

Ion Ratio Lower Upper



264 100


260 25.8 20.4 30.6

265 28.4 22.5 33.7



Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E









#35
Benzo(b)fluoranthene
Concen: 0.431 ng
RT: 22.93 min Scan# 1936
Delta R.T. 0.01 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4EC

Tot Ion:252 Resp: 12939
Ion Ratio Lower Upper
252 100
253 24.9 19.0 28.4
125 8.7 6.3 9.5




Abundance Ion 252.00 (251.70 to 252.70): E
10000 Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E






#36
Benzo(k)fluoranthene
Concen: 0.360 ng
RT: 22.98 min Scan# 1950
Delta R.T. 0.00 min
Lab File: BE100224.D
Acq: 28 Jun 2019 2:53

Tgt Ion:252 Resp: 12933
Ion Ratio Lower Upper
252 100
253 24.0 18.4 27.6
125 7.0 5.0 7.4



Abundance Ion 252.00 (251.70 to 252.70): E
10000 Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#37

Benzo(a)pyrene

Concen: 0.396 ng

RT: 23.57 min Scan# 2114

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion:252 Resp: 12295

Ion Ratio Lower Upper

252 100

253 25.2 19.2 28.8




125 12.0 8.2 12.2



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#38

Dibenzo(a,h)anthracene

Concen: 0.381 ng

RT: 26.26 min Scan# 2865

Delta R.T. 0.00 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53


Tgt Ion:278 Resp: 12655

Ion Ratio Lower Upper

278 100

139 9.6 9.0 13.4



279 26.2 21.1 31.7




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#39

Benzo(a,h,i)perylene

Concen: 0.379 ng

RT: 27.02 min Scan# 3079

Delta R.T. 0.01 min

Lab File: BE100224.D

Acq: 28 Jun 2019 2:53

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4EC


Tot Ion: 276 Resp: 13142

Ion Ratio Lower Upper

276 100

277 26.2 19.8 29.6

138 12.3 10.2 15.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

