

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062915\
 Data File : BE089995.D
 Acq On : 29 Jun 2015 22:43
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 30 07:03:48 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	91	0.00
2	1,4-Dioxane	1.000	1.013	-1.3	96	0.00
3	n-Nitrosodimethylamine	1.000	0.914	8.6	86	0.00
4 S	2-Fluorophenol	1.000	0.804	19.6	73	0.00
5 S	Phenol-d6	1.000	0.819	18.1	76	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	90	0.00
7 S	Nitrobenzene-d5	1.000	0.878	12.2	80	0.00
8	Nitrobenzene	1.000	0.888	11.2	82	0.00
9	Naphthalene	1.000	0.961	3.9	88	0.00
10	Hexachlorobutadiene	1.000	1.015	-1.5	92	0.00
11	2-Methylnaphthalene	1.000	0.952	4.8	87	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	1.000	0.625	37.5#	56	0.00
14 S	2-Fluorobiphenyl	1.000	0.981	1.9	89	0.00
15	Acenaphthylene	1.000	0.948	5.2	86	0.00
16	Acenaphthene	1.000	0.971	2.9	87	0.00
17	Fluorene	1.000	0.953	4.7	86	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
19	4-Bromophenyl-phenylether	1.000	0.975	2.5	87	0.00
20	Hexachlorobenzene	1.000	0.979	2.1	89	0.00
21	Pentachlorophenol	1.000	0.667	33.3#	51	0.00
22	Phenanthrene	1.000	0.976	2.4	88	0.00
23	Anthracene	1.000	0.939	6.1	83	0.00
24	Fluoranthene	1.000	0.962	3.8	86	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	89	0.00
26	Pyrene	1.000	0.949	5.1	86	0.00
27 S	Terphenyl-d14	1.000	0.958	4.2	85	0.00
28	Benzo(a)anthracene	1.000	0.934	6.6	85	0.00
29	Chrysene	1.000	0.975	2.5	88	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.613	38.7#	46	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.874	12.6	79	0.00
32 I	Perylene-d12	5.000	5.000	0.0	86	0.00
33	Benzo(b)fluoranthene	1.000	0.951	4.9	82	0.00
34	Benzo(k)fluoranthene	1.000	0.984	1.6	85	0.00
35 C	Benzo(a)pyrene	1.000	0.924	7.6	80	0.00
36	Dibenzo(a,h)anthracene	1.000	0.903	9.7	79	0.00
37	Benzo(g,h,i)perylene	1.000	0.901	9.9	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0