

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062915\
 Data File : BE089997.D
 Acq On : 30 Jun 2015 1:09
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jun 30 07:17:49 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	89	0.00
2	1,4-Dioxane	1.000	0.975	2.5	90	0.00
3	n-Nitrosodimethylamine	1.000	0.908	9.2	84	0.00
4 S	2-Fluorophenol	1.000	0.785	21.5	70	0.00
5 S	Phenol-d6	1.000	0.780	22.0	70	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	88	0.00
7 S	Nitrobenzene-d5	1.000	0.867	13.3	78	0.00
8	Nitrobenzene	1.000	0.874	12.6	79	0.00
9	Naphthalene	1.000	0.960	4.0	86	0.00
10	Hexachlorobutadiene	1.000	1.021	-2.1	91	0.00
11	2-Methylnaphthalene	1.000	0.949	5.1	85	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	87	0.00
13 S	2,4,6-Tribromophenol	1.000	0.601	39.9#	54	0.00
14 S	2-Fluorobiphenyl	1.000	0.973	2.7	87	0.00
15	Acenaphthylene	1.000	0.946	5.4	85	0.00
16	Acenaphthene	1.000	0.972	2.8	86	0.00
17	Fluorene	1.000	0.959	4.1	86	0.00
18 I	Phenanthrene-d10	5.000	5.000	0.0	86	0.00
19	4-Bromophenyl-phenylether	1.000	0.967	3.3	85	0.00
20	Hexachlorobenzene	1.000	1.003	-0.3	90	0.00
21	Pentachlorophenol	1.000	0.663	33.7#	50	0.00
22	Phenanthrene	1.000	0.962	3.8	86	0.00
23	Anthracene	1.000	0.963	3.7	85	0.00
24	Fluoranthene	1.000	0.952	4.8	84	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	88	0.00
26	Pyrene	1.000	0.950	5.0	85	0.00
27 S	Terphenyl-d14	1.000	0.953	4.7	84	0.00
28	Benzo(a)anthracene	1.000	0.931	6.9	84	0.00
29	Chrysene	1.000	0.965	3.5	86	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	0.588	41.2#	41	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	0.844	15.6	75	0.00
32 I	Perylene-d12	5.000	5.000	0.0	84	0.00
33	Benzo(b)fluoranthene	1.000	0.966	3.4	81	0.00
34	Benzo(k)fluoranthene	1.000	0.966	3.4	81	0.00
35 C	Benzo(a)pyrene	1.000	0.913	8.7	76	0.00
36	Dibenzo(a,h)anthracene	1.000	0.886	11.4	75	0.00
37	Benzo(g,h,i)perylene	1.000	0.901	9.9	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0