

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062915\
 Data File : BE090008.D
 Acq On : 30 Jun 2015 8:19
 Operator : TP/IZ
 Sample : G2810-28MSD
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0520006-150625MSD

Manual Integrations
 APPROVED

apatel
 6/30/2015 4:55:37 PM

Quant Time: Jun 30 09:16:19 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.63	152	16800	5.00	ng	0.00
6) Naphthalene-d8	10.38	136	82021	5.00	ng	0.00
12) Acenaphthene-d10	14.24	164	49339	5.00	ng	0.00
18) Phenanthrene-d10	16.98	188	112338	5.00	ng	0.00
25) Chrysene-d12	21.14	240	118224	5.00	ng	0.00
32) Perylene-d12	23.47	264	121336	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.26	112	26836	6.95	ng	0.00
5) Phenol-d6	6.81	99	51393	9.51	ng	0.00
7) Nitrobenzene-d5	8.76	82	31648	4.86	ng	0.00
13) 2,4,6-Tribromophenol	15.74	330	18328	12.68	ng	0.02
14) 2-Fluorobiphenyl	12.86	172	67441	4.55	ng	0.00
27) Terphenyl-d14	19.59	244	102665	5.21	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.24	88	5099	2.21	ng	100
3) n-Nitrosodimethylamine	3.53	42	9677	3.10	ng	92
8) Nitrobenzene	8.80	77	33619	4.86	ng	97
9) Naphthalene	10.43	128	101960	5.50	ng	99
10) Hexachlorobutadiene	10.72	225	12989	4.46	ng	99
11) 2-Methylnaphthalene	12.05	142	66620	5.38	ng	99
15) Acenaphthylene	13.95	152	451613	5.07	ng	100
16) Acenaphthene	14.30	154	80444	6.27	ng	97
17) Fluorene	15.29	166	96231	5.52	ng	99
19) 4-Bromophenyl-phenylether	16.18	248	25378	5.31	ng	95
20) Hexachlorobenzene	16.30	284	107133	5.39	ng	97
21) Pentachlorophenol	16.63	266	30349	11.80	ng	94
22) Phenanthrene	17.01	178	537158	18.66	ng	99
23) Anthracene	17.10	178	232413	8.83	ng	99
24) Fluoranthene	19.02	202	1394723	43.39	ng	97
26) Pyrene	19.38	202	1338636	47.50	ng	# 93
28) Benzo(a)anthracene	21.12	228	1000463	33.88	ng	97
29) Chrysene	21.18	228	1089191	35.53	ng	95
30) Bis(2-ethylhexyl)phthalate	21.07	149	123956	8.88	ng	95
31) Indeno(1,2,3-cd)pyrene	25.89	276	819421	27.88	ng	97
33) Benzo(b)fluoranthene	22.76	252	1403874	52.45	ng	97
34) Benzo(k)fluoranthene	22.80	252	654237m	21.88	ng	
35) Benzo(a)pyrene	23.37	252	1050064	42.48	ng	98
36) Dibenzo(a,h)anthracene	25.90	278	311845	12.29	ng	99
37) Benzo(g,h,i)perylene	26.64	276	795585	30.93	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA_E\DATA\BE062915\
Data File : BE090008.D
Acq On : 30 Jun 2015 8:19
Operator : TP/IZ
Sample : G2810-28MSD
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM

Quant Time: Jun 30 09:16:19 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 07:01:01 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 7.63 min Scan# 854
Delta R.T. 0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion	Ratio	Lower	Upper
152	100		
150	157.7	123.9	185.9
115	64.9	50.6	76.0

Manual Integrations
APPROVED




apatel
6/30/2015 4:55:37 PM




#2

1,4-Dioxane
Concen: 2.21 ng
RT: 3.24 min Scan# 95
Delta R.T. 0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Lower	Upper
88	100		
58	88.6	70.9	106.3
43	34.2	27.6	41.4









#3
n-Nitrosodimethylamine
Concen: 3.10 ng
RT: 3.53 min Scan# 138
Delta R.T. 0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion	Ratio	Lower	Upper
42	100		
74	121.3	89.4	134.2
44	7.5	6.9	10.3




Manual Integrations
APPROVED


apatel
6/30/2015 4:55:37 PM



#4
2-Fluorophenol
Concen: 6.95 ng
RT: 5.26 min Scan# 392
Delta R.T. 0.01 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Lower	Upper
112	100		
64	72.1	56.3	84.5
63	37.2	29.8	44.6





#5

Phenol-d6

Concen: 9.51 ng

RT: 6.81 min Scan# 674

Delta R.T. 0.01 min

Lab File: BE090008.D


Acq: 30 Jun 2015 8:19

Instrument :

BNA_E

ClientSampleId :




X1SS0520006-150625MSD



Tgt Ion:	99	Resp:	51393
Ion Ratio	Lower	Upper	
99	100		
42	22.0	18.7	28.1
71	35.0	27.5	41.3

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#6

Naphthalene-d8

Concen: 5.00 ng




RT: 10.38 min Scan# 1377


Delta R.T. 0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion:	136	Resp:	82021
Ion Ratio	Lower	Upper	
136	100		
137	11.1	8.9	13.3
54	10.6	8.2	12.4
68	7.7	5.8	8.8





#7

Nitrobenzene-d5

Concen: 4.86 ng

RT: 8.76 min Scan# 1100

Delta R.T. -0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Instrument :

BNA_E


ClientSampleId :

X1SS0520006-150625MSD




Tgt Ion:	82	Resp:	31648
Ion Ratio	Lower	Upper	
82	100		
128	34.4	28.0	42.0
54	58.3	47.6	71.4

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



Abundance Ion 82.00 (81.70 to 82.70): BE0
Ion 128.00 (127.70 to 128.70): BE0
Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene

Concen: 4.86 ng


RT: 8.80 min Scan# 1109

Delta R.T. -0.00 min



Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion:	77	Resp:	33619
Ion Ratio	Lower	Upper	
77	100		
123	36.6	27.4	41.0
65	13.8	10.6	15.8



Abundance Ion 77.00 (76.70 to 77.70): BE0
Ion 123.00 (122.70 to 123.70): BE0
Ion 65.00 (64.70 to 65.70): BE0





#9

Naphthalene

Concen: 5.50 ng

RT: 10.43 min Scan# 1383

Delta R.T. 0.00 min

Lab File: BE090008.D


Acq: 30 Jun 2015 8:19

Instrument :

BNA_E

ClientSampleId :




X1SS0520006-150625MSD



Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.2	9.3	13.9
127	13.1	10.2	15.4

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#10

Hexachlorobutadiene

Concen: 4.46 ng




RT: 10.72 min Scan# 1422


Delta R.T. 0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.9	50.1	75.1
227	64.2	50.9	76.3









#11
2-Methylnaphthalene
Concen: 5.38 ng
RT: 12.05 min Scan# 1658
Delta R.T. 0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion:142 Resp: 66620
Ion Ratio Lower Upper
142 100
141 88.7 71.3 106.9
115 36.5 30.4 45.6




Manual Integrations
APPROVED


apatel
6/30/2015 4:55:37 PM



#12
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.24 min Scan# 2081
Delta R.T. 0.01 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion:164 Resp: 49339
Ion Ratio Lower Upper
164 100
162 102.0 82.2 123.4
160 44.5 39.0 58.6









#13
 2,4,6-Tribromophenol
 Concen: 12.68 ng
 RT: 15.74 min Scan# 2208
 Delta R.T. 0.02 min
 Lab File: BE090008.D
 Acq: 30 Jun 2015 8:19

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0520006-150625MSD

Tgt Ion	Ratio	Lower	Upper
330	100		
332	96.6	76.4	114.6
141	158.1	130.3	195.5




Manual Integrations
 APPROVED


apatel
 6/30/2015 4:55:37 PM



#14
 2-Fluorobiphenyl
 Concen: 4.55 ng
 RT: 12.86 min Scan# 1836
 Delta R.T. 0.00 min
 Lab File: BE090008.D
 Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Lower	Upper
172	100		
171	34.9	27.8	41.6
170	23.4	18.3	27.5





#15

Acenaphthylene

Concen: 5.07 ng

RT: 13.95 min Scan# 2037

Delta R.T. 0.01 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Instrument :

BNA_E





ClientSampleId :

X1SS0520006-150625MSD

Tgt Ion:	152	Resp:	451613
Ion Ratio	Lower	Upper	
152	100		
151	5.0	3.9	5.9
153	13.2	10.4	15.6

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#16

Acenaphthene

Concen: 6.27 ng




RT: 14.30 min Scan# 2090


Delta R.T. 0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion:	154	Resp:	80444
Ion Ratio	Lower	Upper	
154	100		
153	451.2	358.6	538.0
152	224.4	189.6	284.4





#17

Fluorene

Concen: 5.52 ng

RT: 15.29 min Scan# 2182

Delta R.T. -0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Instrument :

BNA_E





ClientSampleId :

X1SS0520006-150625MSD

Tgt Ion:	166	Resp:	96231
Ion Ratio	Lower	Upper	
166	100		
165	98.3	79.0	118.6
167	14.0	11.8	17.8

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#18

Phenanthrene-d10

Concen: 5.00 ng




RT: 16.98 min Scan# 2279



Delta R.T. -0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion:	188	Resp:	112338
Ion Ratio	Lower	Upper	
188	100		
94	7.8	4.1	6.1#
80	8.4	4.2	6.2#









#19
4-Bromophenyl-phenylether
Concen: 5.31 ng
RT: 16.18 min Scan# 2233
Delta R.T. -0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Manual Integrations
APPROVED


apatel
6/30/2015 4:55:37 PM

Tgt Ion	Ratio	Lower	Upper
248	100		
250	96.7	78.3	117.5
141	322.1	248.1	372.1



#20
Hexachlorobenzene
Concen: 5.39 ng
RT: 16.30 min Scan# 2240
Delta R.T. -0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Lower	Upper
284	100		
142	37.7	32.6	49.0
249	31.1	25.5	38.3









#21
Pentachlorophenol
Concen: 11.80 ng
RT: 16.63 min Scan# 2259
Delta R.T. -0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion	Ratio	Resp	Lower	Upper
266	100	30349		
268	60.5	51.8	77.8	
264	68.4	50.2	75.2	




Manual Integrations
APPROVED


apatel
6/30/2015 4:55:37 PM



#22
Phenanthrene
Concen: 18.66 ng
RT: 17.01 min Scan# 2281
Delta R.T. -0.00 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	537158		
176	19.6	15.3	22.9	
179	16.0	12.4	18.6	





#23

Anthracene

Concen: 8.83 ng

RT: 17.10 min Scan# 2286

Delta R.T. -0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Instrument :

BNA_E





ClientSampleId :

X1SS0520006-150625MSD

Tgt Ion:	178	Resp:	232413
Ion Ratio	Lower	Upper	
178	100		
176	18.4	14.7	22.1
179	14.2	12.1	18.1

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#24

Fluoranthene

Concen: 43.39 ng




RT: 19.02 min Scan# 2513

Delta R.T. 0.01 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion:	202	Resp:	1394723
Ion Ratio	Lower	Upper	
202	100		
101	14.7	10.6	15.8
203	18.3	13.8	20.6









#25
Chrysene-d12
Concen: 5.00 ng
RT: 21.14 min Scan# 2871
Delta R.T. 0.01 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion	Ratio	Resp	Lower	Upper
240	100	118224		
120	9.7	6.9	10.3	
236	26.3	20.8	31.2	




Manual Integrations
APPROVED


apatel
6/30/2015 4:55:37 PM



#26
Pyrene
Concen: 47.50 ng
RT: 19.38 min Scan# 2581
Delta R.T. 0.01 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Resp	Lower	Upper
202	100	1338636		
200	22.2	16.8	25.2	
203	22.4	13.9	20.9	





#27

Terphenyl-d14

Concen: 5.21 ng

RT: 19.59 min Scan# 2620

Delta R.T. -0.00 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Instrument :

BNA_E


ClientSampleId :

X1SS0520006-150625MSD




Tgt Ion:	244	Resp:	102665
Ion Ratio	Lower	Upper	
244	100		
212	8.4	6.6	9.8
122	13.8	9.8	14.8

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



Abundance Ion 244.00 (243.70 to 244.70): E
Ion 212.00 (211.70 to 212.70): E
Ion 122.00 (121.70 to 122.70): E



#28

Benzo(a)anthracene

Concen: 33.88 ng


RT: 21.12 min Scan# 2868

Delta R.T. 0.01 min



Lab File: BE090008.D


Acq: 30 Jun 2015 8:19

Tgt Ion:	228	Resp:	1000463
Ion Ratio	Lower	Upper	
228	100		
226	29.0	21.8	32.6
229	21.2	15.8	23.6



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E









#29
Chrysene
 Concen: 35.53 ng
 RT: 21.18 min Scan# 2876
 Delta R.T. 0.01 min
 Lab File: BE090008.D
 Acq: 30 Jun 2015 8:19

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0520006-150625MSD

Tgt Ion: 228 Resp: 1089191
 Ion Ratio Lower Upper
 228 100
 226 31.2 23.4 35.0
 229 22.8 16.2 24.2




Manual Integrations
 APPROVED


apatel
 6/30/2015 4:55:37 PM



#30
Bis(2-ethylhexyl)phthalate
 Concen: 8.88 ng
 RT: 21.07 min Scan# 2860
 Delta R.T. 0.00 min
 Lab File: BE090008.D
 Acq: 30 Jun 2015 8:19

Tgt Ion: 149 Resp: 123956
 Ion Ratio Lower Upper
 149 100
 167 22.9 20.3 30.5
 279 3.4 2.6 3.8









#31
Indeno(1,2,3-cd)pyrene
Concen: 27.88 ng
RT: 25.89 min Scan# 3851
Delta R.T. 0.02 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion	Ratio	Resp Lower	Resp Upper
276	100		
138	24.3	21.3	31.9
277	24.1	19.7	29.5




Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#32
Perylene-d12
Concen: 5.00 ng
RT: 23.47 min Scan# 3322
Delta R.T. 0.01 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Resp Lower	Resp Upper
264	100		
260	25.4	20.2	30.2
265	26.0	17.1	25.7









#33
Benzo(b)fluoranthene
Concen: 52.45 ng
RT: 22.76 min Scan# 3168
Delta R.T. 0.02 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Instrument :
BNA_E
ClientSampleId :
X1SS0520006-150625MSD

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.6	17.8	26.8
125	12.6	9.0	13.4




Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#34
Benzo(k)fluoranthene
Concen: 21.88 ng m
RT: 22.80 min Scan# 3177
Delta R.T. 0.01 min
Lab File: BE090008.D
Acq: 30 Jun 2015 8:19

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.3	17.5	26.3
125	12.2	9.5	14.3





#35

Benzo(a)pyrene

Concen: 42.48 ng

RT: 23.37 min Scan# 3301

Delta R.T. 0.02 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Instrument :

BNA_E





ClientSampleId :

X1SS0520006-150625MSD

Tgt Ion:	252	Resp:	1050064
Ion Ratio	Lower	Upper	
252	100		
253	23.3	17.4	26.2
125	13.0	10.5	15.7

Manual Integrations
APPROVED

apatel
6/30/2015 4:55:37 PM



#36

Dibenzo(a,h)anthracene

Concen: 12.29 ng




RT: 25.90 min Scan# 3853



Delta R.T. 0.01 min

Lab File: BE090008.D

Acq: 30 Jun 2015 8:19

Tgt Ion:	278	Resp:	311845
Ion Ratio	Lower	Upper	
278	100		
139	18.4	14.4	21.6
279	24.2	19.0	28.6





#37
 Benzo(g,h,i)perylene
 Concen: 30.93 ng
 RT: 26.64 min Scan# 4015
 Delta R.T. 0.03 min
 Lab File: BE090008.D
 Acq: 30 Jun 2015 8:19

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0520006-150625MSD

Tgt Ion:	276	Resp:	795585
Ion Ratio	Lower	Upper	
276	100		
277	24.0	18.8	28.2
138	24.3	18.6	27.8

Manual Integrations
 APPROVED

apatel
 6/30/2015 4:55:37 PM

