

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE063015\
 Data File : BE090016.D
 Acq On : 30 Jun 2015 18:31
 Operator : TP/IZ
 Sample : G2824-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0650624-150626

Quant Time: Jul 01 05:35:26 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	13566	5.00	ng	0.00
6) Naphthalene-d8	10.38	136	58809	5.00	ng	0.00
12) Acenaphthene-d10	14.23	164	33048	5.00	ng	0.00
18) Phenanthrene-d10	16.96	188	74554	5.00	ng	-0.02
25) Chrysene-d12	21.12	240	92231	5.00	ng	-0.01
32) Perylene-d12	23.44	264	92376	5.00	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.25	112	22452	7.20	ng	0.00
5) Phenol-d6	6.80	99	33985	7.79	ng	0.00
7) Nitrobenzene-d5	8.75	82	20329	4.35	ng	0.00
13) 2,4,6-Tribromophenol	15.73	330	9157	9.46	ng	0.00
14) 2-Fluorobiphenyl	12.85	172	43574	4.39	ng	0.00
27) Terphenyl-d14	19.58	244	70668	4.60	ng	0.00
Target Compounds						Qvalue
22) Phenanthrene	17.01	178	2900	0.15	ng	97
23) Anthracene	17.10	178	565m	0.03	ng	
24) Fluoranthene	19.01	202	9891	0.46	ng	99
26) Pyrene	19.37	202	9588	0.44	ng	96
28) Benzo(a)anthracene	21.11	228	7601	0.33	ng	98
29) Chrysene	21.16	228	8095	0.34	ng	97
30) Bis(2-ethylhexyl)phthalate	21.05	149	2207	0.50	ng	100
31) Indeno(1,2,3-cd)pyrene	25.84	276	5841	0.25	ng	98
33) Benzo(b)fluoranthene	22.74	252	11638	0.57	ng	# 94
34) Benzo(k)fluoranthene	22.78	252	4414m	0.19	ng	
35) Benzo(a)pyrene	23.33	252	7559	0.40	ng	# 93
36) Dibenzo(a,h)anthracene	25.85	278	1744	0.09	ng	# 71
37) Benzo(g,h,i)perylene	26.57	276	6092	0.31	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE063015\
Data File : BE090016.D
Acq On : 30 Jun 2015 18:31
Operator : TP/IZ
Sample : G2824-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0650624-150626

Quant Time: Jul 01 05:35:26 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 07:01:01 2015
Response via : Initial Calibration

