

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE063015\
 Data File : BE090021.D
 Acq On : 30 Jun 2015 21:30
 Operator : TP/IZ
 Sample : G2824-15
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0670006-150626

Quant Time: Jul 01 16:50:55 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	15636	5.00	ng	0.00
6) Naphthalene-d8	10.37	136	68882	5.00	ng	-0.02
12) Acenaphthene-d10	14.23	164	38694	5.00	ng	0.00
18) Phenanthrene-d10	16.96	188	90877	5.00	ng	-0.02
25) Chrysene-d12	21.12	240	108545	5.00	ng	-0.01
32) Perylene-d12	23.44	264	102593	5.00	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.24	112	23518	6.55	ng	0.00
5) Phenol-d6	6.80	99	35451	7.05	ng	0.00
7) Nitrobenzene-d5	8.75	82	21115	3.86	ng	0.00
13) 2,4,6-Tribromophenol	15.73	330	9751	8.60	ng	0.00
14) 2-Fluorobiphenyl	12.85	172	45293	3.89	ng	0.00
27) Terphenyl-d14	19.58	244	71713	3.97	ng	0.00
Target Compounds						Qvalue
22) Phenanthrene	17.01	178	2297	0.10	ng	95
24) Fluoranthene	19.01	202	5717	0.22	ng	99
26) Pyrene	19.37	202	5387	0.21	ng	97
28) Benzo(a)anthracene	21.11	228	3923	0.14	ng	95
29) Chrysene	21.15	228	4338	0.15	ng	96
30) Bis(2-ethylhexyl)phthalate	21.05	149	1522	0.42	ng	94
31) Indeno(1,2,3-cd)pyrene	25.84	276	3001	0.11	ng	99
33) Benzo(b)fluoranthene	22.73	252	5785	0.26	ng	# 88
34) Benzo(k)fluoranthene	22.77	252	2260m	0.09	ng	
35) Benzo(a)pyrene	23.33	252	3856	0.18	ng	# 86
36) Dibenzo(a,h)anthracene	25.85	278	978	0.05	ng	# 58
37) Benzo(g,h,i)perylene	26.57	276	3036	0.14	ng	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE063015\
Data File : BE090021.D
Acq On : 30 Jun 2015 21:30
Operator : TP/IZ
Sample : G2824-15
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0670006-150626

Quant Time: Jul 01 16:50:55 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 07:01:01 2015
Response via : Initial Calibration

