

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE063015\
 Data File : BE090013.D
 Acq On : 30 Jun 2015 16:35
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 01 05:16:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	-0.01
2	1,4-Dioxane	0.686	0.608	11.4	83	-0.01
3	n-Nitrosodimethylamine	0.930	0.824	11.4	83	-0.02
4 S	2-Fluorophenol	1.149	1.201	-4.5	95	-0.01
5 S	Phenol-d6	1.608	1.662	-3.4	95	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	92	-0.01
7 S	Nitrobenzene-d5	0.397	0.371	6.5	87	0.00
8	Nitrobenzene	0.421	0.389	7.6	87	0.00
9	Naphthalene	1.130	1.045	7.5	86	-0.01
10	Hexachlorobutadiene	0.178	0.168	5.6	88	0.00
11	2-Methylnaphthalene	0.755	0.677	10.3	84	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
13 S	2,4,6-Tribromophenol	0.147	0.172	-17.0	101	0.00
14 S	2-Fluorobiphenyl	1.503	1.397	7.1	81	0.00
15	Acenaphthylene	9.019	8.502	5.7	83	0.00
16	Acenaphthene	1.300	1.219	6.2	81	0.00
17	Fluorene	1.768	1.613	8.8	80	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	77	-0.02
19	4-Bromophenyl-phenylether	0.213	0.193	9.4	71	0.00
20	Hexachlorobenzene	0.885	0.824	6.9	74	-0.02
21	Pentachlorophenol	0.086	0.094	-9.3	89	0.00
22	Phenanthrene	1.281	1.167	8.9	72	0.00
23	Anthracene	1.172	1.149	2.0	77	0.00
24	Fluoranthene	1.431	1.381	3.5	75	0.00
25 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
26	Pyrene	1.192	1.125	5.6	77	0.00
27 S	Terphenyl-d14	0.833	0.762	8.5	73	0.00
28	Benzo(a)anthracene	1.249	1.205	3.5	79	0.00
29	Chrysene	1.297	1.173	9.6	73	0.00
30	Bis(2-ethylhexyl)phthalate	0.398	0.665	-67.1#	146	0.00
31	Indeno(1,2,3-cd)pyrene	1.243	1.413	-13.7	93	-0.02
32 I	Perylene-d12	1.000	1.000	0.0	83	0.00
33	Benzo(b)fluoranthene	1.103	1.176	-6.6	89	0.00
34	Benzo(k)fluoranthene	1.232	1.248	-1.3	84	0.00
35 C	Benzo(a)pyrene	1.019	1.111	-9.0	90	-0.01
36	Dibenzo(a,h)anthracene	1.046	1.135	-8.5	91	-0.01
37	Benzo(g,h,i)perylene	1.060	1.162	-9.6	92	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0