

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE063015\
 Data File : BE090013.D
 Acq On : 30 Jun 2015 16:35
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 01 05:16:33 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	90	-0.01
2	1,4-Dioxane	1.000	0.885	11.5	83	-0.01
3	n-Nitrosodimethylamine	1.000	0.885	11.5	83	-0.02
4 S	2-Fluorophenol	1.000	1.046	-4.6	95	-0.01
5 S	Phenol-d6	1.000	1.034	-3.4	95	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	92	-0.01
7 S	Nitrobenzene-d5	1.000	0.935	6.5	87	0.00
8	Nitrobenzene	1.000	0.924	7.6	87	0.00
9	Naphthalene	1.000	0.924	7.6	86	-0.01
10	Hexachlorobutadiene	1.000	0.943	5.7	88	0.00
11	2-Methylnaphthalene	1.000	0.896	10.4	84	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	85	0.00
13 S	2,4,6-Tribromophenol	1.000	1.171	-17.1	101	0.00
14 S	2-Fluorobiphenyl	1.000	0.929	7.1	81	0.00
15	Acenaphthylene	1.000	0.943	5.7	83	0.00
16	Acenaphthene	1.000	0.938	6.2	81	0.00
17	Fluorene	1.000	0.912	8.8	80	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	77	-0.02
19	4-Bromophenyl-phenylether	1.000	0.909	9.1	71	0.00
20	Hexachlorobenzene	1.000	0.931	6.9	74	-0.02
21	Pentachlorophenol	1.000	1.065	-6.5	89	0.00
22	Phenanthrene	1.000	0.911	8.9	72	0.00
23	Anthracene	1.000	0.980	2.0	77	0.00
24	Fluoranthene	1.000	0.965	3.5	75	0.00
25 I	Chrysene-d12	5.000	5.000	0.0	80	0.00
26	Pyrene	1.000	0.944	5.6	77	0.00
27 S	Terphenyl-d14	1.000	0.915	8.5	73	0.00
28	Benzo(a)anthracene	1.000	0.965	3.5	79	0.00
29	Chrysene	1.000	0.905	9.5	73	0.00
30	Bis(2-ethylhexyl)phthalate	1.000	1.393	-39.3#	146	0.00
31	Indeno(1,2,3-cd)pyrene	1.000	1.137	-13.7	93	-0.02
32 I	Perylene-d12	5.000	5.000	0.0	83	0.00
33	Benzo(b)fluoranthene	1.000	1.066	-6.6	89	0.00
34	Benzo(k)fluoranthene	1.000	1.013	-1.3	84	0.00
35 C	Benzo(a)pyrene	1.000	1.090	-9.0	90	-0.01
36	Dibenzo(a,h)anthracene	1.000	1.086	-8.6	91	-0.01
37	Benzo(g,h,i)perylene	1.000	1.096	-9.6	92	-0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0