

Data Path : Z:\HPCHEM1\BNA_E\Data\BE063015\
 Data File : BE090029.D
 Acq On : 1 Jul 2015 2:55
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 01 06:20:16 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.686	0.590	14.0	86	-0.01
3	n-Nitrosodimethylamine	0.930	0.749	19.5	81	-0.01
4 S	2-Fluorophenol	1.149	1.108	3.6	93	0.00
5 S	Phenol-d6	1.608	1.551	3.5	94	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	96	-0.01
7 S	Nitrobenzene-d5	0.397	0.351	11.6	87	0.00
8	Nitrobenzene	0.421	0.364	13.5	86	0.00
9	Naphthalene	1.130	1.041	7.9	90	0.00
10	Hexachlorobutadiene	0.178	0.167	6.2	92	0.00
11	2-Methylnaphthalene	0.755	0.671	11.1	87	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	88	0.00
13 S	2,4,6-Tribromophenol	0.147	0.138	6.1	85	0.00
14 S	2-Fluorobiphenyl	1.503	1.387	7.7	84	0.00
15	Acenaphthylene	9.019	8.489	5.9	86	0.00
16	Acenaphthene	1.300	1.230	5.4	85	-0.01
17	Fluorene	1.768	1.595	9.8	82	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	81	-0.02
19	4-Bromophenyl-phenylether	0.213	0.187	12.2	73	0.00
20	Hexachlorobenzene	0.885	0.786	11.2	75	-0.02
21	Pentachlorophenol	0.086	0.073	15.1	73	0.00
22	Phenanthrene	1.281	1.132	11.6	74	0.00
23	Anthracene	1.172	1.092	6.8	77	0.00
24	Fluoranthene	1.431	1.298	9.3	75	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	81	-0.01
26	Pyrene	1.192	1.122	5.9	77	-0.01
27 S	Terphenyl-d14	0.833	0.751	9.8	73	-0.01
28	Benzo(a)anthracene	1.249	1.157	7.4	76	-0.01
29	Chrysene	1.297	1.165	10.2	73	-0.01
30	Bis(2-ethylhexyl)phthalate	0.398	0.484	-21.6	107	-0.01
31	Indeno(1,2,3-cd)pyrene	1.243	1.283	-3.2	85	-0.03
32 I	Perylene-d12	1.000	1.000	0.0	80	-0.02
33	Benzo(b)fluoranthene	1.103	1.129	-2.4	82	-0.01
34	Benzo(k)fluoranthene	1.232	1.226	0.5	80	-0.01
35 C	Benzo(a)pyrene	1.019	1.048	-2.8	82	-0.02
36	Dibenzo(a,h)anthracene	1.046	1.067	-2.0	83	-0.02
37	Benzo(g,h,i)perylene	1.060	1.088	-2.6	83	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0