

Data Path : Z:\HPCHEM1\BNA_E\Data\BE063017\
 Data File : BE093375.D
 Acq On : 30 Jun 2017 10:23
 Operator : SJ/JU
 Sample : SSTDC00.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.428

Quant Time: Jun 30 11:12:33 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE060717.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 30 10:50:58 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	107	0.00
3	Naphthalene	0.992	1.022	-3.0	112	0.00
4 SURR	2-Methylnaphthalene-d10	0.586	0.627	-7.0	114	0.00
5	2-Methylnaphthalene	0.679	0.729	-7.4	116	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	121	0.00
7	Acenaphthylene	1.717	1.827	-6.4	129	0.00
8 C	Acenaphthene	1.385	1.363	1.6	119	0.00
9	Fluorene	1.624	1.671	-2.9	126	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	118	-0.02
11	Pentachlorophenol	0.059	0.034	42.4#	72	0.00
12	Phenanthrene	1.101	1.103	-0.2	122	0.00
13	Anthracene	0.958	1.064	-11.1	133	0.00
14 SURR	Fluoranthene-d10	1.088	1.285	-18.1	141	0.00
15 C	Fluoranthene	1.350	1.452	-7.6	129	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	139	0.00
17	Pyrene	1.180	1.105	6.4	123	0.00
18	Benzo(a)anthracene	1.067	1.132	-6.1	148	0.00
19	Chrysene	1.107	1.111	-0.4	141	0.00
20 I	Perylene-d12	1.000	1.000	0.0	159#	0.00
21	Benzo(b)fluoranthene	1.214	1.199	1.2	162#	0.00
22	Benzo(k)fluoranthene	1.289	1.252	2.9	158#	0.00
23 C	Benzo(a)pyrene	1.148	1.135	1.1	164#	0.00
24	Indeno(1,2,3-cd)pyrene	1.329	1.318	0.8	164#	0.00
25	Dibenzo(a,h)anthracene	1.119	1.116	0.3	163#	0.00
26	Benzo(g,h,i)perylene	1.140	1.122	1.6	164#	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0