

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070115\
 Data File : BE090046.D
 Acq On : 1 Jul 2015 18:22
 Operator : TP/IZ
 Sample : G2811-07
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0720624-150625

Quant Time: Jul 02 05:15:39 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.61	152	14928	5.00	ng	-0.01
6) Naphthalene-d8	10.37	136	64598	5.00	ng	-0.01
12) Acenaphthene-d10	14.23	164	36356	5.00	ng	0.00
18) Phenanthrene-d10	16.96	188	83605	5.00	ng	-0.02
25) Chrysene-d12	21.12	240	104139	5.00	ng	-0.01
32) Perylene-d12	23.44	264	97864	5.00	ng	-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.24	112	26363	7.69	ng	0.00
5) Phenol-d6	6.79	99	39698	8.27	ng	0.00
7) Nitrobenzene-d5	8.75	82	23475	4.57	ng	0.00
13) 2,4,6-Tribromophenol	15.73	330	8512	7.99	ng	0.00
14) 2-Fluorobiphenyl	12.85	172	51271	4.69	ng	-0.01
27) Terphenyl-d14	19.58	244	83286	4.80	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.24	88	186	0.09	ng	# 67
9) Naphthalene	10.42	128	2406	0.16	ng	# 93
11) 2-Methylnaphthalene	12.04	142	465	0.05	ng	# 94
16) Acenaphthene	14.28	154	1805	0.19	ng	94
17) Fluorene	15.27	166	983	0.08	ng	# 94
21) Pentachlorophenol	16.60	266	21	0.27	ng	# 54
22) Phenanthrene	17.01	178	60695	2.83	ng	99
23) Anthracene	17.01	178	60668	3.10	ng	98
24) Fluoranthene	19.01	202	196415	8.21	ng	99
26) Pyrene	19.37	202	183584	7.40	ng	# 96
28) Benzo(a)anthracene	21.10	228	136946	5.27	ng	98
29) Chrysene	21.16	228	142867	5.29	ng	98
30) Bis(2-ethylhexyl)phthalate	21.05	149	693	0.36	ng	97
31) Indeno(1,2,3-cd)pyrene	25.84	276	67388	2.60	ng	97
33) Benzo(b)fluoranthene	22.74	252	172653	8.00	ng	98
34) Benzo(k)fluoranthene	22.74	252	172653	7.16	ng	98
35) Benzo(a)pyrene	23.33	252	113995	5.72	ng	98
36) Dibenzo(a,h)anthracene	25.85	278	17284	0.84	ng	96
37) Benzo(g,h,i)perylene	26.58	276	66770	3.22	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070115\
Data File : BE090046.D
Acq On : 1 Jul 2015 18:22
Operator : TP/IZ
Sample : G2811-07
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0720624-150625

Quant Time: Jul 02 05:15:39 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 07:01:01 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 7.61 min Scan# 852

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:152 Resp: 14928

Ion Ratio Lower Upper

152 100

150 154.8 123.9 185.9




115 63.9 50.6 76.0



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 0.09 ng

RT: 3.24 min Scan# 94

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion: 88 Resp: 186

Ion Ratio Lower Upper

88 100

58 76.3 70.9 106.3



43 83.3 27.6 41.4#




Abundance Ion 88.00 (87.70 to 88.70): BE0

Ion 58.00 (57.70 to 58.70): BE0

Ion 43.00 (42.70 to 43.70): BE0





#4

2-Fluorophenol

Concen: 7.69 ng

RT: 5.24 min Scan# 390

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion: 112 Resp: 26363

Ion Ratio Lower Upper

112 100

64 70.9 56.3 84.5




63 36.8 29.8 44.6



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#5

Phenol-d6

Concen: 8.27 ng

RT: 6.79 min Scan# 672

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion: 99 Resp: 39698

Ion Ratio Lower Upper

99 100

42 20.9 18.7 28.1



71 34.1 27.5 41.3




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#6

Naphthalene-d8

Concen: 5.00 ng

RT: 10.37 min Scan# 1375

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625

Tgt Ion:136 Resp: 64598


Ion Ratio Lower Upper

136 100

137 11.0 8.9 13.3

54 11.4 8.2 12.4

68 8.2 5.8 8.8






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#7

Nitrobenzene-d5

Concen: 4.57 ng

RT: 8.75 min Scan# 1098

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion: 82 Resp: 23475

Ion Ratio Lower Upper

82 100

128 34.8 28.0 42.0



54 58.5 47.6 71.4




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Naphthalene

Concen: 0.16 ng

RT: 10.42 min Scan# 1382

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:128 Resp: 2406

Ion Ratio Lower Upper

128 100

129 14.5 9.3 13.9#




127 15.4 10.2 15.4#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#11

2-Methylnaphthalene

Concen: 0.05 ng

RT: 12.04 min Scan# 1656

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:142 Resp: 465

Ion Ratio Lower Upper

142 100

141 87.5 71.3 106.9



115 48.3 30.4 45.6#




Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E





#12

Acenaphthene-d10

Concen: 5.00 ng

RT: 14.23 min Scan# 2079

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:164 Resp: 36356

Ion Ratio Lower Upper

164 100

162 102.5 82.2 123.4




160 47.4 39.0 58.6



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#13

2,4,6-Tribromophenol

Concen: 7.99 ng

RT: 15.73 min Scan# 2207

Delta R.T. -0.00 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:330 Resp: 8512

Ion Ratio Lower Upper

330 100

332 99.2 76.4 114.6



141 175.8 130.3 195.5




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E





#14

2-Fluorobiphenyl

Concen: 4.69 ng

RT: 12.85 min Scan# 1834

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625





Tgt Ion:172 Resp: 51271

Ion Ratio Lower Upper

172 100

171 34.9 27.8 41.6

170 23.6 18.3 27.5



#16

Acenaphthene

Concen: 0.19 ng

RT: 14.28 min Scan# 2088

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22




Tgt Ion:154 Resp: 1805


Ion Ratio Lower Upper

154 100

153 462.9 358.6 538.0

152 247.5 189.6 284.4





#17

Fluorene

Concen: 0.08 ng

RT: 15.27 min Scan# 2181

Delta R.T. -0.02 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:166 Resp: 983

Ion Ratio Lower Upper

166 100

165 104.1 79.0 118.6




167 20.0 11.8 17.8#



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#18

Phenanthrene-d10

Concen: 5.00 ng

RT: 16.96 min Scan# 2278

Delta R.T. -0.02 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:188 Resp: 83605

Ion Ratio Lower Upper

188 100

94 12.1 4.1 6.1#



80 13.1 4.2 6.2#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0





#21

Pentachlorophenol

Concen: 0.27 ng

RT: 16.60 min Scan# 2257

Delta R.T. -0.04 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:266 Resp: 21

Ion Ratio Lower Upper

266 100

268 102.1 51.8 77.8#




264 97.9 50.2 75.2#



Abundance Ion 265.70 (265.40 to 266.40): E

Ion 268.00 (267.70 to 268.70): E

Ion 264.00 (263.70 to 264.70): E



#22

Phenanthrene

Concen: 2.83 ng

RT: 17.01 min Scan# 2281

Delta R.T. -0.00 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:178 Resp: 60695

Ion Ratio Lower Upper

178 100

176 19.4 15.3 22.9



179 15.7 12.4 18.6




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#23

Anthracene

Concen: 3.10 ng

RT: 17.01 min Scan# 2281

Delta R.T. -0.09 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:178 Resp: 60668

Ion Ratio Lower Upper

178 100

176 19.4 14.7 22.1




179 15.7 12.1 18.1



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#24

Fluoranthene

Concen: 8.21 ng

RT: 19.01 min Scan# 2510

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:202 Resp: 196415

Ion Ratio Lower Upper

202 100

101 13.5 10.6 15.8



203 17.4 13.8 20.6




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#25

Chrysene-d12

Concen: 5.00 ng

RT: 21.12 min Scan# 2868

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:240 Resp: 104139

Ion Ratio Lower Upper

240 100

120 10.1 6.9 10.3




236 26.8 20.8 31.2



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#26

Pyrene

Concen: 7.40 ng

RT: 19.37 min Scan# 2578

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:202 Resp: 183584

Ion Ratio Lower Upper

202 100

200 21.4 16.8 25.2



203 21.0 13.9 20.9#




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E









#27
 Terphenyl-d14
 Concen: 4.80 ng
 RT: 19.58 min Scan# 2618
 Delta R.T. -0.01 min
 Lab File: BE090046.D
 Acq: 1 Jul 2015 18:22




Instrument :
 BNA_E
 ClientSampleId :
 X1SS0720624-150625


Tgt Ion:244 Resp: 83286
 Ion Ratio Lower Upper
 244 100
 212 7.9 6.6 9.8
 122 12.2 9.8 14.8



#28
 Benzo(a)anthracene
 Concen: 5.27 ng
 RT: 21.10 min Scan# 2865
 Delta R.T. -0.01 min
 Lab File: BE090046.D
 Acq: 1 Jul 2015 18:22

Tgt Ion:228 Resp: 136946
 Ion Ratio Lower Upper
 228 100
 226 28.2 21.8 32.6
 229 20.3 15.8 23.6





#29

Chrysene

Concen: 5.29 ng

RT: 21.16 min Scan# 2873

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:228 Resp: 142867

Ion Ratio Lower Upper

228 100

226 30.1 23.4 35.0




229 21.5 16.2 24.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#30

Bis(2-ethylhexyl)phthalate

Concen: 0.36 ng

RT: 21.05 min Scan# 2858

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:149 Resp: 693

Ion Ratio Lower Upper

149 100

167 26.8 20.3 30.5



279 3.5 2.6 3.8




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E





#31

Indeno(1,2,3-cd)pyrene

Concen: 2.60 ng

RT: 25.84 min Scan# 3839

Delta R.T. -0.03 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:276 Resp: 67388

Ion Ratio Lower Upper

276 100

138 24.2 21.3 31.9




277 23.5 19.7 29.5



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#32

Perylene-d12

Concen: 5.00 ng

RT: 23.44 min Scan# 3316

Delta R.T. -0.02 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:264 Resp: 97864

Ion Ratio Lower Upper

264 100

260 25.1 20.2 30.2



265 22.4 17.1 25.7




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#33

Benzo(b)fluoranthene

Concen: 8.00 ng

RT: 22.74 min Scan# 3162

Delta R.T. -0.01 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625


Tgt Ion:252 Resp: 172653

Ion Ratio Lower Upper

252 100

253 23.0 17.8 26.8




125 12.1 9.0 13.4



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#34

Benzo(k)fluoranthene

Concen: 7.16 ng

RT: 22.74 min Scan# 3162

Delta R.T. -0.05 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22


Tgt Ion:252 Resp: 172653

Ion Ratio Lower Upper

252 100

253 23.0 17.5 26.3



125 12.1 9.5 14.3




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#35

Benzo(a)pyrene

Concen: 5.72 ng

RT: 23.33 min Scan# 3293

Delta R.T. -0.02 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625





Tgt Ion:252 Resp: 113995

Ion Ratio Lower Upper

252 100

253 23.1 17.4 26.2

125 12.6 10.5 15.7



#36

Dibenzo(a,h)anthracene

Concen: 0.84 ng

RT: 25.85 min Scan# 3843

Delta R.T. -0.03 min

Lab File: BE090046.D

Acq: 1 Jul 2015 18:22




Tgt Ion:278 Resp: 17284


Ion Ratio Lower Upper

278 100

139 19.5 14.4 21.6

279 25.6 19.0 28.6





#37

Benzo(g,h,i)perylene

Concen: 3.22 ng

RT: 26.58 min Scan# 4002

Delta R.T. -0.03 min

Lab File: BE090046.D


Acq: 1 Jul 2015 18:22

Instrument :

BNA_E

ClientSampleId :

X1SS0720624-150625



Tgt Ion: 276 Resp: 66770

Ion Ratio Lower Upper

276 100

277 24.5 18.8 28.2

138 24.8 18.6 27.8

