

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070115\
 Data File : BE090049.D
 Acq On : 1 Jul 2015 20:12
 Operator : TP/IZ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 02 05:16:31 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2	1,4-Dioxane	1.000	0.859	14.1	91	0.00
3	n-Nitrosodimethylamine	1.000	0.825	17.5	88	-0.01
4 S	2-Fluorophenol	1.000	0.974	2.6	99	0.00
5 S	Phenol-d6	1.000	0.988	1.2	102	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	106	-0.02
7 S	Nitrobenzene-d5	1.000	0.871	12.9	94	0.00
8	Nitrobenzene	1.000	0.867	13.3	95	0.00
9	Naphthalene	1.000	0.920	8.0	99	0.00
10	Hexachlorobutadiene	1.000	0.935	6.5	100	0.00
11	2-Methylnaphthalene	1.000	0.896	10.4	96	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	100	0.00
13 S	2,4,6-Tribromophenol	1.000	0.878	12.2	89	0.00
14 S	2-Fluorobiphenyl	1.000	0.918	8.2	94	-0.01
15	Acenaphthylene	1.000	0.938	6.2	96	0.00
16	Acenaphthene	1.000	0.951	4.9	96	-0.01
17	Fluorene	1.000	0.910	9.0	93	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	93	-0.02
19	4-Bromophenyl-phenylether	1.000	0.875	12.5	83	-0.02
20	Hexachlorobenzene	1.000	0.940	6.0	91	-0.02
21	Pentachlorophenol	1.000	0.770	23.0	68	0.00
22	Phenanthrene	1.000	0.901	9.9	86	0.00
23	Anthracene	1.000	0.916	8.4	87	0.00
24	Fluoranthene	1.000	0.918	8.2	87	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	90	-0.01
26	Pyrene	1.000	0.952	4.8	87	-0.01
27 S	Terphenyl-d14	1.000	0.921	7.9	83	-0.01
28	Benzo(a)anthracene	1.000	0.957	4.3	88	-0.01
29	Chrysene	1.000	0.909	9.1	83	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	0.995	0.5	104	-0.01
31	Indeno(1,2,3-cd)pyrene	1.000	1.001	-0.1	92	-0.03
32 I	Perylene-d12	5.000	5.000	0.0	87	-0.02
33	Benzo(b)fluoranthene	1.000	1.090	-9.0	96	-0.01
34	Benzo(k)fluoranthene	1.000	1.008	-0.8	88	-0.01
35 C	Benzo(a)pyrene	1.000	1.070	-7.0	93	-0.02
36	Dibenzo(a,h)anthracene	1.000	1.008	-0.8	89	-0.03
37	Benzo(g,h,i)perylene	1.000	1.005	-0.5	88	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0