


n-Nitrosodimethylamine

Response Ratio




Resp Ratio = $6.31 \times 10^{-1} \times \text{Amt} - 2.03 \times 10^{-1}$
Coef of Det (r^2) = 0.999 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070117.M
Calibration Table Last Updated: Sat Jul 01 15:43:24 2017

2,4,6-Tribromophenol

Response Ratio




Resp Ratio = $2.22 \times 10^{-1} \times \text{Amt} - 6.06 \times 10^{-2}$
Coef of Det (r^2) = 0.996 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070117.M
Calibration Table Last Updated: Sat Jul 01 15:43:24 2017


4-Bromophenyl-phenylether

Response Ratio



Resp Ratio = $1.16 \times 10^{-1} \times \text{Amt} - 3.16 \times 10^{-2}$
Coef of Det (r^2) = 0.993 Curve Fit: Linear

Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070117.M
Calibration Table Last Updated: Sat Jul 01 15:43:24 2017



Method Name: Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070117.M
Calibration Table Last Updated: Sat Jul 01 15:43:24 2017