

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070117\
 Data File : BE093384.D
 Acq On : 30 Jun 2017 19:40
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE070117

Quant Time: Jul 03 05:08:19 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070117.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 01 15:21:30 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	1,4-Dioxane	0.826	0.836	-1.2	100	0.00
3	n-Nitrosodimethylamine	0.497	0.439	11.7	102	0.00
4 S	2-Fluorophenol	1.247	1.263	-1.3	100	0.00
5 S	Phenol-d6	1.861	1.851	0.5	103	0.00
6	bis(2-Chloroethyl)ether	1.312	1.312	0.0	104	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.238	0.269	-13.0	123	0.00
9	Naphthalene	4.155	4.332	-4.3	103	0.00
10	Hexachlorobutadiene	0.167	0.177	-6.0	101	0.00
11 SURR	2-Methylnaphthalene-d10	0.639	0.660	-3.3	102	0.00
12	2-Methylnaphthalene	0.708	0.733	-3.5	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	95	-0.02
14 S	2,4,6-Tribromophenol	0.172	0.161	6.4	98	0.00
15 S	2-Fluorobiphenyl	1.566	1.638	-4.6	100	0.00
16	Acenaphthylene	1.739	1.712	1.6	99	0.00
17	Acenaphthene	1.224	1.256	-2.6	101	0.00
18	Fluorene	1.683	1.736	-3.1	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.091	0.113	-24.2	111	0.00
21	Hexachlorobenzene	0.153	0.187	-22.2	109	0.00
22	Pentachlorophenol	0.018	0.016	11.1	100	0.00
23	Phenanthrene	0.968	1.110	-14.7	99	0.00
24	Anthracene	0.974	0.907	6.9	91	0.00
25 SURR	Fluoranthene-d10	3.688	3.943	-6.9	102	0.00
26	Fluoranthene	1.049	1.117	-6.5	102	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	97	-0.01
28	Pyrene	1.056	1.086	-2.8	102	0.00
29 S	Terphenyl-d14	0.742	0.777	-4.7	103	0.00
30	Benzo(a)anthracene	1.088	1.108	-1.8	103	0.00
31	Chrysene	1.137	1.200	-5.5	104	0.00
32	Bis(2-ethylhexyl)phthalate	1.612	1.448	10.2	104	-0.01
33	Indeno(1,2,3-cd)pyrene	1.112	1.160	-4.3	104	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.256	1.304	-3.8	104	0.00
36	Benzo(k)fluoranthene	1.260	1.312	-4.1	105	0.00
37 C	Benzo(a)pyrene	1.116	1.148	-2.9	105	0.00
38	Dibenzo(a,h)anthracene	1.137	1.171	-3.0	104	0.00
39	Benzo(g,h,i)perylene	1.159	1.208	-4.2	103	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070117\
Data File : BE093384.D
Acq On : 30 Jun 2017 19:40
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE070117

Quant Time: Jul 03 05:08:19 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070117.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 01 15:21:30 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		