

Data File : BE093418.D
Acq On : 2 Jul 2017 17:44
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4

Quant Time: Jul 03 06:31:22 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jul 02 13:43:38 2017
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.94	152	3239	0.40	ng	0.00
7) Naphthalene-d8	10.72	136	15700	0.40	ng	0.00
13) Acenaphthene-d10	14.54	164	10251	0.40	ng	0.00
19) Phenanthrene-d10	17.27	188	29134	0.40	ng	0.00
27) Chrysene-d12	21.43	240	36091	0.40	ng	0.00
34) Perylene-d12	23.94	264	29449	0.40	ng	0.00

System Monitoring Compounds						
4) 2-Fluorophenol	5.53	112	3694	0.38	ng	0.00
5) Phenol-d6	7.10	99	5540	0.38	ng	0.00
8) Nitrobenzene-d5	9.08	82	3844	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.31	152	9758	0.39	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	1044	0.32	ng	0.00
15) 2-Fluorobiphenyl	13.18	172	15330	0.39	ng	0.00
25) Fluoranthene-d10	19.29	212	136380	0.39	ng	0.00
29) Terphenyl-d14	19.87	244	25290	0.39	ng	0.00

Target Compounds				Qvalue		
2) 1,4-Dioxane	3.45	88	2623	0.38	ng	94
3) n-Nitrosodimethylamine	3.76	42	1274	0.38	ng	# 84
6) bis(2-Chloroethyl)ether	7.36	93	4203	0.39	ng	# 97
9) Naphthalene	10.77	128	64063	0.39	ng	# 73
10) Hexachlorobutadiene	11.08	225	2453	0.39	ng	99
12) 2-Methylnaphthalene	12.38	142	10749	0.39	ng	98
16) Acenaphthylene	14.26	152	15470	0.36	ng	# 88
17) Acenaphthene	14.60	154	11909	0.38	ng	99
18) Fluorene	15.59	166	16471	0.38	ng	# 87
20) 4-Bromophenyl-phenylether	16.45	248	6467	0.40	ng	# 26
21) Hexachlorobenzene	16.58	284	6748	0.39	ng	# 100
22) Pentachlorophenol	16.91	266	755	0.33	ng	# 78
23) Phenanthrene	17.30	178	42375	0.40	ng	99
24) Anthracene	17.30	178	43126	0.66	ng	99
26) Fluoranthene	19.32	202	39306	0.39	ng	98
28) Pyrene	19.67	202	39839	0.39	ng	# 99
30) Benzo(a)anthracene	21.40	228	37262	0.38	ng	93
31) Chrysene	21.46	228	39499	0.39	ng	94
32) Bis(2-ethylhexyl)phthalate	21.33	149	46596	0.33	ng	# 66
33) Indeno(1,2,3-cd)pyrene	26.58	276	34924	0.39	ng	95
35) Benzo(b)fluoranthene	23.16	252	36241	0.38	ng	97
36) Benzo(k)fluoranthene	23.21	252	37561	0.40	ng	95
37) Benzo(a)pyrene	23.82	252	31472	0.38	ng	# 96
38) Dibenzo(a,h)anthracene	26.60	278	31578	0.39	ng	# 91
39) Benzo(g,h,i)perylene	27.40	276	31889	0.39	ng	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : BE093418.D

Acq On : 2 Jul 2017 17:44

Operator : SJ/JU

Sample : SSTDCCC0.4

Misc :

ALS Vial : 4 Sample Multiplier: 1

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4


Quant Time: Jul 03 06:31:22 2017


Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Sun Jul 02 13:43:38 2017

Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.94 min Scan# 404
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 152 Resp: 3239
 Ion Ratio Lower Upper
 152 100
 150 152.3 125.1 187.7
 115 67.7 55.7 83.5



#2
 1,4-Dioxane
 Concen: 0.38 ng
 RT: 3.45 min Scan# 47
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion: 88 Resp: 2623
 Ion Ratio Lower Upper
 88 100
 58 85.1 62.2 93.4
 43 28.3 23.2 34.8








#3
 n-Nitrosodimethylamine
 Concen: 0.38 ng
 RT: 3.76 min Scan# 72
 Delta R.T. 0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 1274
 Ion Ratio Lower Upper
 42 100
 74 137.4 99.1 148.7
 44 35.8 7.4 11.2#



#4
 2-Fluorophenol
 Concen: 0.38 ng
 RT: 5.53 min Scan# 212
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion: 112 Resp: 3694
 Ion Ratio Lower Upper
 112 100
 64 69.4 56.4 84.6
 63 144.2 29.3 43.9#








#5
 Phenol-d6
 Concen: 0.38 ng
 RT: 7.10 min Scan# 337
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 99 Resp: 5540
 Ion Ratio Lower Upper
 99 100
 42 19.3 0.0 0.0#
 71 34.1 0.0 0.0#



#6
 bis(2-Chloroethyl)ether
 Concen: 0.39 ng
 RT: 7.36 min Scan# 358
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion: 93 Resp: 4203
 Ion Ratio Lower Upper
 93 100
 63 350.2 0.0 0.0#
 95 33.2 25.4 38.0








#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.72 min Scan# 578
Delta R.T. 0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44



Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	15700
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.8	14.8
54	44.1	10.3	15.5#
68	8.7	9.0	13.4#



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 9.08 min Scan# 480
Delta R.T. 0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44

Tgt Ion:	82	Resp:	3844
Ion	Ratio	Lower	Upper
82	100		
128	103.5	17.0	25.4#
54	237.1	53.8	80.6#








#9
Naphthalene
Concen: 0.39 ng
RT: 10.77 min Scan# 581
Delta R.T. 0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44



Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128 Resp: 64063
Ion Ratio Lower Upper
128 100
129 2.8 11.4 17.0#
127 3.5 11.2 16.8#



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 11.08 min Scan# 599
Delta R.T. 0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44

Tgt Ion:225 Resp: 2453
Ion Ratio Lower Upper
225 100
223 63.5 49.7 74.5
227 62.6 50.3 75.5








#11
 2-Methylnaphthalene-d10
 Concen: 0.39 ng
 RT: 12.31 min Scan# 830
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 9758
 Ion Ratio Lower Upper
 152 100
 151 19.3 15.1 22.7



#12
 2-Methylnaphthalene
 Concen: 0.39 ng
 RT: 12.38 min Scan# 849
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:142 Resp: 10749
 Ion Ratio Lower Upper
 142 100
 141 89.2 72.6 108.8
 115 38.2 32.2 48.2








#13
 Acenaphthene-d10
 Concen: 0.40 ng
 RT: 14.54 min Scan# 1113
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:164 Resp: 10251
 Ion Ratio Lower Upper
 164 100
 162 104.7 84.6 127.0
 160 47.7 41.8 62.6



#14
 2,4,6-Tribromophenol
 Concen: 0.32 ng
 RT: 16.03 min Scan# 1206
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:330 Resp: 1044
 Ion Ratio Lower Upper
 330 100
 332 111.1 77.7 116.5
 141 0.0 56.2 84.4#








#15
2-Fluorobiphenyl
Concen: 0.39 ng
RT: 13.18 min Scan# 1021
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44



Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 15330
Ion Ratio Lower Upper
172 100
171 34.6 29.8 44.6
170 23.0 20.7 31.1



#16
Acenaphthylene
Concen: 0.36 ng
RT: 14.26 min Scan# 1094
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44

Tgt Ion:152 Resp: 15470
Ion Ratio Lower Upper
152 100
151 19.4 4.0 6.0#
153 12.9 10.3 15.5








#17
Acenaphthene
Concen: 0.38 ng
RT: 14.60 min Scan# 1117
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44



Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 11909
Ion Ratio Lower Upper
154 100
153 114.4 91.2 136.8
152 55.2 44.8 67.2



#18
Fluorene
Concen: 0.38 ng
RT: 15.59 min Scan# 1184
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44

Tgt Ion:166 Resp: 16471
Ion Ratio Lower Upper
166 100
165 98.9 78.6 117.8
167 54.9 11.1 16.7#








#19
 Phenanthrene-d10
 Concen: 0.40 ng
 RT: 17.27 min Scan# 1244
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:188 Resp: 29134
 Ion Ratio Lower Upper
 188 100
 94 3.3 11.3 16.9#
 80 2.2 11.7 17.5#



#20
 4-Bromophenyl-phenylether
 Concen: 0.40 ng
 RT: 16.45 min Scan# 1219
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:248 Resp: 6467
 Ion Ratio Lower Upper
 248 100
 250 71.6 40.8 61.2#
 141 72.3 161.6 242.4#








#21
Hexachlorobenzene
Concen: 0.39 ng
RT: 16.58 min Scan# 1223
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44



Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:284 Resp: 6748
Ion Ratio Lower Upper
284 100
142 0.0 0.0 0.0
249 24.3 0.0 0.0#



#22
Pentachlorophenol
Concen: 0.33 ng
RT: 16.91 min Scan# 1233
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44

Tgt Ion:266 Resp: 755
Ion Ratio Lower Upper
266 100
264 80.9 58.2 87.2
268 60.5 71.9 107.9#








#23
Phenanthrene
Concen: 0.40 ng
RT: 17.30 min Scan# 1245
Delta R.T. -0.00 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44



Instrument :
BNA_E
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 42375
Ion Ratio Lower Upper
178 100
176 17.9 15.0 22.4
179 15.8 12.7 19.1



#24
Anthracene
Concen: 0.66 ng
RT: 17.30 min Scan# 1245
Delta R.T. -0.10 min
Lab File: BE093418.D
Acq: 2 Jul 2017 17:44

Tgt Ion:178 Resp: 43126
Ion Ratio Lower Upper
178 100
176 18.1 14.6 22.0
179 15.4 12.0 18.0





#25

Fluoranthene-d10

Concen: 0.39 ng

RT: 19.29 min Scan# 1434

Delta R.T. 0.00 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4




Tgt Ion:212 Resp: 136380

Ion Ratio Lower Upper

212 100

106 3.5 2.4 3.6

104 2.0 1.6 2.4



#26

Fluoranthene

Concen: 0.39 ng

RT: 19.32 min Scan# 1438

Delta R.T. 0.00 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44



Tgt Ion:202 Resp: 39306


Ion Ratio Lower Upper

202 100

101 12.0 9.6 14.4

203 16.8 14.4 21.6








#27
 Chrysene-d12
 Concen: 0.40 ng
 RT: 21.43 min Scan# 1656
 Delta R.T. 0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:240 Resp: 36091
 Ion Ratio Lower Upper
 240 100
 120 7.3 4.6 7.0#
 236 25.3 20.8 31.2



#28
 Pyrene
 Concen: 0.39 ng
 RT: 19.67 min Scan# 1486
 Delta R.T. -0.01 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:202 Resp: 39839
 Ion Ratio Lower Upper
 202 100
 200 20.7 0.0 0.0#
 203 17.7 13.7 20.5








#29
 Terphenyl-d14
 Concen: 0.39 ng
 RT: 19.87 min Scan# 1513
 Delta R.T. -0.01 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 25290
 Ion Ratio Lower Upper
 244 100
 212 35.8 10.6 16.0#
 122 15.0 15.4 23.0#



#30
 Benzo(a)anthracene
 Concen: 0.38 ng
 RT: 21.40 min Scan# 1654
 Delta R.T. 0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:228 Resp: 37262
 Ion Ratio Lower Upper
 228 100
 226 27.0 19.7 29.5
 229 19.5 19.2 28.8





#31

Chrysene

Concen: 0.39 ng

RT: 21.46 min Scan# 1659

Delta R.T. 0.00 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4




Tgt Ion:228 Resp: 39499

Ion Ratio Lower Upper

228 100

226 29.7 21.5 32.3

229 19.7 18.6 28.0



#32

Bis(2-ethylhexyl)phthalate

Concen: 0.33 ng

RT: 21.33 min Scan# 1647

Delta R.T. -0.01 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44



Tgt Ion:149 Resp: 46596


Ion Ratio Lower Upper

149 100

167 6.7 20.1 30.1#

279 1.0 2.2 3.4#








#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.39 ng
 RT: 26.58 min Scan# 2705
 Delta R.T. -0.01 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:276 Resp: 34924
 Ion Ratio Lower Upper
 276 100
 138 29.6 27.4 41.2
 277 25.1 20.1 30.1



#34
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.94 min Scan# 2128
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:264 Resp: 29449
 Ion Ratio Lower Upper
 264 100
 260 24.0 21.0 31.4
 265 29.4 24.6 36.8








#35
 Benzo(b)fluoranthene
 Concen: 0.38 ng
 RT: 23.16 min Scan# 1958
 Delta R.T. -0.00 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44



Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:252 Resp: 36241
 Ion Ratio Lower Upper
 252 100
 253 23.5 18.5 27.7
 125 13.6 13.4 20.0



#36
 Benzo(k)fluoranthene
 Concen: 0.40 ng
 RT: 23.21 min Scan# 1969
 Delta R.T. -0.01 min
 Lab File: BE093418.D
 Acq: 2 Jul 2017 17:44

Tgt Ion:252 Resp: 37561
 Ion Ratio Lower Upper
 252 100
 253 23.4 20.3 30.5
 125 12.3 12.2 18.4





#37

Benzo(a)pyrene

Concen: 0.38 ng

RT: 23.82 min Scan# 2102

Delta R.T. -0.01 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4




Tgt Ion:252 Resp: 31472

Ion Ratio Lower Upper

252 100

253 25.3 19.9 29.9

125 14.5 14.6 21.8#



#38

Dibenzo(a,h)anthracene

Concen: 0.39 ng

RT: 26.60 min Scan# 2710

Delta R.T. -0.01 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44



Tgt Ion:278 Resp: 31578


Ion Ratio Lower Upper

278 100

139 19.9 21.4 32.0#

279 24.7 22.2 33.2





#39

Benzo(g,h,i)perylene

Concen: 0.39 ng

RT: 27.40 min Scan# 2884

Delta R.T. -0.00 min

Lab File: BE093418.D

Acq: 2 Jul 2017 17:44

Instrument :

BNA_E

ClientSampleId :

SSTDCCC0.4

Tgt Ion:276 Resp: 31889

Ion Ratio Lower Upper

276 100

277 24.3 21.2 31.8

138 25.1 24.5 36.7

