

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070217\
 Data File : BE093411.D
 Acq On : 2 Jul 2017 13:21
 Operator : SJ/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE070217

Quant Time: Jul 02 13:55:52 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jul 02 13:43:38 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.847	0.751	11.3	88	0.00
3	n-Nitrosodimethylamine	0.444	0.396	10.8	111	0.00
4 S	2-Fluorophenol	1.207	1.159	4.0	97	0.00
5 S	Phenol-d6	1.780	1.725	3.1	102	0.00
6	bis(2-Chloroethyl)ether	1.339	1.293	3.4	98	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	0.00
8 S	Nitrobenzene-d5	0.256	0.251	2.0	105	0.00
9	Naphthalene	4.190	4.144	1.1	101	0.00
10	Hexachlorobutadiene	0.159	0.162	-1.9	103	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.625	2.0	100	0.00
12	2-Methylnaphthalene	0.704	0.684	2.8	100	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.126	0.107	15.1	92	0.00
15 S	2-Fluorobiphenyl	1.553	1.551	0.1	103	0.00
16	Acenaphthylene	1.666	1.527	8.3	96	0.00
17	Acenaphthene	1.220	1.173	3.9	99	0.00
18	Fluorene	1.675	1.616	3.5	99	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.219	0.224	-2.3	100	0.00
21	Hexachlorobenzene	0.237	0.234	1.3	100	0.00
22	Pentachlorophenol	0.035	0.030	14.3	94	0.00
23	Phenanthrene	1.454	1.462	-0.6	99	0.00
24	Anthracene	0.891	0.852	4.4	91	0.00
25 SURR	Fluoranthene-d10	4.823	4.757	1.4	97	0.00
26	Fluoranthene	1.367	1.349	1.3	98	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	98	0.00
28	Pyrene	1.124	1.091	2.9	97	0.00
29 S	Terphenyl-d14	0.714	0.698	2.2	97	0.00
30	Benzo(a)anthracene	1.080	1.026	5.0	98	0.00
31	Chrysene	1.130	1.105	2.2	97	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.318	15.2	101	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	0.961	3.1	98	0.00
34 I	Perylene-d12	1.000	1.000	0.0	97	0.00
35	Benzo(b)fluoranthene	1.292	1.231	4.7	95	0.00
36	Benzo(k)fluoranthene	1.278	1.231	3.7	98	-0.05
37 C	Benzo(a)pyrene	1.132	1.065	5.9	97	0.00
38	Dibenzo(a,h)anthracene	1.099	1.064	3.2	98	0.00
39	Benzo(g,h,i)perylene	1.110	1.075	3.2	99	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070217\
Data File : BE093411.D
Acq On : 2 Jul 2017 13:21
Operator : SJ/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE070217

Quant Time: Jul 02 13:55:52 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jul 02 13:43:38 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			