

Data Path : Z:\HPCHEM1\BNA E\DATA\BE070317\
 Data File : BE093421.D
 Acq On : 3 Jul 2017 9:57
 Operator : SJ/JU
 Sample : I3842-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Jul 03 15:05:09 2017
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.94	152	3278	0.40	ng	0.00
7) Naphthalene-d8	10.72	136	15964	0.40	ng	0.00
13) Acenaphthene-d10	14.54	164	10272	0.40	ng	0.00
19) Phenanthrene-d10	17.27	188	30314	0.40	ng	0.00
27) Chrysene-d12	21.43	240	36743	0.40	ng	0.00
34) Perylene-d12	23.94	264	31418	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.53	112	3868	0.39	ng	0.00
5) Phenol-d6	7.10	99	6040	0.41	ng	0.00
8) Nitrobenzene-d5	9.08	82	4000	0.39	ng	0.00
11) 2-Methylnaphthalene-d10	12.31	152	8203	0.32	ng	0.00
14) 2,4,6-Tribromophenol	16.03	330	1146	0.35	ng	0.00
15) 2-Fluorobiphenyl	13.18	172	12400	0.31	ng	0.00
25) Fluoranthene-d10	19.29	212	95187	0.26	ng	0.00
29) Terphenyl-d14	19.88	244	17168	0.26	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) n-Nitrosodimethylamine	3.68	42	709	0.26	ng	# 1
9) Naphthalene	10.77	128	14628	0.09	ng	# 75
12) 2-Methylnaphthalene	12.38	142	2037	0.07	ng	96
16) Acenaphthylene	14.26	152	10233	0.24	ng	# 86
17) Acenaphthene	14.60	154	3238	0.10	ng	97
18) Fluorene	15.59	166	7852	0.18	ng	# 80
23) Phenanthrene	17.30	178	170399	1.55	ng	98
24) Anthracene	17.40	178	16338	0.24	ng	95
26) Fluoranthene	19.32	202	230846	2.23	ng	99
28) Pvrene	19.67	202	241339	2.34	ng	# 91
30) Benzo(a)anthracene	21.40	228	115778	1.17	ng	93
31) Chrysene	21.46	228	120320	1.16	ng	96
32) Bis(2-ethylhexyl)phthalate	21.33	149	315680	2.21	ng	# 66
33) Indeno(1,2,3-cd)pyrene	26.59	276	61181	0.67	ng	# 91
35) Benzo(b)fluoranthene	23.17	252	131617	1.30	ng	95
36) Benzo(k)fluoranthene	23.17	252	131617	1.31	ng	# 96
37) Benzo(a)pvrene	23.82	252	94997	1.07	ng	96
38) Dibenzo(a,h)anthracene	26.60	278	16770	0.19	ng	# 87
39) Benzo(g,h,i)perylene	27.40	276	65094	0.75	ng	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE070317\
Data File : BE093421.D
Acq On : 3 Jul 2017 9:57
Operator : SJ/JU
Sample : I3842-01
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Jul 03 15:05:09 2017
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 0.40 ng

RT: 7.94 min Scan# 404

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :


Tot Ion:152 Resp: 3278

Ion Ratio Lower Upper

152 100

150 160.9 125.1 187.7




115 71.2 55.7 83.5



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#3

n-Nitrosodimethylamine

Concen: 0.26 ng

RT: 3.68 min Scan# 65

Delta R.T. -0.09 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Tgt Ion: 42 Resp: 709

Ion Ratio Lower Upper

42 100

74 1.1 99.1 148.7#



44 164.0 7.4 11.2#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0





#4

2-Fluorophenol

Concen: 0.39 ng

RT: 5.53 min Scan# 212

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :


Tot Ion:112 Resp: 3868

Ion Ratio Lower Upper

112 100

64 70.5 56.4 84.6




63 149.4 29.3 43.9#



Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#5

Phenol-d6

Concen: 0.41 ng

RT: 7.10 min Scan# 337

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Tgt Ion: 99 Resp: 6040

Ion Ratio Lower Upper

99 100

42 25.4 0.0 0.0#



71 35.7 0.0 0.0#




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#7

Naphthalene-d8

Concen: 0.40 ng

RT: 10.72 min Scan# 578

Delta R.T. 0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :

Tot Ion:136 Resp: 15964



Ion Ratio Lower Upper

136 100

137 11.3 9.8 14.8

54 45.6 10.3 15.5#

68 9.3 9.0 13.4





Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0

Ion 68.00 (67.70 to 68.70): BE0



#8

Nitrobenzene-d5

Concen: 0.39 ng

RT: 9.08 min Scan# 480

Delta R.T. 0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57



Tgt Ion: 82 Resp: 4000

Ion Ratio Lower Upper

82 100

128 106.0 17.0 25.4#


54 236.9 53.8 80.6#




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0









#9
Naphthalene
Concen: 0.09 ng
RT: 10.77 min Scan# 581
Delta R.T. 0.00 min
Lab File: BE093421.D
Acq: 3 Jul 2017 9:57




Instrument :
BNA_E
ClientSampleId :


Tot Ion:128 Resp: 14628
Ion Ratio Lower Upper
128 100
129 3.4 11.4 17.0#
127 4.3 11.2 16.8#



#11
2-Methylnaphthalene-d10
Concen: 0.32 ng
RT: 12.31 min Scan# 829
Delta R.T. -0.00 min
Lab File: BE093421.D
Acq: 3 Jul 2017 9:57

Tgt Ion:152 Resp: 8203
Ion Ratio Lower Upper
152 100
151 19.0 15.1 22.7





#12

2-Methylnaphthalene

Concen: 0.07 ng

RT: 12.38 min Scan# 848

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :





Tot Ion:142 Resp: 2037

Ion Ratio Lower Upper

142 100

141 93.3 72.6 108.8

115 43.6 32.2 48.2



#13

Acenaphthene-d10

Concen: 0.40 ng

RT: 14.54 min Scan# 1112

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57




Tgt Ion:164 Resp: 10272


Ion Ratio Lower Upper

164 100

162 106.5 84.6 127.0

160 48.5 41.8 62.6





#14

2,4,6-Tribromophenol

Concen: 0.35 ng

RT: 16.03 min Scan# 1205

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :


Tot Ion:330 Resp: 1146

Ion Ratio Lower Upper




330 100

332 108.7 77.7 116.5

141 0.0 56.2 84.4#



Abundance Ion 330.00 (329.70 to 330.70): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E



#15

2-Fluorobiphenyl

Concen: 0.31 ng

RT: 13.18 min Scan# 1020

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Tgt Ion:172 Resp: 12400

Ion Ratio Lower Upper



172 100


171 33.7 29.8 44.6

170 22.4 20.7 31.1



Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E





#16

Acenaphthylene

Concen: 0.24 ng

RT: 14.26 min Scan# 1093

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :


Tot Ion:152 Resp: 10233

Ion Ratio Lower Upper

152 100

151 20.1 4.0 6.0#




153 13.5 10.3 15.5



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#17

Acenaphthene

Concen: 0.10 ng

RT: 14.60 min Scan# 1116

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Tgt Ion:154 Resp: 3238

Ion Ratio Lower Upper

154 100

153 111.1 91.2 136.8



152 54.3 44.8 67.2




Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E





#18

Fluorene

Concen: 0.18 ng

RT: 15.59 min Scan# 1183

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :


Tot Ion:166 Resp: 7852

Ion Ratio Lower Upper

166 100

165 104.0 78.6 117.8




167 62.1 11.1 16.7#



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#19

Phenanthrene-d10

Concen: 0.40 ng

RT: 17.27 min Scan# 1243

Delta R.T. -0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Tgt Ion:188 Resp: 30314

Ion Ratio Lower Upper

188 100

94 3.9 11.3 16.9#



80 3.1 11.7 17.5#




Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0









#23
Phenanthrene
Concen: 1.55 ng
RT: 17.30 min Scan# 1244
Delta R.T. -0.00 min
Lab File: BE093421.D
Acq: 3 Jul 2017 9:57

Instrument :
BNA_E
ClientSampled :

Tot Ion:178 Resp: 170399
Ion Ratio Lower Upper
178 100
176 17.6 15.0 22.4
179 16.1 12.7 19.1




Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E






#24
Anthracene
Concen: 0.24 ng
RT: 17.40 min Scan# 1247
Delta R.T. -0.00 min
Lab File: BE093421.D
Acq: 3 Jul 2017 9:57

Tgt Ion:178 Resp: 16338
Ion Ratio Lower Upper
178 100
176 15.0 14.6 22.0
179 13.8 12.0 18.0



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 176.00 (175.70 to 176.70): E
Ion 179.00 (178.70 to 179.70): E





#25

Fluoranthene-d10

Concen: 0.26 ng

RT: 19.29 min Scan# 1433

Delta R.T. 0.00 min

Lab File: BE093421.D





Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :

Tot Ion:212	Resp:	95187
Ion Ratio	Lower	Upper
212 100		
106 4.0	2.4	3.6#
104 2.5	1.6	2.4#



#26

Fluoranthene

Concen: 2.23 ng




RT: 19.32 min Scan# 1437


Delta R.T. 0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Tgt Ion:202	Resp:	230846
Ion Ratio	Lower	Upper
202 100		
101 12.7	9.6	14.4
203 17.4	14.4	21.6









#27
 Chrysene-d12
 Concen: 0.40 ng
 RT: 21.43 min Scan# 1655
 Delta R.T. 0.00 min
 Lab File: BE093421.D
 Acq: 3 Jul 2017 9:57




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:240 Resp: 36743
 Ion Ratio Lower Upper
 240 100
 120 10.4 4.6 7.0#
 236 26.9 20.8 31.2



#28
 Pyrene
 Concen: 2.34 ng
 RT: 19.67 min Scan# 1485
 Delta R.T. -0.01 min
 Lab File: BE093421.D
 Acq: 3 Jul 2017 9:57

Tgt Ion:202 Resp: 241339
 Ion Ratio Lower Upper
 202 100
 200 21.1 0.0 0.0#
 203 21.1 13.7 20.5#





#29

Terphenyl-d14

Concen: 0.26 ng

RT: 19.88 min Scan# 1513

Delta R.T. 0.00 min


Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampled :






Tot Ion:244 Resp: 17168

Ion Ratio Lower Upper

244 100

212 35.8 10.6 16.0#

122 12.5 15.4 23.0#



#30

Benzo(a)anthracene

Concen: 1.17 ng

RT: 21.40 min Scan# 1653

Delta R.T. 0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57




Tgt Ion:228 Resp: 115778


Ion Ratio Lower Upper

228 100

226 28.7 19.7 29.5

229 20.9 19.2 28.8





#31

Chrysene

Concen: 1.16 ng

RT: 21.46 min Scan# 1658

Delta R.T. 0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Instrument :

BNA_E

ClientSampleId :

Tot Ion:228 Resp: 120320




Ion	Ratio	Lower	Upper
228	100		
226	30.3	21.5	32.3
229	22.6	18.6	28.0



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#32

Bis(2-ethylhexyl)phthalate

Concen: 2.21 ng

RT: 21.33 min Scan# 1646


Delta R.T. -0.01 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Tgt Ion:149 Resp: 315680



Ion	Ratio	Lower	Upper
149	100		
167	6.7	20.1	30.1#
279	0.9	2.2	3.4#




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.67 ng
 RT: 26.59 min Scan# 2705
 Delta R.T. -0.01 min
 Lab File: BE093421.D
 Acq: 3 Jul 2017 9:57




Instrument :
 BNA_E
 ClientSampleId :


Tot Ion:276 Resp: 61181
 Ion Ratio Lower Upper
 276 100
 138 25.6 27.4 41.2#
 277 24.7 20.1 30.1



#34
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.94 min Scan# 2127
 Delta R.T. -0.00 min
 Lab File: BE093421.D
 Acq: 3 Jul 2017 9:57

Tgt Ion:264 Resp: 31418
 Ion Ratio Lower Upper
 264 100
 260 26.0 21.0 31.4
 265 31.6 24.6 36.8





#35

Benzo(b)fluoranthene

Concen: 1.30 ng

RT: 23.17 min Scan# 1958

Delta R.T. 0.00 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :


Tot Ion:252 Resp: 131617

Ion Ratio Lower Upper




252 100

253 25.0 18.5 27.7

125 19.5 13.4 20.0



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#36

Benzo(k)fluoranthene

Concen: 1.31 ng

RT: 23.17 min Scan# 1958

Delta R.T. -0.05 min

Lab File: BE093421.D

Acq: 3 Jul 2017 9:57


Tgt Ion:252 Resp: 131617

Ion Ratio Lower Upper



252 100


253 25.0 20.3 30.5

125 19.5 12.2 18.4#



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E









#37
Benzo(a)pyrene
Concen: 1.07 ng
RT: 23.82 min Scan# 2102
Delta R.T. -0.00 min
Lab File: BE093421.D
Acq: 3 Jul 2017 9:57




Instrument :
BNA_E
ClientSampleId :


Tot Ion:252 Resp: 94997
Ion Ratio Lower Upper
252 100
253 24.2 19.9 29.9
125 14.8 14.6 21.8



#38
Dibenzo(a,h)anthracene
Concen: 0.19 ng
RT: 26.60 min Scan# 2708
Delta R.T. -0.01 min
Lab File: BE093421.D
Acq: 3 Jul 2017 9:57

Tgt Ion:278 Resp: 16770
Ion Ratio Lower Upper
278 100
139 34.0 21.4 32.0#
279 34.2 22.2 33.2#





#39

Benzo(a,h,i)perylene

Concen: 0.75 ng

RT: 27.40 min Scan# 2884

Delta R.T. 0.00 min


Lab File: BE093421.D

Acq: 3 Jul 2017 9:57

Instrument :

BNA_E

ClientSampleId :




Tot Ion:276 Resp: 65094

Ion Ratio Lower Upper

276 100

277 24.8 21.2 31.8

138 25.9 24.5 36.7



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

