

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070317\
 Data File : BE093420.D
 Acq On : 3 Jul 2017 9:18
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 04 04:27:41 2017
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 03 06:40:23 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
2	1,4-Dioxane	0.847	0.792	6.5	96	0.00
3	n-Nitrosodimethylamine	0.444	0.383	13.7	111	0.00
4 S	2-Fluorophenol	1.207	1.127	6.6	98	0.00
5 S	Phenol-d6	1.780	1.616	9.2	99	0.00
6	bis(2-Chloroethyl)ether	1.339	1.272	5.0	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.256	0.243	5.1	100	0.00
9	Naphthalene	4.190	4.106	2.0	98	0.00
10	Hexachlorobutadiene	0.159	0.153	3.8	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.638	0.597	6.4	94	0.00
12	2-Methylnaphthalene	0.704	0.656	6.8	94	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.126	0.093	26.2#	70	0.00
15 S	2-Fluorobiphenyl	1.553	1.577	-1.5	90	0.00
16	Acenaphthylene	1.666	1.520	8.8	83	0.00
17	Acenaphthene	1.220	1.166	4.4	86	0.00
18	Fluorene	1.675	1.588	5.2	84	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
20	4-Bromophenyl-phenylether	0.219	0.211	3.7	85	0.00
21	Hexachlorobenzene	0.237	0.218	8.0	84	0.00
22	Pentachlorophenol	0.035	0.023	34.3#	65	0.00
23	Phenanthrene	1.454	1.380	5.1	85	0.00
24	Anthracene	0.891	0.794	10.9	77	0.00
25 SURR	Fluoranthene-d10	4.823	4.670	3.2	87	0.00
26	Fluoranthene	1.367	1.335	2.3	88	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
28	Pyrene	1.124	1.017	9.5	88	0.00
29 S	Terphenyl-d14	0.714	0.660	7.6	89	0.00
30	Benzo(a)anthracene	1.080	1.022	5.4	95	0.00
31	Chrysene	1.130	1.093	3.3	93	0.00
32	Bis(2-ethylhexyl)phthalate	1.555	1.301	16.3	97	-0.01
33	Indeno(1,2,3-cd)pyrene	0.992	1.047	-5.5	103	-0.01
34 I	Perylene-d12	1.000	1.000	0.0	100	0.00
35	Benzo(b)fluoranthene	1.292	1.212	6.2	96	0.00
36	Benzo(k)fluoranthene	1.278	1.256	1.7	103	0.00
37 C	Benzo(a)pyrene	1.132	1.066	5.8	99	0.00
38	Dibenzo(a,h)anthracene	1.099	1.093	0.5	103	0.00
39	Benzo(g,h,i)perylene	1.110	1.100	0.9	104	0.00

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070317\
Data File : BE093420.D
Acq On : 3 Jul 2017 9:18
Operator : SJ/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 04 04:27:41 2017
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\8270-SIM-BE070217.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 03 06:40:23 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		