

Data Path : Z:\HPCHEM1\BNA E\DATA\BE070516\
 Data File : BE091980.D
 Acq On : 6 Jul 2016 2:48
 Operator : UM/SJ
 Sample : H3826-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 KY038PS037-160627

Quant Time: Jul 06 06:50:03 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 05 16:14:58 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	21905	5.00	ng	0.00
8) Naphthalene-d8	10.99	136	102065	5.00	ng	0.00
15) Acenaphthene-d10	14.86	164	78245	5.00	ng	0.00
24) Phenanthrene-d10	17.59	188	168401	5.00	ng	0.00
31) Chrysene-d12	21.76	240	301499	5.00	ng	0.00
40) Perylene-d12	24.18	264	211718	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.72	112	6635	1.27	ng	0.00
5) Phenol-d6	7.37	99	5439	0.75	ng	0.00
9) Nitrobenzene-d5	9.37	82	19276	2.77	ng	-0.02
16) 2,4,6-Tribromophenol	16.35	330	9544	3.49	ng	0.00
17) 2-Fluorobiphenyl	13.49	172	54860	2.87	ng	0.00
34) Terphenyl-d14	20.22	244	112759	3.17	ng	0.00
Target Compounds						
38) Bis(2-ethylhexyl)phthalate	21.64	149	13543	0.23	ng	Qvalue # 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE070516\
Data File : BE091980.D
Acq On : 6 Jul 2016 2:48
Operator : UM/SJ
Sample : H3826-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
KY038PS037-160627

Quant Time: Jul 06 06:50:03 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 05 16:14:58 2016
Response via : Initial Calibration

