

Data Path : Z:\HPCHEM1\BNA E\DATA\BE070516\
 Data File : BE091964.D
 Acq On : 5 Jul 2016 12:38
 Operator : UM/SJ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Manual Integrations

umangi
 7/6/2016 7:19:19 PM

Quant Time: Jul 05 13:35:14 2016

Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Jul 05 12:36:52 2016

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.20	152	23847	5.00	ng	0.00
8) Naphthalene-d8	10.99	136	119093	5.00	ng	0.00
15) Acenaphthene-d10	14.86	164	89525	5.00	ng	0.00
24) Phenanthrene-d10	17.60	188	181460	5.00	ng	0.00
31) Chrysene-d12	21.76	240	303004	5.00	ng	0.00
40) Perylene-d12	24.18	264	216934	5.00	ng	0.00

System Monitoring Compounds

4) 2-Fluorophenol	5.72	112	31116	6.67	ng	0.00
5) Phenol-d6	7.35	99	44946	7.44	ng	-0.02
9) Nitrobenzene-d5	9.36	82	43663	5.31	ng	-0.02
16) 2,4,6-Tribromophenol	16.35	330	15881	8.70	ng	0.00
17) 2-Fluorobiphenyl	13.49	172	110502	4.86	ng	0.00
34) Terphenyl-d14	20.22	244	168097	4.53	ng	0.00

Target Compounds

						Qvalue
2) 1,4-Dioxane	3.47	88	14541	4.32	ng	90
3) n-Nitrosodimethylamine	3.80	42	22161	5.80	ng	84
6) bis(2-Chloroethyl)ether	7.61	93	37779	5.53	ng	98
7) n-Nitroso-di-n-propylamine	9.00	70	36355	6.70	ng	# 99
10) Nitrobenzene	9.40	77	49611	6.07	ng	# 100
11) bis(2-Chloroethoxy)methane	10.40	93	50651	4.86	ng	# 1
12) Naphthalene	11.04	128	133881	5.15	ng	97
13) Hexachlorobutadiene	11.34	225	19495	4.63	ng	98
14) 2-Methylnaphthalene	12.66	142	88677	5.03	ng	95
18) Acenaphthylene	14.56	152	637633	7.02	ng	# 72
19) 2,6-Dinitrotoluene	14.44	165	135691	9.35	ng	# 10
20) Acenaphthene	14.92	154	114606	4.58	ng	# 1
21) 2,4-Dinitrotoluene	15.22	165	159688m	11.77	ng	
22) Fluorene	15.90	166	131943	5.30	ng	99
23) 4-Chlorophenyl-phenylether	15.90	204	63753	5.20	ng	98
25) 4-Bromophenyl-phenylether	16.78	248	39476	5.63	ng	99
26) Hexachlorobenzene	16.90	284	38294	5.32	ng	99
27) Pentachlorophenol	17.25	266	18516	9.39	ng	# 88
28) Phenanthrene	17.63	178	228592	5.04	ng	99
29) Anthracene	17.72	178	227186	5.66	ng	100
30) Fluoranthene	19.65	202	1008992	5.19	ng	# 87
32) Benzidine	19.83	184	107524m	10.46	ng	
33) Pyrene	20.02	202	1074017	5.21	ng	# 75
35) Benzo(a)anthracene	21.73	228	248935m	5.23	ng	
36) 3,3'-Dichlorobenzidine	21.67	252	65302	8.47	ng	# 58
37) Chrysene	21.79	228	346120m	4.98	ng	
38) Bis(2-ethylhexyl)phthalate	21.64	149	285240	8.11	ng	# 100
39) Indeno(1,2,3-cd)pyrene	26.70	276	1257484	5.29	ng	99
41) Benzo(b)fluoranthene	23.43	252	305814m	5.49	ng	
42) Benzo(k)fluoranthene	23.48	252	280122	5.50	ng	97
43) Benzo(a)pyrene	24.06	252	281243	5.78	ng	95
44) Dibenzo(a,h)anthracene	26.72	278	261655	5.53	ng	97
45) Benzo(a,h,i)perylene	27.47	276	1051844	5.38	ng	99

Data Path : Z:\HPCHEM1\BNA E\DATA\BE070516\
Data File : BE091964.D
Acq On : 5 Jul 2016 12:38
Operator : UM/SJ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Manual Integrations

APPROVED
umangi
7/6/2016 7:19:19 PM

Quant Time: Jul 05 13:35:14 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 05 12:36:52 2016
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA E\DATA\BE070516\
Data File : BE091964.D
Acq On : 5 Jul 2016 12:38
Operator : UM/SJ
Sample : SSTDIC005
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
SSTDICC005

Quant Time: Jul 05 13:35:14 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE070516.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 05 12:36:52 2016
Response via : Initial Calibration

Manual Integrations

umangi
7/6/2016 7:19:19 PM





#1





1,4-Dichlorobenzene-d4
Concen: 5.00 ng
RT: 8.20 min Scan# 323
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
152	100		
150	157.4	123.9	185.9
115	60.4	48.3	72.5

Manual Integrations




umangi
7/6/2016 7:19:19 PM




#2

1,4-Dioxane
Concen: 4.32 ng
RT: 3.47 min Scan# 51
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
88	100		
58	90.8	63.0	94.4
43	37.4	29.1	43.7





#3





n-Nitrosodimethylamine
Concen: 5.80 ng
RT: 3.80 min Scan# 70
Delta R.T. -0.04 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
42	100		
74	98.0	93.4	140.0
44	6.4	5.0	7.6

Manual Integrations




umangi
7/6/2016 7:19:19 PM




#4

2-Fluorophenol
Concen: 6.67 ng
RT: 5.72 min Scan# 181
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
112	100		
64	66.6	53.7	80.5
63	35.8	29.5	44.3





#5

Phenol-d6

Concen: 7.44 ng

RT: 7.35 min Scan# 274

Delta R.T. -0.02 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion: 99 Resp: 44946

Ion Ratio Lower Upper

99 100





42 25.0 19.6 29.4

71 35.6 28.1 42.1

Manual Integrations

umangi

7/6/2016 7:19:19 PM



#6

bis(2-Chloroethyl)ether

Concen: 5.53 ng

RT: 7.61 min Scan# 289

Delta R.T. -0.02 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38




Tgt Ion: 93 Resp: 37779




Ion Ratio Lower Upper

93 100

63 84.3 66.2 99.4

95 32.6 25.5 38.3





#7

n-Nitroso-di-n-propylamine

Concen: 6.70 ng

RT: 9.00 min Scan# 365

Delta R.T. -0.02 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion: 70 Resp: 36355

Ion Ratio Lower Upper

70 100

42 71.8 58.4 87.6





101 8.2 6.4 9.6

130 16.7 0.0 0.0#

Manual Integrations

umangi

7/6/2016 7:19:19 PM



#8

Naphthalene-d8

Concen: 5.00 ng

RT: 10.99 min Scan# 456

Delta R.T. -0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Tgt Ion: 136 Resp: 119093


Ion Ratio Lower Upper


136 100

137 10.2 8.3 12.5

54 7.5 8.2 12.4#

68 5.8 5.9 8.9#





#9

Nitrobenzene-d5

Concen: 5.31 ng

RT: 9.36 min Scan# 384

Delta R.T. -0.02 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDICC005

Tot Ion: 82 Resp: 43663

Ion Ratio Lower Upper

82 100


128 39.2 39.5 59.3#

54 38.2 30.3 45.5

Manual Integrations

umangi




7/6/2016 7:19:19 PM



Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): BE0

Ion 54.00 (53.70 to 54.70): BE0



#10

Nitrobenzene

Concen: 6.07 ng

RT: 9.40 min Scan# 386

Delta R.T. -0.02 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38


Tgt Ion: 77 Resp: 49611

Ion Ratio Lower Upper

77 100

123 39.0 0.0 0.0#



65 14.0 11.1 16.7






Abundance Ion 77.00 (76.70 to 77.70): BE0

Ion 123.00 (122.70 to 123.70): BE0

Ion 65.00 (64.70 to 65.70): BE0









#11
bis(2-Chloroethoxy)methane
Concen: 4.86 ng
RT: 10.40 min Scan# 432
Delta R.T. -0.05 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
93	100		
95	0.0	57.6	86.4#
123	0.0	100.2	150.4#


Manual Integrations


umangi
7/6/2016 7:19:19 PM



#12
Naphthalene
Concen: 5.15 ng
RT: 11.04 min Scan# 458
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.7	10.4	15.6
127	13.2	10.2	15.2






#13
Hexachlorobutadiene
Concen: 4.63 ng
RT: 11.34 min Scan# 470
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
225	100		
223	61.2	49.8	74.8
227	65.1	50.6	75.8

Manual Integrations

umangi
7/6/2016 7:19:19 PM






Abundance

Ion 225.00 (224.70 to 225.70): E


Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E



#14
2-Methylnaphthalene
Concen: 5.03 ng
RT: 12.66 min Scan# 551
Delta R.T. -0.02 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
142	100		
141	94.9	71.9	107.9
115	40.5	29.9	44.9






Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E









#15
Acenaphthene-d10
Concen: 5.00 ng
RT: 14.86 min Scan# 694
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
164	100		
162	98.7	71.8	107.8
160	40.5	28.0	42.0




Manual Integrations


umangi
7/6/2016 7:19:19 PM



#16
2,4,6-Tribromophenol
Concen: 8.70 ng
RT: 16.35 min Scan# 762
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
330	100		
332	98.4	77.0	115.6
141	172.9	138.4	207.6









#17
2-Fluorobiphenyl
Concen: 4.86 ng
RT: 13.49 min Scan# 622
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
172	100		
171	30.2	24.3	36.5
170	17.7	14.9	22.3




Manual Integrations


umangi
7/6/2016 7:19:19 PM



#18
Acenaphthylene
Concen: 7.02 ng
RT: 14.56 min Scan# 684
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
152	100		
151	5.6	19.4	29.2#
153	11.1	14.4	21.6#









#19
2,6-Dinitrotoluene
Concen: 9.35 ng
RT: 14.44 min Scan# 680
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion	Ratio	Lower	Upper
165	100		
63	0.0	65.9	98.9#
89	0.0	59.1	88.7#




Manual Integrations


umangi
7/6/2016 7:19:19 PM



#20
Acenaphthene
Concen: 4.58 ng
RT: 14.92 min Scan# 696
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
154	100		
153	377.7	88.1	132.1#
152	156.0	44.2	66.2#





#21

2,4-Dinitrotoluene

Concen: 11.77 ng m

RT: 15.22 min Scan# 706

Delta R.T. -0.03 min

Lab File: BE091964.D


Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005



Tot Ion:165 Resp: 159688

Ion Ratio Lower Upper

165 100

63 0.0 0.0 0.0


89 63.5 99.8 149.8#

182 0.4 0.0 0.0#

Manual Integrations

umangi

7/6/2016 7:19:19 PM



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BE0

Ion 89.00 (88.70 to 89.70): BE0

Ion 182.00 (181.70 to 182.70): E

120000

100000

80000

60000


40000

20000

0

15.20 15.30

Time-->



#22

Fluorene

Concen: 5.30 ng

RT: 15.90 min Scan# 736

Delta R.T. 0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38


Tgt Ion:166 Resp: 131943

Ion Ratio Lower Upper

166 100

165 99.0 78.6 117.8

167 13.4 10.8 16.2



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E

120000

100000

80000

60000


40000




20000

0

15.80 16.00

Time-->









#23
4-Chlorophenyl-phenylether
Concen: 5.20 ng
RT: 15.90 min Scan# 736
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampled :
SSTDIC005

Tot Ion:204 Resp: 63753
Ion Ratio Lower Upper
204 100
206 33.2 27.8 41.6
141 243.4 197.6 296.4


Manual Integrations


umangi
7/6/2016 7:19:19 PM



#24
Phenanthrene-d10
Concen: 5.00 ng
RT: 17.60 min Scan# 836
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion:188 Resp: 181460
Ion Ratio Lower Upper
188 100
94 9.4 4.1 6.1#
80 8.9 3.4 5.2#





#25

4-Bromophenyl-phenylether

Concen: 5.63 ng

RT: 16.78 min Scan# 785

Delta R.T. -0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion:248 Resp: 39476

Ion Ratio Lower Upper

248 100





250 95.7 75.7 113.5

141 80.7 64.6 97.0

Manual Integrations

umangi

7/6/2016 7:19:19 PM



#26

Hexachlorobenzene

Concen: 5.32 ng

RT: 16.90 min Scan# 792

Delta R.T. -0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38




Tgt Ion:284 Resp: 38294


Ion Ratio Lower Upper

284 100

142 39.8 31.4 47.2


249 31.9 25.3 37.9





#27
 Pentachlorophenol
 Concen: 9.39 ng
 RT: 17.25 min Scan# 812
 Delta R.T. -0.02 min
 Lab File: BE091964.D
 Acq: 5 Jul 2016 12:38




Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005



Tot Ion	Ratio	Lower	Upper
266	100		
268	58.8	60.5	90.7#
264	66.3	51.0	76.4




Manual Integrations


umangi
 7/6/2016 7:19:19 PM



#28
 Phenanthrene
 Concen: 5.04 ng
 RT: 17.63 min Scan# 839
 Delta R.T. -0.01 min
 Lab File: BE091964.D
 Acq: 5 Jul 2016 12:38

Tgt Ion	Ratio	Lower	Upper
178	100		
176	19.5	15.4	23.2
179	15.3	12.9	19.3





#29

Anthracene

Concen: 5.66 ng

RT: 17.72 min Scan# 847

Delta R.T. -0.01 min

Lab File: BE091964.D


Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005



Tot Ion:178 Resp: 227186

Ion Ratio Lower Upper

178 100




176 19.0 15.1 22.7

179 15.4 12.1 18.1

Manual Integrations

umangi

7/6/2016 7:19:19 PM



#30

Fluoranthene

Concen: 5.19 ng

RT: 19.65 min Scan# 985

Delta R.T. -0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38




Tgt Ion:202 Resp: 1008992


Ion Ratio Lower Upper

202 100

101 0.0 9.4 14.0#

203 18.8 14.3 21.5









#31
Chrysene-d12
Concen: 5.00 ng
RT: 21.76 min Scan# 1065
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion:240 Resp: 303004
Ion Ratio Lower Upper
240 100
120 1.9 1.2 1.8#
236 27.9 19.8 29.8




Manual Integrations


umangi
7/6/2016 7:19:19 PM



#32
Benzidine
Concen: 10.46 ng m
RT: 19.83 min Scan# 993
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion:184 Resp: 107524
Ion Ratio Lower Upper
184 100
185 16.6 11.4 17.2
183 8.8 11.8 17.6#









#33
Pvrene
Concen: 5.21 ng
RT: 20.02 min Scan# 1001
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion:202 Resp: 1074017
Ion Ratio Lower Upper
202 100
200 0.0 16.9 25.3#
203 17.6 14.1 21.1




Manual Integrations


umangi
7/6/2016 7:19:19 PM



#34
Terphenyl-d14
Concen: 4.53 ng
RT: 20.22 min Scan# 1010
Delta R.T. -0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion:244 Resp: 168097
Ion Ratio Lower Upper
244 100
212 6.7 6.5 9.7
122 0.4 2.9 4.3#





#35

Benzo(a)anthracene

Concen: 5.23 ng m

RT: 21.73 min Scan# 1064

Delta R.T. 0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005

Tot Ion:228 Resp: 248935

Ion Ratio Lower Upper

228 100


226 35.2 21.0 31.4#

229 14.8 15.8 23.8#

Manual Integrations

umangi




7/6/2016 7:19:19 PM



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#36

3,3'-Dichlorobenzidine

Concen: 8.47 ng

RT: 21.67 min Scan# 1062

Delta R.T. -0.03 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38


Tgt Ion:252 Resp: 65302

Ion Ratio Lower Upper

252 100

254 41.9 61.5 92.3#



126 26.3 2.5 3.7#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E









#37
Chrysene
Concen: 4.98 ng/ml
RT: 21.79 min Scan# 1066
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005

Tot Ion: 228 Resp: 346120
Ion Ratio Lower Upper
228 100
226 33.7 27.0 40.6
229 17.1 13.8 20.8




Manual Integrations


umangi
7/6/2016 7:19:19 PM



#38
Bis(2-ethylhexyl)phthalate
Concen: 8.11 ng
RT: 21.64 min Scan# 1061
Delta R.T. -0.03 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion: 149 Resp: 285240
Ion Ratio Lower Upper
149 100
167 0.0 0.0 0.0
279 0.0 0.0 0.0






#39
Indeno(1,2,3-cd)pyrene
Concen: 5.29 ng
RT: 26.70 min Scan# 1403
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Instrument :
BNA_E
ClientSampleId :
SSTDIC005




Tot Ion:276 Resp: 1257484
Ion Ratio Lower Upper
276 100
138 25.8 19.7 29.5
277 25.1 20.1 30.1

Manual Integrations

umangi
7/6/2016 7:19:19 PM






Abundance Ion 276.00 (275.70 to 276.70): E
Ion 138.00 (137.70 to 138.70): E
Ion 277.00 (276.70 to 277.70): E




#40
Perylene-d12
Concen: 5.00 ng
RT: 24.18 min Scan# 1233
Delta R.T. 0.00 min
Lab File: BE091964.D
Acq: 5 Jul 2016 12:38

Tgt Ion:264 Resp: 216934
Ion Ratio Lower Upper
264 100
260 24.9 20.3 30.5
265 22.0 17.6 26.4

Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E





#41

Benzo(b)fluoranthene

Concen: 5.49 ng m

RT: 23.43 min Scan# 1169

Delta R.T. 0.00 min

Lab File: BE091964.D


Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :

SSTDIC005





Tot Ion	252	Resp	305814
Ion	Ratio	Lower	Upper
252	100		
253	23.5	17.8	26.6
125	9.2	9.4	14.2#

Manual Integrations

umangi

7/6/2016 7:19:19 PM



#42

Benzo(k)fluoranthene


Concen: 5.50 ng

RT: 23.48 min Scan# 1173



Delta R.T. 0.00 min


Lab File: BE091964.D

Acq: 5 Jul 2016 12:38



Tgt Ion	252	Resp	280122
Ion	Ratio	Lower	Upper
252	100		
253	22.5	19.4	29.2
125	9.5	8.3	12.5





#43

Benzo(a)pyrene

Concen: 5.78 ng

RT: 24.06 min Scan# 1223

Delta R.T. 0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Instrument :

BNA_E

ClientSampleId :





SSTDIC005

Tot Ion:	252	Resp:	281243
Ion Ratio	Lower	Upper	
252	100		
253	21.3	18.9	28.3
125	11.2	10.5	15.7

Manual Integrations

umangi

7/6/2016 7:19:19 PM



#44

Dibenzo(a,h)anthracene

Concen: 5.53 ng




RT: 26.72 min Scan# 1404




Delta R.T. 0.00 min

Lab File: BE091964.D

Acq: 5 Jul 2016 12:38

Tgt Ion:	278	Resp:	261655
Ion Ratio	Lower	Upper	
278	100		
139	14.6	12.6	18.8
279	23.7	20.0	30.0





#45
 Benzo(a,h,i)perylene
 Concen: 5.38 ng
 RT: 27.47 min Scan# 1448
 Delta R.T. 0.00 min
 Lab File: BE091964.D
 Acq: 5 Jul 2016 12:38

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC005

Tot Ion	Ratio	Lower	Upper
276	100		
277	23.4	19.5	29.3
138	22.2	17.7	26.5

Manual Integrations

umangi
 7/6/2016 7:19:19 PM

