

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
 Data File : BE100376.D
 Acq On : 5 Jul 2019 11:57
 Operator : JU/SJ
 Sample : K3560-01
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 EXT-009-SW149-062419

Manual Integrations
 APPROVED

mohammad
 7/8/2019 2:13:30 PM

Quant Time: Jul 08 05:42:10 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	536	0.40	ng	-0.02
7) Naphthalene-d8	10.42	136	1647	0.40	ng	-0.03
13) Acenaphthene-d10	14.30	164	1346	0.40	ng	-0.01
19) Phenanthrene-d10	17.05	188	4657	0.40	ng	0.00
27) Chrysene-d12	21.23	240	6863	0.40	ng	-0.01
34) Perylene-d12	23.65	264	8199	0.40	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.19	112	591	0.43	ng	-0.02
5) Phenol-d6	6.82	99	704	0.39	ng	-0.02
8) Nitrobenzene-d5	8.81	82	511	0.31	ng	-0.01
11) 2-Methylnaphthalene-d10	12.04	152	1163	0.38	ng	-0.01
14) 2,4,6-Tribromophenol	15.81	330	488	0.52	ng	-0.01
15) 2-Fluorobiphenyl	12.93	172	1934	0.37	ng	-0.01
25) Fluoranthene-d10	19.09	212	25789	0.38	ng	-0.01
29) Terphenyl-d14	19.68	244	5038	0.35	ng	-0.02
Target Compounds						
16) Acenaphthylene	14.02	152	193	0.030	ng	# 83
23) Phenanthrene	17.09	178	1754	0.155	ng	98
26) Fluoranthene	19.12	202	4150	0.231	ng	99
28) Pyrene	19.48	202	4993	0.275	ng	98
30) Benzo(a)anthracene	21.22	228	2303	0.102	ng	# 88
31) Chrysene	21.27	228	2980	0.132	ng	# 92
32) Bis(2-ethylhexyl)phthalate	21.14	149	3313	0.046	ng	98
33) Indeno(1,2,3-cd)pyrene	26.21	276	1752	0.058	ng	# 93
35) Benzo(b)fluoranthene	22.92	252	3354	0.151	ng	# 85
36) Benzo(k)fluoranthene	22.95	252	1601m	0.065	ng	
37) Benzo(a)pyrene	23.54	252	2262	0.102	ng	# 70
39) Benzo(a,h,i)perylene	26.99	276	2097	0.088	ng	# 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
Data File : BE100376.D
Acq On : 5 Jul 2019 11:57
Operator : JU/SJ
Sample : K3560-01
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
EXT-009-SW149-062419

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:30 PM

Quant Time: Jul 08 05:42:10 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

