

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\  
 Data File : BE100375.D  
 Acq On : 5 Jul 2019 11:18  
 Operator : JU/SJ  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_E  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Jul 06 00:10:31 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jun 28 14:06:53 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	76	-0.02
2	1,4-Dioxane	0.621	0.667	-7.4	79	-0.02
3	n-Nitrosodimethylamine	0.289	0.407	-40.8#	129	-0.04
4 S	2-Fluorophenol	1.020	1.222	-19.8	90	-0.02
5 S	Phenol-d6	1.342	1.152	14.2	65	0.00
6	bis(2-Chloroethyl)ether	1.102	0.968	12.2	60	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	67	-0.03
8 S	Nitrobenzene-d5	0.397	0.320	19.4	55	-0.01
9	Naphthalene	4.411	4.406	0.1	66	-0.01
10	Hexachlorobutadiene	0.425	0.500	-17.6	76	-0.03
11 SURR	2-Methylnaphthalene-d10	0.752	0.734	2.4	64	-0.01
12	2-Methylnaphthalene	0.774	0.678	12.4	62	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	71	-0.01
14 S	2,4,6-Tribromophenol	0.280	0.267	4.6	69	0.00
15 S	2-Fluorobiphenyl	1.559	1.505	3.5	66	-0.01
16	Acenaphthylene	1.920	1.803	6.1	64	-0.01
17	Acenaphthene	1.088	1.036	4.8	64	-0.01
18	Fluorene	1.609	1.670	-3.8	71	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
20	4-Bromophenyl-phenylether	0.255	0.235	7.8	73	-0.01
21	Hexachlorobenzene	0.265	0.263	0.8	78	-0.01
22	Pentachlorophenol	0.080	0.004	95.0#	5#	0.00
23	Phenanthrene	0.969	0.921	5.0	78	-0.01
24	Anthracene	0.873	0.775	11.2	74	-0.01
25 SURR	Fluoranthene-d10	5.805	5.901	-1.7	83	-0.01
26	Fluoranthene	1.544	1.592	-3.1	83	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	77	-0.01
28	Pyrene	1.059	1.171	-10.6	83	-0.01
29 S	Terphenyl-d14	0.829	0.863	-4.1	79	-0.02
30	Benzo(a)anthracene	1.315	1.192	9.4	68	0.00
31	Chrysene	1.321	1.317	0.3	74	-0.01
32	Bis(2-ethylhexyl)phthalate	2.255	2.021	10.4	71	-0.01
33	Indeno(1,2,3-cd)pyrene	1.747	1.552	11.2	63	0.00
34 I	Perylene-d12	1.000	1.000	0.0	72	0.00
35	Benzo(b)fluoranthene	1.084	0.972	10.3	64	-0.01
36	Benzo(k)fluoranthene	1.211	1.314	-8.5	75	-0.01
37 C	Benzo(a)pyrene	1.087	1.043	4.0	66	0.00
38	Dibenzo(a,h)anthracene	1.135	0.971	14.4	61	0.00
39	Benzo(a,h,i)perylene	1.166	1.208	-3.6	72	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\  
 Data File : BE100375.D  
 Acq On : 5 Jul 2019 11:18  
 Operator : JU/SJ  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

**Instrument :**  
 BNA\_E  
**LabSampleId :**  
 SSTDCCC0.4

Quant Time: Jul 06 00:10:31 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Fri Jun 28 14:06:53 2019  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
-----				
(#) = Out of Range	SPCC's out = 0		CCC's out = 0	