

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
 Data File : BE100392.D
 Acq On : 5 Jul 2019 21:36
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 05 22:41:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	118	-0.02
2	1,4-Dioxane	0.400	0.329	17.8	94	-0.02
3	n-Nitrosodimethylamine	0.400	0.381	4.8	136	-0.03
4 S	2-Fluorophenol	0.400	0.392	2.0	115	-0.02
5 S	Phenol-d6	0.400	0.393	1.8	116	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.364	9.0	97	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	112	-0.03
8 S	Nitrobenzene-d5	0.400	0.320	20.0	91	-0.03
9	Naphthalene	0.400	0.391	2.3	107	-0.01
10	Hexachlorobutadiene	0.400	0.451	-12.7	122	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.386	3.5	106	-0.01
12	2-Methylnaphthalene	0.400	0.355	11.3	104	-0.01
13 I	Acenaphthene-d10	0.400	0.400	0.0	105	-0.01
14 S	2,4,6-Tribromophenol	0.400	0.401	-0.3	107	-0.01
15 S	2-Fluorobiphenyl	0.400	0.403	-0.8	102	-0.01
16	Acenaphthylene	0.400	0.385	3.8	97	-0.01
17	Acenaphthene	0.400	0.389	2.8	96	-0.01
18	Fluorene	0.400	0.385	3.8	97	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	99	0.00
20	4-Bromophenyl-phenylether	0.400	0.423	-5.7	100	-0.03
21	Hexachlorobenzene	0.400	0.433	-8.2	102	0.00
22	Pentachlorophenol	0.400	0.333	16.8	92	-0.03
23	Phenanthrene	0.400	0.383	4.3	94	-0.01
24	Anthracene	0.400	0.389	2.8	97	-0.01
25 SURR	Fluoranthene-d10	0.400	0.376	6.0	92	-0.01
26	Fluoranthene	0.400	0.375	6.3	90	-0.01
27 I	Chrysene-d12	0.400	0.400	0.0	81	-0.01
28	Pvrene	0.400	0.462	-15.5	92	-0.01
29 S	Terphenyl-d14	0.400	0.431	-7.7	87	-0.01
30	Benzo(a)anthracene	0.400	0.377	5.8	75	0.00
31	Chrysene	0.400	0.391	2.3	78	-0.01
32	Bis(2-ethylhexyl)phthalate	0.400	0.445	-11.2	87	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	73	0.00
34 I	Perylene-d12	0.400	0.400	0.0	81	0.00
35	Benzo(b)fluoranthene	0.400	0.359	10.3	72	0.00
36	Benzo(k)fluoranthene	0.400	0.359	10.3	70	0.00
37 C	Benzo(a)pyrene	0.400	0.370	7.5	72	0.00
38	Dibenzo(a,h)anthracene	0.400	0.354	11.5	71	0.00
39	Benzo(a,h,i)perylene	0.400	0.396	1.0	77	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
Data File : BE100392.D
Acq On : 5 Jul 2019 21:36
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 05 22:41:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				