

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
 Data File : BE100394.D
 Acq On : 5 Jul 2019 23:24
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 05 23:58:41 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jun 28 14:06:53 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	-0.02
2	1,4-Dioxane	0.621	0.607	2.3	104	-0.02
3	n-Nitrosodimethylamine	0.289	0.296	-2.4	137	-0.02
4 S	2-Fluorophenol	1.020	0.994	2.5	107	-0.01
5 S	Phenol-d6	1.342	1.293	3.7	106	-0.01
6	bis(2-Chloroethyl)ether	1.102	0.930	15.6	84	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	105	-0.03
8 S	Nitrobenzene-d5	0.397	0.326	17.9	88	-0.03
9	Naphthalene	4.411	4.314	2.2	101	-0.01
10	Hexachlorobutadiene	0.425	0.497	-16.9	119	-0.03
11 SURR	2-Methylnaphthalene-d10	0.752	0.723	3.9	99	-0.01
12	2-Methylnaphthalene	0.774	0.666	14.0	95	-0.01
13 I	Acenaphthene-d10	1.000	1.000	0.0	100	-0.01
14 S	2,4,6-Tribromophenol	0.280	0.245	12.5	89	-0.01
15 S	2-Fluorobiphenyl	1.559	1.613	-3.5	101	-0.01
16	Acenaphthylene	1.920	1.852	3.5	94	-0.01
17	Acenaphthene	1.088	1.078	0.9	94	-0.01
18	Fluorene	1.609	1.547	3.9	93	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
20	4-Bromophenyl-phenylether	0.255	0.269	-5.5	96	-0.01
21	Hexachlorobenzene	0.265	0.282	-6.4	97	0.00
22	Pentachlorophenol	0.080	0.020	75.0#	27#	0.01
23	Phenanthrene	0.969	0.900	7.1	88	-0.01
24	Anthracene	0.873	0.785	10.1	87	-0.01
25 SURR	Fluoranthene-d10	5.805	5.218	10.1	85	-0.01
26	Fluoranthene	1.544	1.365	11.6	82	-0.01
27 I	Chrysene-d12	1.000	1.000	0.0	75	-0.01
28	Pvrene	1.059	1.210	-14.3	84	-0.01
29 S	Terphenyl-d14	0.829	0.879	-6.0	79	-0.01
30	Benzo(a)anthracene	1.315	1.181	10.2	66	0.00
31	Chrysene	1.321	1.352	-2.3	75	-0.01
32	Bis(2-ethylhexyl)phthalate	2.255	2.048	9.2	70	-0.01
33	Indeno(1,2,3-cd)pyrene	1.747	1.701	2.6	68	0.00
34 I	Perylene-d12	1.000	1.000	0.0	73	0.00
35	Benzo(b)fluoranthene	1.084	1.013	6.5	68	-0.01
36	Benzo(k)fluoranthene	1.211	1.209	0.2	70	0.00
37 C	Benzo(a)pyrene	1.087	1.056	2.9	68	0.00
38	Dibenzo(a,h)anthracene	1.135	1.067	6.0	68	-0.01
39	Benzo(a,h,i)perylene	1.166	1.198	-2.7	72	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
Data File : BE100394.D
Acq On : 5 Jul 2019 23:24
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 05 23:58:41 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			