

Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
 Data File : BE100383.D
 Acq On : 5 Jul 2019 16:11
 Operator : JU/SJ
 Sample : K3558-17
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 EXT-016-SW100-062619

Manual Integrations
 APPROVED

mohammad
 7/8/2019 2:13:50 PM

Quant Time: Jul 06 00:39:00 2019

Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Fri Jun 28 14:06:53 2019

Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	700	0.40	ng	-0.02
7) Naphthalene-d8	10.42	136	2272	0.40	ng	-0.03
13) Acenaphthene-d10	14.30	164	1879	0.40	ng	-0.01
19) Phenanthrene-d10	17.04	188	5518	0.40	ng	-0.01
27) Chrysene-d12	21.23	240	7395	0.40	ng	-0.01
34) Perylene-d12	23.65	264	8981	0.40	ng	-0.01

System Monitoring Compounds

4) 2-Fluorophenol	5.20	112	346	0.19	ng	-0.01
5) Phenol-d6	6.83	99	418	0.18	ng	-0.01
8) Nitrobenzene-d5	8.81	82	325	0.14	ng	-0.01
11) 2-Methylnaphthalene-d10	12.02	152	797	0.19	ng	-0.03
14) 2,4,6-Tribromophenol	15.80	330	296	0.23	ng	-0.03
15) 2-Fluorobiphenyl	12.91	172	1156	0.16	ng	-0.03
25) Fluoranthene-d10	19.08	212	11788	0.15	ng	-0.02
29) Terphenyl-d14	19.68	244	2365	0.15	ng	-0.02

Target Compounds

						Qvalue
9) Naphthalene	10.47	128	4555	0.182	ng	# 97
12) 2-Methylnaphthalene	12.09	142	1002	0.228	ng	95
16) Acenaphthylene	14.02	152	1870	0.207	ng	97
17) Acenaphthene	14.37	154	180	0.035	ng	# 82
18) Fluorene	15.34	166	519	0.069	ng	# 84
23) Phenanthrene	17.08	178	15999	1.196	ng	99
24) Anthracene	17.17	178	2225	0.185	ng	# 92
26) Fluoranthene	19.11	202	36079	1.694	ng	99
28) Pyrene	19.48	202	49152	2.511	ng	96
30) Benzo(a)anthracene	21.22	228	27158m	1.117	ng	
31) Chrysene	21.27	228	32014	1.311	ng	97
32) Bis(2-ethylhexyl)phthalate	21.14	149	13299	0.329	ng	99
33) Indeno(1,2,3-cd)pyrene	26.20	276	21102	0.653	ng	# 97
35) Benzo(b)fluoranthene	22.91	252	32579	1.338	ng	# 98
36) Benzo(k)fluoranthene	22.95	252	13351m	0.491	ng	
37) Benzo(a)pyrene	23.54	252	26365	1.080	ng	# 96
38) Dibenzo(a,h)anthracene	26.22	278	5607	0.220	ng	# 79
39) Benzo(a,h,i)perylene	26.98	276	25605	0.978	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA E\DATA\BE070519\
Data File : BE100383.D
Acq On : 5 Jul 2019 16:11
Operator : JU/SJ
Sample : K3558-17
Misc :
ALS Vial : 10 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM

Quant Time: Jul 06 00:39:00 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA E\METHODS\8270-SIM-BE062819.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jun 28 14:06:53 2019
Response via : Initial Calibration





#1


1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.62 min Scan# 413
Delta R.T. -0.02 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
152	100		
150	133.2	114.5	171.7
115	68.2	57.1	85.7

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E


Ion 115.00 (114.70 to 115.70): E



#4

2-Fluorophenol
Concen: 0.194 ng
RT: 5.20 min Scan# 203
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
112	100		
64	55.8	39.4	59.0
63	200.0	104.5	156.7






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BE1

Ion 63.00 (62.70 to 63.70): BE1





#5

Phenol-d6

Concen: 0.178 ng

RT: 6.83 min Scan# 344

Delta R.T. -0.01 min

Lab File: BE100383.D

Acq: 5 Jul 2019 16:11

Instrument :

BNA_E





ClientSampleId :

EXT-016-SW100-062619

Tot Ion:	99	Resp:	418
Ion	Ratio	Lower	Upper
99	100		
42	15.6	11.2	16.8
71	38.3	37.2	55.8

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM



#7

Naphthalene-d8

Concen: 0.400 ng




RT: 10.42 min Scan# 635


Delta R.T. -0.03 min

Lab File: BE100383.D

Acq: 5 Jul 2019 16:11

Tgt Ion:	136	Resp:	2272
Ion	Ratio	Lower	Upper
136	100		
137	15.2	11.8	17.8
54	39.3	27.7	41.5
68	9.4	6.8	10.2





#8

Nitrobenzene-d5

Concen: 0.144 ng

RT: 8.81 min Scan# 511

Delta R.T. -0.01 min

Lab File: BE100383.D

Acq: 5 Jul 2019 16:11

Instrument :

BNA_E





ClientSampleId :

EXT-016-SW100-062619

Tot Ion:	82	Resp:	325
Ion	Ratio	Lower	Upper
82	100		
128	206.4	140.9	211.3
54	233.1	146.2	219.2#

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM



#9

Naphthalene

Concen: 0.182 ng




RT: 10.47 min Scan# 639


Delta R.T. -0.01 min

Lab File: BE100383.D

Acq: 5 Jul 2019 16:11

Tgt Ion:	128	Resp:	4555
Ion	Ratio	Lower	Upper
128	100		
129	4.9	3.0	4.6#
127	5.0	3.5	5.3









#11
 2-Methylnaphthalene-d10
 Concen: 0.187 ng
 RT: 12.02 min Scan# 750
 Delta R.T. -0.03 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Instrument :
 BNA_E
 ClientSampleId :
 EXT-016-SW100-062619

Tot Ion:152 Resp: 797
 Ion Ratio Lower Upper
 152 100
 151 23.6 15.1 22.7#




Manual Integrations
 APPROVED


mohammad
 7/8/2019 2:13:50 PM



#12
 2-Methylnaphthalene
 Concen: 0.228 ng
 RT: 12.09 min Scan# 755
 Delta R.T. -0.03 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Tgt Ion:142 Resp: 1002
 Ion Ratio Lower Upper
 142 100
 141 91.3 72.4 108.6
 115 59.0 40.4 60.6






#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.30 min Scan# 908
 Delta R.T. -0.01 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Instrument :
 BNA_E
 ClientSampleId :
 EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
164	100		
162	95.3	83.4	125.0
160	53.1	45.0	67.6

Manual Integrations
 APPROVED

mohammad
 7/8/2019 2:13:50 PM






Abundance

Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14
 2,4,6-Tribromophenol
 Concen: 0.225 ng
 RT: 15.80 min Scan# 1013
 Delta R.T. -0.03 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11




Tgt Ion	Ratio	Lower	Upper
330	100		
332	90.9	77.0	115.4
141	48.3	22.0	33.0#


Abundance

Ion 330.00 (329.70 to 330.70): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E






#15
2-Fluorobiphenyl
Concen: 0.158 ng
RT: 12.91 min Scan# 812
Delta R.T. -0.03 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
172	100		
171	40.5	31.9	47.9
170	32.8	22.2	33.2

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM






Abundance

Ion 172.00 (171.70 to 172.70): E


Ion 171.00 (170.70 to 171.70): E

Ion 170.00 (169.70 to 170.70): E



#16
Acenaphthylene
Concen: 0.207 ng
RT: 14.02 min Scan# 889
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
152	100		
151	22.2	16.6	25.0
153	14.2	10.5	15.7






Abundance

Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E






#17
Acenaphthene
Concen: 0.035 ng
RT: 14.37 min Scan# 913
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
154	100		
153	121.7	93.4	140.2
152	96.1	49.4	74.2#

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM






Abundance

Ion 154.00 (153.70 to 154.70): E


Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#18
Fluorene
Concen: 0.069 ng
RT: 15.34 min Scan# 979
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
166	100		
165	116.2	81.1	121.7
167	23.1	10.7	16.1#






Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.04 min Scan# 1106
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Resp	Lower	Upper
188	100	5518		
94	0.0	0.0	0.0	0.0
80	7.7	7.6	11.4	




Manual Integrations
APPROVED


mohammad
7/8/2019 2:13:50 PM



#23
Phenanthrene
Concen: 1.196 ng
RT: 17.08 min Scan# 1109
Delta R.T. -0.03 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	15999		
176	20.3	15.9	23.9	
179	15.8	12.2	18.2	





#24

Anthracene

Concen: 0.185 ng

RT: 17.17 min Scan# 1116

Delta R.T. -0.03 min

Lab File: BE100383.D

Acq: 5 Jul 2019 16:11

Instrument :

BNA_E





ClientSampleId :

EXT-016-SW100-062619

Tot Ion:	178	Resp:	2225
Ion Ratio	Lower	Upper	
178	100		
176	18.2	15.3	22.9
179	8.7	12.2	18.2#

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM



#25

Fluoranthene-d10

Concen: 0.147 ng




RT: 19.08 min Scan# 1360


Delta R.T. -0.02 min

Lab File: BE100383.D

Acq: 5 Jul 2019 16:11

Tgt Ion:	212	Resp:	11788
Ion Ratio	Lower	Upper	
212	100		
106	1.6	1.3	1.9
104	1.2	0.8	1.2









#26
Fluoranthene
Concen: 1.694 ng
RT: 19.11 min Scan# 1366
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
202	100		
101	5.4	4.6	7.0
203	17.8	13.7	20.5




Manual Integrations
APPROVED


mohammad
7/8/2019 2:13:50 PM



#27
Chrysene-d12
Concen: 0.400 ng
RT: 21.23 min Scan# 1619
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
240	100		
120	4.7	3.3	4.9
236	29.4	22.6	34.0









#28
Pvrene
Concen: 2.511 ng
RT: 19.48 min Scan# 1429
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Resp	Lower	Upper
202	100			
200	19.8	15.6	23.4	
203	20.5	14.0	21.0	




Manual Integrations
APPROVED


mohammad
7/8/2019 2:13:50 PM



#29
Terphenyl-d14
Concen: 0.154 ng
RT: 19.68 min Scan# 1465
Delta R.T. -0.02 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Resp	Lower	Upper
244	100			
212	43.8	27.3	40.9#	
122	8.0	4.7	7.1#	






#30
Benzo(a)anthracene
Concen: 1.117 ng m
RT: 21.22 min Scan# 1618
Delta R.T. -0.00 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
228	100		
226	31.7	23.2	34.8
229	24.0	15.7	23.5

Manual Integrations
APPROVED

mohammad
7/8/2019 2:13:50 PM






Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#31
Chrysene
Concen: 1.311 ng
RT: 21.27 min Scan# 1622
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11




Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.5	23.9	35.9
229	22.2	16.5	24.7


Abundance

Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E






#32
 Bis(2-ethylhexyl)phthalate
 Concen: 0.329 ng
 RT: 21.14 min Scan# 1611
 Delta R.T. -0.01 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Instrument :
 BNA_E
 ClientSampleId :
 EXT-016-SW100-062619




Tot Ion:	149	Resp:	13299
Ion Ratio	Lower	Upper	
149	100		
167	7.6	5.8	8.8
279	1.7	1.1	1.7

Manual Integrations
 APPROVED

mohammad
 7/8/2019 2:13:50 PM






Abundance Ion 149.00 (148.70 to 149.70): E
 Ion 167.00 (166.70 to 167.70): E
 Ion 279.00 (278.70 to 279.70): E




#33
 Indeno(1,2,3-cd)pyrene
 Concen: 0.653 ng
 RT: 26.20 min Scan# 2848
 Delta R.T. -0.02 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Tgt Ion:	276	Resp:	21102
Ion Ratio	Lower	Upper	
276	100		
138	8.9	9.4	14.0#
277	24.1	19.2	28.8

Abundance Ion 276.00 (275.70 to 276.70): E
 Ion 138.00 (137.70 to 138.70): E
 Ion 277.00 (276.70 to 277.70): E









#34
Pervlene-d12
Concen: 0.400 ng
RT: 23.65 min Scan# 2138
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
264	100		
260	27.4	20.8	31.2
265	34.0	21.8	32.6




Manual Integrations
APPROVED


mohammad
7/8/2019 2:13:50 PM



#35
Benzo(b)fluoranthene
Concen: 1.338 ng
RT: 22.91 min Scan# 1931
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.0	19.2	28.8
125	7.0	7.4	11.0









#36
Benzo(k)fluoranthene
Concen: 0.491 ng m
RT: 22.95 min Scan# 1942
Delta R.T. -0.02 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Instrument :
BNA_E
ClientSampleId :
EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
252	100		
253	25.8	17.9	26.9
125	8.0	5.0	7.4




Manual Integrations
APPROVED


mohammad
7/8/2019 2:13:50 PM



#37
Benzo(a)pyrene
Concen: 1.080 ng
RT: 23.54 min Scan# 2107
Delta R.T. -0.01 min
Lab File: BE100383.D
Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.1	19.6	29.4
125	7.6	9.0	13.6









#38
 Dibenzo(a,h)anthracene
 Concen: 0.220 ng
 RT: 26.22 min Scan# 2854
 Delta R.T. -0.02 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Instrument :
 BNA_E
 ClientSampleId :
 EXT-016-SW100-062619

Tot Ion	Ratio	Lower	Upper
278	100		
139	15.5	8.2	12.2#
279	37.2	20.3	30.5#

Manual Integrations
 APPROVED

mohammad
 7/8/2019 2:13:50 PM



#39
 Benzo(g,h,i)perylene
 Concen: 0.978 ng
 RT: 26.98 min Scan# 3069
 Delta R.T. -0.02 min
 Lab File: BE100383.D
 Acq: 5 Jul 2019 16:11

Tgt Ion	Ratio	Lower	Upper
276	100		
277	25.7	20.2	30.2
138	10.7	9.0	13.6

