

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070615\
 Data File : BE090057.D
 Acq On : 6 Jul 2015 17:27
 Operator : TP/IZ
 Sample : G2824-01DL 25X
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 X1SS0650006-150626DL

Quant Time: Jul 07 01:25:11 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.62	152	18658	5.00	ng	0.00
6) Naphthalene-d8	10.37	136	82217	5.00	ng	-0.02
12) Acenaphthene-d10	14.23	164	47781	5.00	ng	0.00
18) Phenanthrene-d10	16.96	188	112817	5.00	ng	-0.02
25) Chrysene-d12	21.12	240	134807	5.00	ng	-0.01
32) Perylene-d12	23.43	264	121274	5.00	ng	-0.02

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.25	112	892	0.21	ng	0.00
5) Phenol-d6	6.80	99	1233	0.21	ng	0.00
7) Nitrobenzene-d5	8.76	82	857	0.13	ng	0.00
13) 2,4,6-Tribromophenol	15.73	330	178	0.13	ng	0.00
14) 2-Fluorobiphenyl	12.85	172	2255	0.16	ng	0.00
27) Terphenyl-d14	19.58	244	3316	0.15	ng	-0.01

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Naphthalene	10.42	128	1226	0.07	ng	# 87
11) 2-Methylnaphthalene	12.04	142	261	0.02	ng	# 87
16) Acenaphthene	14.28	154	1190	0.10	ng	88
17) Fluorene	15.27	166	1044	0.06	ng	# 86
22) Phenanthrene	17.00	178	53842	1.86	ng	100
23) Anthracene	17.10	178	11328m	0.43	ng	
24) Fluoranthene	19.00	202	128674	3.99	ng	99
26) Pyrene	19.36	202	118885	3.70	ng	97
28) Benzo(a)anthracene	21.10	228	76055	2.26	ng	99
29) Chrysene	21.16	228	88087	2.52	ng	98
31) Indeno(1,2,3-cd)pyrene	25.83	276	34302	1.02	ng	97
33) Benzo(b)fluoranthene	22.73	252	91384	3.42	ng	98
34) Benzo(k)fluoranthene	22.77	252	41123m	1.38	ng	
35) Benzo(a)pyrene	23.33	252	63203	2.56	ng	98
36) Dibenzo(a,h)anthracene	25.85	278	8781	0.35	ng	97
37) Benzo(g,h,i)perylene	26.57	276	35558	1.38	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE070615\
Data File : BE090057.D
Acq On : 6 Jul 2015 17:27
Operator : TP/IZ
Sample : G2824-01DL 25X
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
X1SS0650006-150626DL

Quant Time: Jul 07 01:25:11 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jun 30 07:01:01 2015
Response via : Initial Calibration

