

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070615\
 Data File : BE090060.D
 Acq On : 6 Jul 2015 19:18
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 07 01:17:15 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	118	0.00
2	1,4-Dioxane	1.000	0.814	18.6	100	-0.01
3	n-Nitrosodimethylamine	1.000	0.815	18.5	100	-0.01
4 S	2-Fluorophenol	1.000	0.920	8.0	109	0.00
5 S	Phenol-d6	1.000	0.951	4.9	114	0.00
6 I	Naphthalene-d8	5.000	5.000	0.0	121	-0.01
7 S	Nitrobenzene-d5	1.000	0.858	14.2	106	0.00
8	Nitrobenzene	1.000	0.854	14.6	106	0.00
9	Naphthalene	1.000	0.921	7.9	113	-0.01
10	Hexachlorobutadiene	1.000	0.944	5.6	116	0.00
11	2-Methylnaphthalene	1.000	0.882	11.8	109	0.00
12 I	Acenaphthene-d10	5.000	5.000	0.0	113	-0.01
13 S	2,4,6-Tribromophenol	1.000	0.789	21.1	91	0.00
14 S	2-Fluorobiphenyl	1.000	0.909	9.1	105	-0.01
15	Acenaphthylene	1.000	0.918	8.2	107	-0.01
16	Acenaphthene	1.000	0.932	6.8	106	-0.01
17	Fluorene	1.000	0.889	11.1	103	-0.02
18 I	Phenanthrene-d10	5.000	5.000	0.0	104	-0.02
19	4-Bromophenyl-phenylether	1.000	0.870	13.0	92	-0.02
20	Hexachlorobenzene	1.000	0.896	10.4	97	-0.02
21	Pentachlorophenol	1.000	0.727	27.3#	70	0.00
22	Phenanthrene	1.000	0.873	12.7	93	-0.02
23	Anthracene	1.000	0.900	10.0	95	0.00
24	Fluoranthene	1.000	0.880	12.0	93	-0.01
25 I	Chrysene-d12	5.000	5.000	0.0	96	-0.01
26	Pyrene	1.000	0.952	4.8	93	-0.01
27 S	Terphenyl-d14	1.000	0.922	7.8	89	-0.01
28	Benzo(a)anthracene	1.000	0.919	8.1	91	-0.01
29	Chrysene	1.000	0.902	9.8	88	-0.01
30	Bis(2-ethylhexyl)phthalate	1.000	0.870	13.0	91	-0.01
31	Indeno(1,2,3-cd)pyrene	1.000	0.915	8.5	90	-0.04
32 I	Perylene-d12	5.000	5.000	0.0	91	-0.02
33	Benzo(b)fluoranthene	1.000	1.020	-2.0	94	-0.02
34	Benzo(k)fluoranthene	1.000	0.981	1.9	90	-0.02
35 C	Benzo(a)pyrene	1.000	1.011	-1.1	92	-0.03
36	Dibenzo(a,h)anthracene	1.000	0.950	5.0	88	-0.04
37	Benzo(g,h,i)perylene	1.000	0.948	5.2	87	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0