

Data Path : Z:\HPCHEM1\BNA_E\Data\BE070615\
 Data File : BE090060.D
 Acq On : 6 Jul 2015 19:18
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Jul 07 01:17:15 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BE062915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jun 30 07:01:01 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	118	0.00
2	1,4-Dioxane	0.686	0.558	18.7	100	-0.01
3	n-Nitrosodimethylamine	0.930	0.758	18.5	100	-0.01
4 S	2-Fluorophenol	1.149	1.057	8.0	109	0.00
5 S	Phenol-d6	1.608	1.528	5.0	114	0.00
6 I	Naphthalene-d8	1.000	1.000	0.0	121	-0.01
7 S	Nitrobenzene-d5	0.397	0.341	14.1	106	0.00
8	Nitrobenzene	0.421	0.360	14.5	106	0.00
9	Naphthalene	1.130	1.041	7.9	113	-0.01
10	Hexachlorobutadiene	0.178	0.168	5.6	116	0.00
11	2-Methylnaphthalene	0.755	0.666	11.8	109	0.00
12 I	Acenaphthene-d10	1.000	1.000	0.0	113	-0.01
13 S	2,4,6-Tribromophenol	0.147	0.116	21.1	91	0.00
14 S	2-Fluorobiphenyl	1.503	1.365	9.2	105	-0.01
15	Acenaphthylene	9.019	8.277	8.2	107	-0.01
16	Acenaphthene	1.300	1.212	6.8	106	-0.01
17	Fluorene	1.768	1.571	11.1	103	-0.02
18 I	Phenanthrene-d10	1.000	1.000	0.0	104	-0.02
19	4-Bromophenyl-phenylether	0.213	0.185	13.1	92	-0.02
20	Hexachlorobenzene	0.885	0.792	10.5	97	-0.02
21	Pentachlorophenol	0.086	0.055	36.0#	70	0.00
22	Phenanthrene	1.281	1.118	12.7	93	-0.02
23	Anthracene	1.172	1.054	10.1	95	0.00
24	Fluoranthene	1.431	1.259	12.0	93	-0.01
25 I	Chrysene-d12	1.000	1.000	0.0	96	-0.01
26	Pyrene	1.192	1.134	4.9	93	-0.01
27 S	Terphenyl-d14	0.833	0.768	7.8	89	-0.01
28	Benzo(a)anthracene	1.249	1.148	8.1	91	-0.01
29	Chrysene	1.297	1.170	9.8	88	-0.01
30	Bis(2-ethylhexyl)phthalate	0.398	0.344	13.6	91	-0.01
31	Indeno(1,2,3-cd)pyrene	1.243	1.138	8.4	90	-0.04
32 I	Perylene-d12	1.000	1.000	0.0	91	-0.02
33	Benzo(b)fluoranthene	1.103	1.125	-2.0	94	-0.02
34	Benzo(k)fluoranthene	1.232	1.208	1.9	90	-0.02
35 C	Benzo(a)pyrene	1.019	1.030	-1.1	92	-0.03
36	Dibenzo(a,h)anthracene	1.046	0.994	5.0	88	-0.04
37	Benzo(g,h,i)perylene	1.060	1.005	5.2	87	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0